



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
2100 RENAISSANCE BLVD.
KING OF PRUSSIA, PA 19406-2713

September 5, 2017

Peggy Pust, CRA, FAHRA
Vice President
Monongalia General Hospital
1200 J.D. Anderson Drive
Morgantown, WV 26505-3486

SUBJECT: VOIDANCE OF YOUR REQUEST FOR AMENDED LICENSE, MAIL CONTROL
NO. 599829

Dear Ms. Pust:

This concerns your request for an amended license dated May 18, 2017. Because you are not currently able to provide the additional information required to complete the requested action as noted by your response of "pending" for several items, we have voided your request. This action is taken without prejudice to the resubmission of your request. In particular, the additional information required to complete your request includes the following:

Possession Only of HDR

1. Dr. Saconn is licensed to practice medicine in NC; however is not listed on the West Virginia (WV) medical board site. Provide evidence that Dr. Saconn is licensed to practice medicine in WV.
2. We note that Mr. Murray and Ms. Pervola were previously listed on a license for high dose rate remote afterloader (HDR) use (LN 47-17282-01). Please confirm that this is correct.
3. Provide the serial number of the HDR unit to be supplied. If the exact number is unknown, provide a range of numbers that the unit is expected to fall within.

Medical Use of HDR

1. Confirm that the "person qualified" to perform survey instrument calibrations will have an NRC or Agreement State license to perform this service.
2. Confirm that annual emergency response training will be "hands on."
3. Provide a revised preceptor attestation statement for Dr. Saconn for HDR use that states that Dr. Saconn "has achieved a level of competency sufficient to function independently as an authorized user for remote afterloader units."
4. Submit the emergency procedures that will be posted near the treatment console, including the contact names and numbers for the authorized medical physicist (AMP), Radiation Safety Officer (RSO), and Authorized User (AU).
5. Clarify the training provided to AMPs, the AU, and the RSO as described in Section 8 of your application. For instance, Item A appears to apply to the AU and the RSO in addition to the AMPs and numbers xiv and xv of Item B appear to also apply to the AU. Finally, the audience for Item C of the training is not specified. Will this training be provided to AMPs and the AU?

6. Provide documentation of the vendor training provided to Dr. Saconn, Mr. Murray, Ms. Pervola, and Mr. Perna.
7. Provide step-by-step procedures for all spot checks.
8. In your response to Item 8, you indicated that HDR security was already described. However, the security measures taken when a source for exchange is stored in another room was not described. Please describe.

10 CFR 35.394 authorization for Dr. Saconn

We understand that you are pursuing the additional information to support this authorization (license/permit for 2 preceptors and third I-131 case > 33 mCi). Please provide.

In accordance with 10 CFR 2.390 of the NRC's "Agency Rules of Practice and Procedure," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records component of NRC's Agencywide Documents Access and Management System (ADAMS). ADAMS is accessible from the NRC Web Site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,



Penny Lanzisera, Senior Health Physicist
Medical Branch
Division of Nuclear Materials Safety
Region I

License No. 47-16259-01
Docket No. 03010683
Mail Control No. 599829

cc: Mark T. Perna, M.S., Radiation Safety Officer

6. Provide documentation of the vendor training provided to Dr. Saconn, Mr. Murray, Ms. Pervola, and Mr. Perna.
7. Provide step-by-step procedures for all spot checks.
8. In your response to Item 8, you indicated that HDR security was already described. However, the security measures taken when a source for exchange is stored in another room was not described. Please describe.

10 CFR 35.394 authorization for Dr. Saconn

We understand that you are pursuing the additional information to support this authorization (license/permit for 2 preceptors and third I-131 case > 33 mCi). Please provide.

In accordance with 10 CFR 2.390 of the NRC's "Agency Rules of Practice and Procedure," a copy of this letter will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records component of NRC's Agencywide Documents Access and Management System (ADAMS). ADAMS is accessible from the NRC Web Site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,

Penny Lanzisera, Senior Health Physicist
Medical Branch
Division of Nuclear Materials Safety
Region I

License No. 47-16259-01
Docket No. 03010683
Mail Control No. 599829

cc: Mark T. Perna, M.S., Radiation Safety Officer

DOCUMENT NAME: G:\WBL Documents\WBL License Cover Letter\L47-16259-01.599829.docx

[LAT to enter ADAMS Accession No.]

SUNSI Review Complete: PLanzisera

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	RI:DNMS	N					
NAME	PLanzisera <i>PL</i>						
DATE	9/5/17						

OFFICIAL RECORD COPY