Community Hospital St. Catherine Hospital St. Mary Medical Center

September 1, 2017

United States Nuclear Regulatory Commission Region III, Materials Licensing 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352

RE: Amendment NRC License No. <u>13-32478-01</u>
Remove location of use

Dear Sir/Madam:

We would like to request the removal of the Portage Heart Care, 6375 U.S. Highway 6, Suite B, Portage, Indiana facility from our NRC license number <u>13-32478-01</u>. We are no longer providing nuclear medicine services at this facility. Prior to performing a closeout survey, all radioactive sealed sources were transferred on August 7, 2017 to the 3545 Arbors Boulevard, Portage, Indiana facility listed on St. Mary Medical Center's NRC License # <u>13-03459-03</u>.

We have included the confirmation of receipt from St. Mary Medical Center that the sources were accepted into their inventory. In addition, please find the most recent sealed source leak test, the close-out survey results and diagram. Please note that Cobalt-57 flood source (s/n 1777-181) listed on the sealed source leak test was returned to the manufacturer prior to the closeout survey. We have enclosed the acknowledgement of receipt from Eckert & Ziegler.


If you have any questions, please contact our radiation safety officer, Sharon Updike, MHP, DABR at 517-795-8786.

Respectfully Yours,

Teresa A. Kempfer, MBA Director, Physician Practices

ivesalkempfr

Community Care Network



ST. MARY Medical Center

ACKNOWLEDGEMENT OF SEALED SOURCE RECEIPT

Please let this letter serve as an acknowledgement that the following sealed sources were received from Portage Heart Care and accepted into inventory at the 3545 Arbors Boulevard, Portage, Indiana facility under St. Mary Medical Center's NRC license #13-03459-03 on August 7, 2017:

Cs-137 Vial Source (s/n 63537) Cs-137 Check Source (s/n 52629) Ba-133 Vial Source (s/n 1218-45-2) Co-57 Flood Source (s/n 1934-021)

Respectfully Yours,

09/06/2017

Santosh K. Kar Radiation Safety Officer St. Mary Medical Center, 1500 South Lake Park Avenue, Hobart, IN, 46342



Sealed Source Leak Test

Licensee: Community Care Network-Portage Heart Care

Date: 04/24/17

Performed by: Sharon Updike

		Calibration	Calibratio	on		
Nuclide	Type	Activity	Date	Location	M/N	S/N
Cs-137	Vial	233.8 uCi	04/01/05	RAM Closet	MED3550	63537
Current Ac	tivity: 17	6.91uCi				
Co-57	Flood	20 mCi	12/01/14	RAM Closet	MED3709	1777-181
Current Ac	tivity: 2	.133mCi				
Co-57	Flood	20 mCi	03/01/17	RAM Closet	MED3709	1934-021
Current Ac	tivity: 1	7.42mCi				
Ba-133	Vial	244.5 uCi	04/01/07	RAM Closet	RV-133-250U	1218-45-2
Current Ac	tivity: 12	5.98uCi				

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The following Minimum Detectable Activities are based upon a background at the indicated value. Background was at or below these levels when the above tests were completed. Well Counter: Captus 3000

<u>Nuclide</u>	MDA	Background
Cs-137	2.0 x 10 ⁻⁴ uCi	145 counts/1 min
Ba-133	8.8 x 10 ⁻⁵ uCi	204 counts/1 min
Co-57	1.7 x 10 ⁻⁵ uCì	53 counts/1 min

RADIATION SAFETY OFFICER: _



Receiving 1800 Facility 1800 N. Keystone St. Burbank, CA 91504 US

Shipped From:
Portage Heart Care
Sharon Updike MHP
(517)795-8786
supdike@mpcphysics.com
6375 US HWY 6 Suite B
Portage, IN 46368
United States

Return acknowledgement

 Return Number:
 298703

 RMA number.
 : RMA_029538

 Date
 : 8/24/2017

 Page
 : 1 of 1

 Customer account.
 : PINTEC01

 Customer reference
 : Wendy Koehnen

 Customer requisition
 : Wendy Koehnen

Dear Valued Customer, Eckert & Ziegler Isotope Products Laboratories has received your radioactive source(s) and takes esponsibility for tracking and storage of the source(s) listed below. If you have any further questions about the source(s), please contact EZIP and reference the return number listed above. Thank you or your business.

tem number	Batch number	Serial number	Quantity returned	Received on	Description
43525-2-20M		1777-181	1.000000	8/21/2017	MED3709

Nuclide Info: Co-57 20.000000 mCi Ref. Date: 12/1/2014

Community Hospital St. Catherine Hospital St. Mary Medical Center

Close-out Survey/Information

Date performed: August 17, 2017

Performed by: Tracy Day, CNMT

Comments: Sealed sources and waste were removed from our department prior to performing the close-out survey.

Instruments

Wipe tests analyzed with a <u>Atomlab 500</u> Well Counter s/n: 13100172

Date Efficiency performed: 04/25/17

Radionuclide: <u>Cobalt-57</u> MDA for Co-57: <u>55.7 dpm</u> Conversion factor: <u>1.11 dpm/cpm</u> Radionuclide: <u>Cesium-137</u> MDA for Cs-137: <u>287 dpm</u> Conversion factor: <u>8.11 dpm/cpm</u>

Area survey performed with the following survey meter:

Manufacturer: Ludlum
Type: GM
Model Number: 14C
Serial Number: 208891

Probe Model: EW

Annual Calibration Performed: 1/13/17

Battery check acceptable: YES
Operational check acceptable: YES

Current reading: 5.5 mR/hr



Community Hospital St. Catherine Hospital St. Mary Medical Center

Historical Use of Radionuclides from inception of NRC license

Unsealed

Tc99m

Sealed

Co-57, Ba-133, Cs-137 – last date of use – 7/12/17

Last date of byproduct radiopharmaceutical use: Tc 99m - 7/12/17

Visual Check: The area was checked to ensure that all sealed sources and radioactive waste had been removed. No evidence of radioactive material was noted.

Sealed Source Leak Testing: No history of leaks from Sealed Sources.

Radiation Level Survey: No area within the department demonstrated radiation levels in excess of <u>0.02</u> mR/hr.

Removable Contamination: No area within the department demonstrated removable contamination in excess of <u>167</u> dpm's.

Imaging and Hot Lab, Stress Room See Diagram Attachment

Conclusion: No radioactive materials remain in this area. No removable contamination is present.

1500 S. Lake Park Avenue, Hobart, Indiana 46342 Phone: 219-947-6012 Fax: 219-947-6095

Portage Heart Care

License # 13-32478-01

CLOSEOUT SURVEY RESULTS

Date: 8/17/2017

Performed By: Tracy Day, CNMT


		Bkg. 496 cpm		Bkg. 0.02 mR/hr
Wipe			net dpm	gross mR/hr
1	Camera Gantry Base		0	0.02
2	Camera Gantry Base		66.00	0.02
3	Imaging Room floor		96.00	0.02
4	Imaging Room floor		52.00	0.02
5	Imaging Room floor		88.00	0.02
6	Imaging Room floor		102.00	0.02
7	Imaging Room floor		0	0.02
8	Imaging Room desk		99.00	0.02
9	Imaging Room floor		167.00	0.02
10	Imaging Room floor		58.00	0.02
11	Hot Lab Cabinet		99.00	0.02
12	Hot Lab Cabinet		63.00	0.02
13	Hot Lab counter		41.00	0.02
14	Imaging Room floor		167.00	0.02
15	Imaging Room floor		41.00	0.02
16	Imaging Room floor		66.00	0.02
17	Imaging Room floor		104.00	0.02
18	Imaging Room floor		80.00	0.02
19	RAM Storage Closet		0	0.02
20	RAM Storage Closet		33.00	0.02
21	Stress Room Bed		30.00	0.02
22	Stress Room Counter		82.00	0.02
23	Stress Room floor		0	0.02
24	Stress Room floor		0	0.02
25	Stress Room floor		0	0.02

Comments: No evidence of removable contamination

ACTION LEVELS: 2000 dpm/100 cm^2

< 0.2 mR/hr

COMMENTS: No evidence of removable contamination.



PORTAGE MEARTCARE (218) 762-0400 THE UPS STORE 03601 6212 US HIGHWAY 8 PORTAGE IN 46368-5057

1 LBS 1 OF 1 SHP NT: 1 LBS DATE: 12 SEP 2017

SHIP U.S. NUCLEAR REGULATORY COMMISSION TO: STE 210 2443 WARRENVILLE RD

UPS GROUND TRACKING #: 1Z 296 YX3 03 1023 9892



BILLING: PAP

RECEIVED SEP 13 2017