

# ACCELERATED DISTRIBUTION DEMONSTRATION SYSTEM

## REGULATORY INFORMATION DISTRIBUTION SYSTEM (RIDS)

ACCESSION NBR: 9203180013      DOC. DATE: 92/02/24      NOTARIZED: NO      DOCKET #  
 FACIL: 50-244 Robert Emmet Ginna Nuclear Plant, Unit 1, Rochester G      05000244  
 AUTH. NAME      AUTHOR AFFILIATION  
 MECREDY, R.C.      Rochester Gas & Electric Corp.  
 RECIP. NAME      RECIPIENT AFFILIATION  
                          NRC - No Detailed Affiliation Given

SUBJECT: Initial part 21 rept re Victoreen field calibr source S/N  
 109, model 878-10. Source accepted w/out meeting purchase  
 order requirements, creating potential to allow use of source  
 to calibr containment radiation monitors.

DISTRIBUTION CODE: IE19D      COPIES RECEIVED: LTR 1 ENCL 1 SIZE: 3  
 TITLE: Part 21 Rept (50 DKT)

NOTES: License Exp date in accordance with 10CFR2,2.109(9/19/72).      05000244

	RECIPIENT ID CODE/NAME	COPIES LTTR ENCL	RECIPIENT ID CODE/NAME	COPIES LTTR ENCL
	PD1-3 LA	1 0	PD1-3 PD	1 1
	JOHNSON, A	1 1		
INTERNAL:	AEOD/DSP/ROAB	1 1	NRR/DOEA/OGCB11	2 2
	NRR/DRIS/RVIB9D	1 1	<del>REG-FILE</del> 01	1 1
	RES/DSIR/EIB	1 1	RGN1	1 1
	RGN2	1 1	RGN3	1 1
	RGN4	1 1	RGN5	1 1
	SECY VANDERMEL	1 1		
EXTERNAL:	INPO RECORD CTR	1 1	NRC PDR	1 1
	NSIC SILVER, E	1 1		

NOTE TO ALL "RIDS" RECIPIENTS:

PLEASE HELP US TO REDUCE WASTE! CONTACT THE DOCUMENT CONTROL DESK,  
 ROOM P1-37 (EXT. 20079) TO ELIMINATE YOUR NAME FROM DISTRIBUTION  
 LISTS FOR DOCUMENTS YOU DON'T NEED!

TOTAL NUMBER OF COPIES REQUIRED: LTTR 18 ENCL 17



ROCHESTER GAS AND ELECTRIC CORPORATION • 89 EAST AVENUE, ROCHESTER N.Y. 14649-0001

ROBERT C. MECREDY  
Vice President  
Ginna Nuclear Production

TELEPHONE  
AREA CODE 716 546-2700



February 24, 1992

U.S. Nuclear Regulatory Commission  
Operations Center

SUBJECT: Notification of 10 CFR Part 21 Defect  
R.E. Ginna Nuclear Power Plant  
Docket No. 50-244

In accordance with 10 CFR Part 21, Section 21 (c)(3)(i), the enclosed initial notification is provided via facsimile, to the NRC Operations Center, at (301) 492-8187. Written notification, in accordance with 10 CFR Part 21, Section 21 (c)(3)(ii), will be provided by March 23, 1992.

Very truly yours,

  
Robert C. Mecredy

Enclosures

9203180013 920224  
PDR ADOCK 05000244  
S PDR

TE19

190126

Controlled Copy #4  
Refer to A-61 for  
Instructions

FORM 2CONTENT OF TRANSMITTAL TO NRCDate: 2/20/92

1. NAME: Robert E. Smith  
TITLE: V. P. Production and Engineering  
ADDRESS: Rochester Gas & Electric  
89 East Avenue  
Rochester, NY 14649
2. FACILITY: R.E. Ginna Nuclear Power Station  
DOCKET NO.: 50-244
3. Activity, or basic component supplied which fails to comply or contains a defect: \_\_\_\_\_  
Field calibration source S/N 109, Model 878-10
4. Firm supplying the basic component which fails to comply or contains a defect: \_\_\_\_\_  
Victoreen Inc.
5. Nature of the defect or failure to comply and the safety hazard which is created or could be created by such a defect or failure to comply: \_\_\_\_\_  
The field calibration source was accepted without meeting the purchase  
order requirements. This created the potential to allow use of the source  
to calibrate containment radiation monitors.
6. Date on which the information of such defect or failure to comply was obtained: December 10, 1991
7. In the case of a basic component which contains a defect or fails to comply, the number and locations of all such components in use at, supplied for, or being supplied at Ginna Station:  
NUMBER: One  
LOCATIONS: Radiochemistry, Hot Lab

Refer to A-61 for  
Instructions

FORM 2 (con't)

8. Corrective Action which has been, is being, or will be taken:

Nonconformance Report 91-613

Identified Deficiency Report 0109-91

- 8.1 Name of individual or organization responsible for the  
action:

Health Physics Dept/Nuclear Assurance

- 8.2 Length of time that has been or will be taken to complete the  
action: Field source will be calibrated and returned in April, 1992.

9. Any advice related to the defect or failure to comply about  
the facility, activity, or basic component that has been, is  
being, or will be given to Rochester Gas and Electric:

Recalibration of field source is being completed at this time.

For any additional information, contact Thomas Morlow,

Superintendent Ginna Production, at 315-524-4446 EXT. 315



25