

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:  Insight Health Corporation 5775 Wayzata Blvd., Suite 400 Minneapolis, MN 55416 Location Inspected: Stamford, CT  REPORT NUMBER(S) 2017002	2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S)  030-38590	4. LICENSE NUMBER(S)  22-29403-01	5. DATE(S) OF INSPECTION  October 4, 2017
--------------------------------------	---	---

**LICENSEE:**  
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):  
 \_\_\_\_\_

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
 (Violations and Corrective Actions)

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Lester Tripp	<i>[Signature]</i> (for LT)	11/01/2017
BRANCH CHIEF	Aaron McCraw	<i>[Signature]</i>	11/02/2017

**Docket File Information**

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: Insight Health Corporation 5775 Wayzata Blvd., Suite 400 Minneapolis, MN 55416 Location Inspected: Stamford, CT REPORT NUMBER(S) 2017002	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
---	---

3. DOCKET NUMBER(S) 030-38590	4. LICENSE NUMBER(S) 22-29403-01	5. DATE(S) OF INSPECTION October 4, 2017
----------------------------------	-------------------------------------	---

6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 3.01 - 3.07
--	--

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S) 02220	2. PRIORITY 3	3. LICENSEE CONTACT Marc Renn - Area Manager, PET/CT	4. TELEPHONE NUMBER (570) 778-4350
-----------------------------	------------------	---	---------------------------------------

Main Office Inspection                      Next Inspection Date:                      No Change

Field Office Inspection

Temporary Job Site Inspection 32 Strawberry Hill Court, Stamford, CT

**PROGRAM SCOPE**

This was an unannounced assist inspection of a mobile nuclear medicine service provider authorized to use fluorine-18 for PET scans. The scope of this inspection was limited to the conduct of licensed activities on the premises of Stamford Hospital's Tully Health Center in Stamford, Connecticut. The licensee performed 4-6 studies here every Wednesday, Saturday, and Sunday. The mobile nuclear medicine van, in which the licensee performed these studies, was located on a pad behind the parking garage on the right side of the facility, and was most easily accessed from Morgan Street in the loading/facilities area. Program oversight was provided by corporate radiation officer and West Physics, a medical physics consulting company. Oversight also included an area manager and local radiation safety officer.

The inspector interviewed the nuclear medicine technologist (NMT) and the patient manger/orderly and found them to be knowledgeable with regard to the regulations, radiation safety practices, and security practices. The NMT and patient manager were observed during patient interaction, dose assay, patient scans, and end of the day contamination and exposure surveys. Independent inspector radiation measurements were consistent with licensee postings and regulatory requirements. A review of the following records confirmed statements made by licensee staff: semi-annual audits, annual program review, dosimetry, instrument calibrations, surveys, leak tests, inventory, and training. All records are kept electronically with the ability to print copies as needed.

The inspector toured the coach to evaluate the license's measures for radioactive material security and exposure control. Material security and control of exposure were both found to be adequate. During observation of the NMT and the patient manager/orderly, both were observed wearing the proper dosimetry, laboratory coats and gloves. The inspector conducted confirmatory exposure surveys of the van and the surrounding areas. The results of these survey were similar to the licensee's results.

Through observations of licensee personnel, demonstrations and discussion, the inspector found the licensee's personnel to be knowledgeable with regards to radiation safety principles, and regulatory requirements.

Within the scope of this inspection, no violations were identified.