ACCELERATED DISTRIBUTION DEMONSTRATION SYSTEM

REGULATORY INFORMATION DISTRIBUTION SYSTEM (RIDS) ACCESSION NBR:9404140199 DOC.DATE: 94/03/29 NOTARIZED: NO DOCKET # FACIL:STN-50-528 Palo Verde Nuclear Station, Unit 1, Arizona Publi STN-50-529 Palo Verde Nuclear Station, Unit 2, Arizona Publi STN-50-530 Palo Verde Nuclear Station, Unit 3, Arizona Publi 05000528 05000529 05000530 AUTHOR AFFILIATION AUTH.NAME R CONWAY, W.F. Arizona Public Service Co. (formerly Arizona Nuclear Power RECIPIENT AFFILIATION RECIP.NAME I Document Control Branch (Document Control Desk) SUBJECT: Forwards technical justification for containment high-range D radiation monitor calibr methodology. S. DISTRIBUTION CODE: A001D COPIES RECEIVED:LTR SIZE: ~ ENCL TITLE: OR Submittal: General Distribution NOTES: STANDARDIZED PLANT 05000528 Α 05000529 Standardized plant. Standardized plant. 05000530 D Ď RECIPIENT COPIES RECIPIENT COPIES LTTR ENCL LTTR ENCL ID CODE/NAME ID CODE/NAME S 4 PDIV-3 LA 1 1 PDIV-3 PD 1 1 HOLIAN, B 2 2 TRAN, L . 2 2 INTERNAL: ACRS 6 6 NRR/DE/EELB 1 1 NRR/DORS/OTSB 1 1 NRR/DRCH/HICB 1 1 1 NRR/DRPW 1 1 NRR/DSSA/SPLB 1 NRR/DSSA/SRXB 1 1 NUDOCS-ABSTRACT 1 1 1 0 OC/LFDGB 0 OGC/HDS1 1 REG-FILE 01 * 1 1 EXTERNAL: NRC PDR 1 1 NSIC 1 1

NOTE TO ALL "RIDS" RECIPIENTS:

PLEASE HELP US TO REDUCE WASTE! CONTACT THE DOCUMENT CONTROL DESK, ROOM PI-37 (EXT. 20079) TO ELIMINATE YOUR NAME FROM DISTRIBUTION LISTS FOR DOCUMENTS YOU DON'T NEED!

TOTAL NUMBER OF COPIES REQUIRED: LTTR 24 ENCL 22

R

Ι

D

S

/

Α

D

D

S

Arizona Public Service Company P.O. BOX 53999 • PHOENIX, ARIZONA 85072-3999

WILLIAM F. CONWAY EXECUTIVE VICE PRESIDENT NUCLEAR 102-02886-WFC/AKK/DLK March 29, 1994

U. S. Nuclear Regulatory Commission ATTN: Document Control Desk Mail Station P1-37 Washington, DC 20555

References: 1. Letter dated November 30, 1993, from B. E. Holian, USNRC, to W. F. Conway, APS, "Summary of Meeting Held on October 19, 1993, to Discuss Radiation Monitor Calibration Methods"

2. Letter 102-02741, dated November 29, 1993, from W. F. Conway, APS, to USNRC, "Reply to Notice of Deviation 50-528/93-03-01"

Dear Sirs:

9404140199 940329

000528 PDR

Subject: Palo Verde Nuclear Generating Station (PVNGS) Units 1, 2, and 3 Docket Nos. STN 50-528/529/530 Technical Justification for the Containment High-Range Radiation Monitor Calibration Methodology File: 94-016-545; 94-056-026

Representatives from Arizona Public Service Company (APS) met in Washington DC with the Office of Nuclear Reactor Regulation (NRR) to present the calibration methodology, used by APS, for the containment high-range radiation monitors and to provide a forum, at a working group level, for questions and answers. A summary of the meeting and the overheads used are included in Reference 1. By letter dated November 29, 1993 (Reference 2), APS committed to forward to NRR the technical justification, along with surveillance test results that correlated the performance of the containment high-range radiation monitors to the primary calibration (performed by the manufacturer). Enclosure 1 to this letter is the technical justification. The surveillance test results that correlate the performance to the primary calibration (performed by the primary calibration (performed by the manufacturer) are included in Enclosure 2.



·

.

U. S. Nuclear Regulatory Commission ATTN: Document Control Desk Containment High-Range Radiation Monitor Calibration Methodology Page 2

Should you have any questions, please contact Angela K. Krainik at (602) 393-5421.

Sincerely,

won

5

"

WFC/AKK/DLK/rv

Enclosures:

- 1. Technical Justification
- 2. Test Results Correlation to the Primary Calibration
- cc: K. E. Perkins, Jr. K. E. Johnston

· ,' · · ·

х Х Х Х Х

¥.

. . .

ENCLOSURE 1

TECHNICAL JUSTIFICATION

FOR

CONTAINMENT HIGH-RANGE RADIATION MONITOR CALIBRATION

METHODOLOGY

•

. ,

т. ч. В 94 – Ф • \$ **6**=

аг т 1 т ٠ţ

•

۰,

٠ f

TECHNICAL JUSTIFICATION FOR CONTAINMENT HIGH-RANGE MONITOR CALIBRATION METHODOLOGY

Arizona Public Service Company (APS) uses a transfer calibration methodology to verify detector accuracy and operability through the use of a transfer radiation source. APS' transfer calibration procedure is essentially the procedure specified by the equipment manufacturer to comply with the NUREG-0737 recommendations, and is consistent with the transfer calibration methodology used by the two other major manufacturers of containment radiation monitors.

No stipulation is made in Table II.F.1-3 of NUREG-0737 that an <u>in situ</u> calibration needs to fulfill the same requirements of a primary calibration. Moreover, the NUREG does not specifically endorse the use of an external calibration source, nor does it prohibit the use of an internal calibration source. It simply recommends that "<u>in situ</u> calibration for at least one decade below 10 R/hr shall be by means of a calibrated radiation source." Based on correspondence and discussions with the Office of Nuclear Reactor Regulation (NRR), APS recognizes that NRR interprets this recommendation to imply the use of a calibrated external source. This submittal presents the technical basis and justification supporting the use of an alternate <u>in situ</u> transfer calibration methodology using a source internal to the detector. It demonstrates that using an internal source is technically equivalent to using an external source when performing transfer calibrations and meets the overall intent of the <u>in situ</u> calibration recommendations contained in Table II.F.1-3 of NUREG-0737.

The basis for the APS' transfer calibration methodology can be explained by examining the following three areas:

- the differences between primary and transfer in <u>situ</u> calibrations, and the ability of the transfer calibration source to serve as a "calibrated radiation source";
- the dual role of the Am-241 source and the specific procedural requirements imposed that allow it to be used as a transfer calibration standard; and
- the principles of ion chamber operation that allow the use of an internally mounted alpha emitter (Am-241) to demonstrate the accuracy and functionality of an ion chamber used to measure external beta/gamma radiation.

Primary vs. Transfer Calibration

The purpose of a primary calibration is to <u>quantify</u> the performance characteristics of a particular design of radiation detector. Primary calibrations of each detector having the same design are not necessary as long as subsequent calibrations correlate each detector's response to radiation to the response measured on the specific detector that underwent the primary calibration. This is typically accomplished using radioactive solid

Page 1 of 6

р ја Р 4

* ** ** · · · • . **'**•

· · . 4

• • • • • •



• •

. . 8 N а , ч

" • lasources measured in a fixed geometry. During the primary calibration, the detector's response to these sources is measured and recorded. Subsequent calibrations on other detectors of the same design require that these detectors have the same response to the same sources (or other sources made with the same isotopes and geometry), as measured on the detector that underwent the primary calibration. This calibration methodology is known as a transfer calibration. The purpose of a transfer calibration is only to verify that the response of each detector to radiation is similar to the response measured on the detector that underwent a primary calibration. The sources used are referred to as "transfer calibration" sources. If the transfer calibration is performed in the field, as opposed to removing the detector from the field and transporting it to a calibration facility, the transfer calibration is referred to as an <u>in situ</u> calibration.

The following describes the five basic parts of the primary calibration performed on the containment high-range radiation monitors used at Palo Verde.

- 1) Electronic Alignment In the case of containment high-range area monitors, this consists of setting detector operating voltage and performing an electronic calibration of the detection channel using a suitable pico amp current source. This methodology is described in NUREG-0737 recommendations.
- Energy Calibration Since collection efficiency of an ion chamber depends 2) on ion density, and not on how the ionization is produced, the linearity characteristics of the ion chamber should be independent of the type or spectrum of the radiation involved (provided that the intensity of the radiation field is reasonably uniform over the volume of the ionization chamber). Furthermore, since ion chambers of a particular design can be manufactured so that they are essentially identical, this calibration can be performed on a "type" basis. For the containment high-range monitors, this calibration was performed to demonstrate a sensitivity to photon energies as low as 60 KeV and an essentially flat response of the detection system in terms of exposure rate in the incident photon energy range of 0.1 to 3.0 MeV. The energy calibration was performed as part of the primary calibration and is documented in "Report of Calibration, Model KDA-HR Ion Detector (P/N 824636-001)", Kaman Report Number: Chamber K-82-70-U(R).
- 3) Linearity Testing This consists of exposing the detector to widely spaced points in its measurement range and verifying linear response. For containment high-range detectors, this test is accomplished for exposure rates through 10⁷ R/hr on a type basis for the same reasons explained above. Linearity testing was performed as part of the primary calibration and is documented in "Report of Calibration, Model KDA-HR Ion Chamber Detector (P/N 824636-001)", Kaman Report Number: K-82-70-U(R). Every

. ¥. a * • 1 N

، ۲ ٩ и • ` 2 ſ

. د

,

• ₽- ***** • 10 . di + + ***** U • у

• *** • ِنْہِ *

ji 🖡 5 -

• • •

,

.

•

subsequently manufactured detector that is purchased is tested through When these detectors are purchased, calibration data 10³ R/hr. documenting this test is supplied with the detector. This methodology follows the guidance recommended in NUREG-0737 and later clarified by D. G. Eisenhut.¹

4) Primary Calibration - For containment high-range detectors, this consists of exposing the detector in an NIST (formerly NBS) traceable gamma radiation field and establishing the output of the detector in terms of R/hr/amp referenced to a particular isotope (generally, Cesium (Cs-137) or Cobalt (Co-60)). This part of the calibration was performed in conjunction with the linearity testing described above. Therefore, each detector installed in the field undergoes a primary calibration (exposure through 10³ R/hr) before it is shipped by the manufacturer to ensure correct response.

Transfer Source Testing - This test consists of exposing the detector to a 5) source that will serve as a transfer standard to ensure proper operation at subsequent field in situ calibrations. The response to the transfer calibration source observed during the primary calibration is correlated to the expected response in the field. As long as the responses are similar, the detector is expected to respond to radiation in the field as it did during the primary calibration. The transfer calibration source becomes a "calibrated source" when it is linked to the primary calibration by this method. The transfer source is used only as a link of traceability to the primary calibration and need not be NIST traceable. In the case of the containment high-range monitor, the manufacturer records each detector response to the internal Am-241 source observed during linearity and primary calibration testing (through 10³ R/hr) performed at the factory. This information is supplied with every detector. The Am-241 source response measured and recorded as part of the primary calibration forms the link between the primary calibration and subsequent field calibrations making the Am-241 acceptable as a transfer calibration source.

While all parts of the primary calibration process discussed above are important to quantify the performance characteristics of the detector, not all of the parts are required to be repeated in the field in order to verify detector accuracy and operability. Due to design configuration control, full range linearity and energy response testing are performed by the manufacture on a "type" basis. Because of ALARA considerations, the detection efficiency and linearity of each production detector is verified to be the same

Letter dated August 16, 1982, from D. G. Eisenhut, USNRC, to Regional Administrators, "Proposed Guidance for Calibration and Surveillance Requirements for Equipment Provided to Meet Item II.F.1, Attachments 1, 2, and 3, NUREG-0737"

· · · ·

•

· · · · ·

• • •

, ,

as the type tested detectors through 10³ R/hr before they are installed in the field. This calibration may be performed by the manufacturer. Therefore, the <u>in situ</u> transfer calibration performed on the containment high-range monitors need only consist of electronic alignment and verification of the transfer source response in order to ensure that the monitor is functioning accurately for its intended purpose. This overall methodology is consistent with the specific recommendations presented in NUREG-0737.

All three major US manufacturers of containment high-range area monitoring systems follow this calibration methodology and utilize a transfer calibration method to verify that the detector and monitor are functioning accurately after field installation. Palo Verde Nuclear Generating Station is no exception.

By way of comparison, Victoreen Instrument Corporation follows the methodology as stated establishing traceability to the primary calibration through the use of a modified version of the Victoreen Field Calibration Kit (FCK). The Victoreen FCK consists of a Cs-137 source mounted in a shielded holder that fixes the detector to source geometry. The FCK is not directly NIST traceable in terms of exposure rates. Traceability is linked only to the source response observed during the primary calibration.

GA/Sorrento Electronics uses a similar method in the form of their RT-11 calibrator. Again, a Cs-137 source in a shielded holder is used. The holder is subsequently hung on the side of the detector in a known geometry. The RT-11 is not NIST traceable in terms of exposure rate. Traceability is only linked to the primary calibration.

Although other manufacturers decided to utilize external transfer calibration sources, Kaman elected to use an internal source. The Palo Verde Nuclear Generating Station calibration method is essentially the method prescribed by the original equipment manufacturer (Kaman) with tighter tolerances applied to the transfer source response. In this case, the transfer calibration source consists of an Am-241 source mounted in a fixed location within the detector itself. This source is present at all times and its output is directly traceable to the primary calibration of the detector.

NUREG-0737 recommended that the detector's response be verified with a "calibrated source". All three major manufacturers of containment high-range monitors use a transfer calibration methodology to accomplish this requirement and to routinely assess detector performance in the field. The use of any transfer source, externally or internally mounted, satisfies the intent of the NUREG-0737 recommendation to verify detector response, provided that it is traceable to the factory calibration and that it satisfactorily demonstrates the detector and monitor are functioning accurately after field installation.

č

Dual Role of Am-241

Am-241 is used as a check source to verify detector operability during continuous operation. The containment high-range monitor uses the continuous current response generated by the internal source to detect detector failure (i.e., a loss of detector signal). This is only one application of the internal source. PVNGS calibration procedures require the Am-241 source response fall within a prescribed range (\pm 30%) of the response recorded during the primary calibration. Under this application, the PVNGS calibration procedures use the internal source response as a quantitative assessment which is directly traceable to the primary calibration. As such the PVNGS calibration procedures a "calibration." This is consistent with the transfer calibration methodology discussed above.

Use of Alpha Emitter to verify Ion Chamber Operation

An ion chamber detects radiation via its interaction with gas contained within the chamber. This interaction produces ions and electrons in an electric field that eventually are detected as a current response between the detector's electrodes. The type of gas in the detector, the type of radiation, or the energy of the radiation does not greatly effect the energy dissipation per ion pair produced by the interaction between the radiation and the gas. In other words, the energy dissipation of fast electrons and alpha particles is essentially the same, and the amount of energy deposited within the chamber is proportional to the number of ion pairs formed (ion.density) by the interaction of the radiation with the gas.² Therefore, collection efficiency of the ion chamber depends on ion density produced within the chamber, and not on how the ionization originated.

The overall detector sensitivity to external radiation is a function of the mechanical design of the ion chamber (i.e., construction material, wall thickness, detector diameter, etc.). Since ion chambers of a particular model can be manufactured so that they are essentially identical with respect to mechanical design, the practice of type testing detectors becomes applicable. This is the basis for performing energy response and linearity testing through 10⁷ R/hr in order to quantify response characteristics for a particular model of detector rather than repeat the test on each detector manufactured. Therefore, as long as the detectors are manufactured to the same mechanical design, the only parameters that can effect the ion collection efficiency within the chamber is the fill gas, its purity, fill pressure, and operating voltage.

Performing an electronic alignment and subsequent linearity testing through 10³ R/hr on each detector manufactured ensures that the mechanical design of the detector and parameters affecting ion collection efficiency are consistent with that of the detectors that

² Knoll, Glenn F., <u>Radiation Detection and Measurement</u>, John Wiley and Sons, New York, NY, 1979 (pages 152-153).

*

• .

, ,

, **•**

γ

, ,

were type tested for linearity and energy response. As part of this factory test, the response to the internal alpha source is recorded and the information is used later in subsequent field calibrations. When the detector is calibrated in the field it undergoes an electronic alignment and verification of its recorded response to the transfer calibration source (Am-241). The field test not only verifies that the ion collection efficiency of the detector has not changed but also verifies that the response of the detector is not affected by any inaccuracy of the electronic components used to measure the detector output or by any factor that could challenge connector and signal cable integrity. Because the detector mechanical design has already been verified in the laboratory and is not altered by field installation, it is not necessary to reverify the mechanical design by performing an in situ calibration in a known radiation field. Since ion collection efficiency within the chamber is independent of the type or energy of the radiation, either an alpha, beta, gamma, or beta/gamma source may be used to produce the ionization. Therefore, the use of an internally mounted alpha source rather than an external beta/gamma or gamma source can be used to test the parameters that affect ion collection efficiency within the chamber and operating characteristics of the monitor.

The use of a high activity beta/gamma or gamma source mounted externally in close proximity to the detector is not an adequate test to verify the mechanical design of the detector with respect to overall energy response because of the highly variable radiation fields. To test the overall energy response of an ion chamber detector, the detector should be exposed in a known radiation field that is relatively uniform with respect to the internal dimensions of the detector. Therefore, the use of an external transfer calibration source under these circumstances provides no advantage over the use of an internal Am-241 source. The same information regarding detector performance can be obtained from either type of source. In fact, the use of a high activity external source can induce a non conservative bias in the response if back scatter (a phenomenon caused by the interaction of the radiation with structures and equipment located near the detector) is not quantified or measured. The use of the internal Am-241 source eliminates back scatter effects.

Finally, NUREG-0737 recommends that the response of the detector be checked at one point below the range of 10 R/hr. The detector response from the internal Am-241 source produces a current that is within the range of 1E-11 to 5E-11 amps. Based on the detector efficiency of 1E+11 R/hr/amps, this response is equivalent to a response in the range of 1 to 5 R/hr. Therefore, the current response falls within the range recommended in NUREG-0737.

. . . .

.

· · · · · · •

• , • • • • •

. • • · · ·

i i

, ,

. ***** .

· · • •

•

ENCLOSURE 2

SURVEILLANCE TEST RESULTS CORRELATION

ТО

THE PRIMARY CALIBRATION

. Тар

• •

· · · · · •

ň * . а 5. 6. 7. и

•

• ••

SURVEILLANCE TEST RESULTS CORRELATION TO THE PRIMARY CALIBRATION

As explained in the technical justification section, the particular model of containment high range detector used at Palo Verde was "type" tested to determine its detection efficiency to gamma radiation; its energy response characteristics; and its linearity characteristics through the range 10⁷ R/hr. Results of the type test are documented in "Report of Calibration, Model KDA-HR Ion Chamber Detector (P/N 824636-001)", Kaman Report Number: K-82-70-U(R). Every subsequently manufactured detector purchased for field use was tested at the factory by exposing it to a calibrated radiation field in order to verify (relative to the type test) its detection efficiency and linearity through 10³ R/hr. During this factory calibration, the detector response to the various radiation fields, along with the baseline current from the internal Am-241 source, were measured and recorded on factory calibration data sheets. The calibration data documenting this factory test were supplied with each detector purchased and are traceable by the detector serial number.

Figures 1 through 6 are copies of the data sheets supplied with the detectors currently installed at Palo Verde. The data sheets contain both the keep alive source reading observed at the time of calibration (identified on the data sheet for convenience) and the data collected when the detector was placed in a calibrated radiation field. The specific keep alive current response recorded during the factory calibration was incorporated into surveillance test procedure 74ST-9SQ23, "Radiation Monitoring Calibration Test for New Scope Monitors." The surveillance test procedure acceptance criteria are \pm 30% of the keep alive source response measured during the factory calibration. The specific criteria used in the procedure are summarized in Table 1.

UNIT	MONITOR	DETECTOR SERIAL NUMBER	KEEP ALIVE CURRENT DURING FACTORY CALIBRATION (IN AMPS)	ALLOWABLE RANGE (IN AMPS)
1	RU-148	22706	1.38E-11	9.66E-12 TO 1.79E-11
_	RU-149	22711	1.16E-11	8.12E-12 TO 1.51E-11
	RU-148	22709	1.00E-11	7.00E-12 TO 1.30E-11
2	RU-149	53364	2.00E-11	1.40E-11 TO 2.60E-11
	RU-148	62165	2.38E-11	1.67E-11 TO 3.09E-11
3	RU-149	62163	2.00E-11	1.40E-11 TO 2.60E-11

Table 1 - Keep Alive Source Current Data Summary

Page 1 of 14

ی

* A

Finally, Figures 7-12 are copies of the applicable pages of 74ST-9SQ23 that were obtained from the most current calibrations performed on the containment high range monitors in all of the units. These data pages from the procedure demonstrate that the keep alive source response of the currently installed detectors fell within \pm 30% of the source response measured during the factory calibration.

• ÷ 1 – . . 1 ۰ ۲.,۴ ¥ * u '-

•

F

DOCUMENT

STANDARD PRACTICE PROCEDURE

18-104 AEV.

PADE 14 OF14 - A

10.0: DATA SHEET -: KDI-1000 PRIMARY CALIBRATICN, C1-137

laman Instrumentation Carperation

· · · · · · · · · · ·

.

Calibration: Date 09 MAY 85. Customer/Channel H-SQA-RE-150 KI Detector Part No. <u>824636-003</u> Afgr. Sarial Ho. _22706 ITEN. HODEL S/# CAL DUE Picoameter KEITHLEN 616 627239 28UNSO 10784052 HV Supply. CANBERRA 3002 OZSEPHS KENTHLEY DVH 130 64927 · 13 JAN 86 Xeep-Alive Current (Acceptance Limits = 0.9 to 5.5 x E-11A): 18. Initial Reading .10 (E-9 Coulombs) 20. Final Reading _____ (E-9 Coulcusts) 21. Time Interval 79 Second (60 seconds ariniams) 22. .Current:= (Stap 20 - Step 18)/Step 21 = 1.38 E-1 Apparas 3.04. R/hr readings: 28. Initial Reading / / OC (E-9) Coulombs 0A
 30. Final Reading
 3.53
 (E-9) Coulombs

 31. Time Interval
 60
 Seconds (60 seconds minimum)
 33. Current = (Step 30 - Stap 28)/Step 31 = <u>A.30 E-11</u> Amperos 34. Net current (Step 33 - Step 22) = 2.925-11 Appares 35. Efficiency (Step_34/3.04)) = ______96.E-+1 A/R/hr = 30.1 R/hr readings: 41. Gross current = 3,20 .: E-10 Amoras 43. Net current (Step 41 - Step 22) = 3.06 _ E-10 Amperas 44. Ifficiancy (Stap 43/30.1) = 1.02. E-11 A/R/hr = 299 R/hr. readings: 50. Gross current = 3.19 E-9 Appares . · 52. Not current (Step 50 - Step 22) = 3.13 E-9 Amperes 53. Efficiency (Step 52/299) * 1.Ch E-11 A/R/hr * * Acceptance Limits = 0,8 to 1,2 x E-11 A/R/hr By Kaliberrow Bata 09 Antros QA Califin Data 5-16-85 Figure 1 - Factory Calibration Data Sheet for Detector Serial Number 22706

Page 3 of 14



< . Way the • .

***** 7 *

۰ ۲

.

. •

.

и **Х**

٩ , , , •

•

DECUMENT

KN7 20.

STANDARD PRACTICE PROCEDURE

10.0 DATA SHEET - KDI-IGOD PRIMARY CALIBRATICH. Ca-137

Calibration Date 09 May 85 . Customer/Channel 1-J-SQB-RE-49 KI Detector Part No. 8246-003 Hfgr. Serial No. 22711 ITEM AFGR MODEL S/1 CAL DUE Picoametor REITHLEY 627234 مال SLUN25 HV Supply CANGEROA 3002 10784052 02 560 85 DYN KATHERY -LV427 . 130 13 JAN 86 Reep-Alive Current (Acceptance Limits' = 0.9 to 5:5 x'E-11A): 18. Initial Reading ____/O (E-9 Coulcas) 20. Final Reading _____/4" (E-9 Coulombs) 21. Time-Interval 90 Second (60 Seconds minimum) 22. Current = (Step 20 - Step 18)/Step 21 = 1.16 E-11 Amounts 3.04 R/hr readings: 28. Initial Reading 1.00 (E-9) Coulcubs 30. Final Reading 3.38 (E-9) Coulcuts .31. Time: Interval: 60 Seconds (60 seconds minimum) 33. Current = (Stap 30 - Stap 28)/Stap 31 = 397 E-11 Amparas 34. Het current (Stap 33 - Step 22) = 2,81 E-1 Apperes 35. Efficiency (Step 34/3.04)) = ;92 E-11 A/R/hr * 30.1' R/hr readings: 41. Gross current = 3,13 E-10 Ampores 43. Net current (Step 41 - Step 22) = 3.01 E-10 Ampares 44. Efficiency (Stop 43/30.1) = ____ (100 E-4) A/R/hr * ... 299 R/hr readings: 50. Gross current = 3.03 E-9 Amperos 52. Net current (Step 50 - Step 22) = 3.07. E-9 Appares 53. Efficiency (Step 52/299) = 1.01 E-11 A/R/hr * * Acceptance Limits = 0.8 to 1.2 x E-11 WR/hr ve litzano Dato DAMY 55 ON SFride Data 5-16-85 Figure 2 - Factory Calibration Data Sheet For Detector Serial Number 22711

• • · · · · . • .

· · · .

* . .

5 4 ga . . , 1

· • • • • · • • • • • • • •

· ·

. بر 9 . •

21

.

e 16 ay 16 2.J.JOA.RE-150 151593.003 AREA HONITOR CALIBRATION: DATA SHEET Je 11520-4 Pala Voirie Unit 2, ch. 1.5-150 Detector Type Jon Chamber lligh Voltage Hfgr. H/P Hfgr. LND Hodel 6516A Hodel 50314 8/2 22704 S/H 1824A04064 KIC Part No. 824636-002 Cal. Due Functional " Readout Type Piacemmater Check Source/Live Zero -. SAUTES Savae: Huclide Am Activity .084 C. Hfgr. Keishley Hodel 616 Date 1482 S/N. 62723 A Cal. Due 11/85 Calibration Date 6/9/83 Test Location Provesse House Temperature 25'C Pressure 607.5%-14 Relative Humidity . < so to Calibration Source 1.D. Therefrom - SUM Huclide 60Cv Activity 6502G Date 12/77 Beam Heasuring Detector: Hfgr. Franklin Hodel . 6 no 5/H 1D-1000 For Verification of radiation field---HBS traceable) Type lon Ch. Certificate See addendum pg TEST DATA High Voltage Setting + Pour 4 Acceptance Criteria , 1' to 1.2 (10") Other Instrument Settings (if applicable) A/R/hr Check Source/Live Zero-Reading 1.0 - - 11 A KEEP ALIVE CURRENT Heasured Field Instrument Reading GROSS A/R/HR ЯEТ . 706 (10³)mR/hr 1,7 8-11 .7 8-11 199 5-11 .566 (104)mR/hr 6.55-11 5.5ELI +97 5-4 , 584 (10⁵)mR/hr 6.20 5-10 6-10 5-10 1.04 5-11 .754 (10⁶)mR/hr 7.84 5-9 7.53 4.4 1.04 5-11 DUM Fluka Mod. 8020 M SIN LV YF. J, Col day 1012 4/23, at +800 V, 2/8/83 Tarted with

631 Date 6/4/23 Performer C **A**D Oato 🚣

Figure 3 - Factory Calibration Data Sheet for Detector Serial Number 22709

1

· · · ·

		, interv		18-104
aman insiri	menipilan Corporation	STANDARD PRACT	ICE PROCEDURE	. REV.
			21)-149	1.200 120/14 1 4
10.0 DAT	A SHEET - XOI-1000 PRIMA	RY CALIERATION, (Ls-137	÷
Calibrati	on Date 10-25-05	, ° 'Custa	cmer/Channel	2.4
KI Detecț	or Part No. 804636 - pong	Mfgr	. Serial No. <u>s</u>	2364
1		τ 1υ ' =ù		,
ITEM	MEGR	- MODEL	<u>5/N</u>	CAL DUE
Picoammet	er <i>Faithley</i>	GIE	12057	6/10/8-2
HV Supply	Can berra	3002	10466	10/20/86
DYM		· -		×
No 43.4			:	
Keep-Aliv	e Current (Acceptance Li	imits = 0.9 to 5.	5 x E-11A):	
. 10.	Sincl Dealing	· (ε-9 Co	ulonos)	
20.	Final Reading <u>1.40</u>	(E-9 Lo	uionins)	λ
21.	$\frac{1}{64.3}$	Second	(ou seconds min	i i mum ;
44.	current - (step 20 - 5)	.ep 16//3(ep 21 -	_ <u></u> AII	iperes
3.04 R/hr	readings:		•	
28.	Initial Reading	(E-9) C	oulombs	
30.	Final Reading 4.07) (e-3)	oulombs	
31.	Time Interval	Seconds	(60 seconds mi	ninum)
33.	Current = (Step 30 - St		4/5	iperes
34.	Net current (Step 33 -	Step 22) =	Amperes	>
35.	Efficiency (Step 34/3.0	14)) = <u>ورجر</u>	2_ A/R/hr *	
30.1 R/hr	readings:			
.41.	Gross current = <u></u>	E-10 Am	iperes	, I
43.	Net current (Step 41 -	Step 22) = 2.5	<u>a</u> E-10 Ampe	res
44.1	Efficiency (Step 43/30.	.!) = <u></u>	A/R/hr *	•
i.	,	* •		
299 R/hr	readings:			
50.	Gross current =	<u>E-9</u> Amp	961.62 De 1.62	
52.	Net current (Step 50 -	ی بر - (Step 22 -	C-9 Amper	re s
53.	Lificiency (Step 52/29	*) * <u></u>	A/R/hr •	
* Accepta	nce Limits = 0.8 to 1.7	x c-11 h/E/hr		*.
By Lings	J. Doute Date 10/25/	, æ <u>s</u> QA		Date arest
Figure	4 - Factory Calibration	Data Sheet for [Detector Serial I	Number 53364
,	. ,			

• •

.

10.6 DATA	SHEET - 16 (14)	Doo fripaav ci	U-3 RU-	- 148 5-137	3478	
Calibration	n llate 5-3	0 24	Custe	mar/Charnel	NIA	-
XI Detector	Part No. 8	<u></u>	. Mfgr.	Serial No.	1-2165	••
¢-	•• .	······································	F T	,		
<u>[][]</u>	MF	GR	HODEL	<u>578</u>	6.46 50:5	•
Piccannete	r Ke	rruces,	616	12:59	6-1 56	
нү Sabbja	C.A.	NBERKA	3002	11468	13-20-80	2
GVM	FZ FZ	UKE.	BOZOA	LV 302	2-9-87	7
Keep-Ålive	Current (Acce	ptance Limits	; = 0.9 to 5.9	5 x E-11A}:		
-18.	Initial Readin	g10	(E-9 Co	ulomos) .	•	
- 20.	Final Reading	1.53	(E-9 Co:	ulombs)		
21.	Time Interval		Second	(60 seconds min	imut:	
12.	Current = (Ste	c 20 - Step i	18)/Step 21 =	.0238E-9 An	iperes 🔨	
,	-					
3.04 R/hr	readings:			•	N/ ta	
23.	Initial Readin	g. <u>(. O</u>	(E-S) C	oulombs	<u> / 6 8</u>	
30.	Final Reading	3.89	(E-9) C	ouiomos		2
31.	Time Interval	60	Seconds	(60 seconds m	inimum) 💦 🤔	
· 33.	Current = (Ste	ep 30 - Step a	28)/Step 31 =	10482E-9 AC	nperes	2.192.19 2.192.19
34.	Net current (S	itep 33 - Ste	p 22) =_ <u>,024</u>	<u>це-ч</u> Aлiperes	•	
35.	Efficiency (St	ep 34/3.04))	= <u>.802E</u> -	<u>//</u> A/R/hr *		
*		4				÷
30.1 F/hr	reading::			·		
41:	Gross current	=	E-10 Ar	iperes		
43.	het current (S	tep 41 - Ste	p 22; = <u>2.7</u>	<u>د ک</u> E-10 Amp	cres	
44.	Efficiency (S	tep 43/30.1}	* <u>.918 E-11</u>	A/R/hr *		
299 R/hr r	eadings:					
50.	Gross current	= <u>Z.80</u>	E-9 Ams	eres		
· 52.	Net.current (S	Step 50 - Ste	p 22) = <u>2.78</u>	8 E-9 Ampe	res	
53,	Efficiency (Si	tep 52/299) =	<u>.92895-1</u>	/ A/R/hr *	•	
* Acceptar	ce Limits = 0.	.2 to 1.2 x E	-11 4/8/86	•		
61	•			(1)		
- YASL 1			/ 0*		Data the second	,

Page 7 of 14

, **,**

-	Kampa Instrumentation Corporation STANDARD PRACTICE PROCEDURE
	10.0 DATA SHEET - NOI-1600 PRIMARY CALICEATION, Cs-137
	Calibration Cate <u>5-30-86</u> Customer/Channel <u>NA</u> KI Detector Part No. <u>824636-003</u> Mfgr. Serial No. <u>62163</u>
•	ITEM MEGR MODEL S/N CAL DUE
	Picoarmeter KEITHLEY 616 12059 6-1-86
	IV SUPPLY CANBERRA 3002 LV468 10-20-86 DVM FLUKE 8020A LV302 2-9-87
,	Keep-Alive Current (Acceptance Limits = 0.9 to 5.5 x E-11A): 18. Initial Reading
	33. Current = (Step 30 - Step 28)/Step 31 = <u>.0468 E-9</u> Amperes 34. Net current (Step 33 - Step 22) <u>0268 E-9</u> Amperes 35. Efficiency (Step 34/3.04)) = <u>.0027 E-044</u> A/R/hr *
	18816E-11
	41. Gross current = <u>3.10</u> E-10 Amperes 43. Net current (Step 41 - Step 22) = <u>2.90</u> E-10 Amperes
* .	44. Efficiency (Step 43/30.1) = <u>.963 E-11</u> A/R/hr *
•	<pre>44. Efficiency (Step 43/30.1) = <u>.963 E-11</u> A/R/hr * . 299 R/hr readings: 50. Gross current = <u>.2.92</u> E-9 Amperes 52. Net current (Step 50 - Step 22) = <u>.2.92</u> E-9 Amperes 53. Efficiency (Step 52/299) = <u>969 E-11</u> A/R/hr *</pre>
	44. Efficiency (Step 43/30.1) = <u>.963 E-11</u> A/R/hr * 299 R/hr readings: 50. Gross current = <u>2.92</u> E-9 Amperes 52. Net current (Step 50' - Step 22) = <u>2.90</u> E-9 Amperes 53. Efficiency (Step 52/299) = <u>.969 E-11</u> A/R/hr * * Acceptance Limits = 0.8 to 1.2 x E-11 A/R/hr By Apple Date <u>5-30-86</u> QA OC Date <u>XY 30 %</u>

. . .

•

•

. . . ≮, .

•

** **

•

)

ŀ

NUCLEAR ADMINISTRATIVE AND TECHNICAL MANUAL

Page 50 of 58

RADIATION TEST FOR	MONITORING CAN	LIBRATION A MONITORS	;	74ST-9SO23	Revision 0.00
<u>148 149</u>				·,	،
	8.4.9 A	s Leit Keep A	live		•
fik fik	8.4.9.1	Using SIMS calibration.	S, verify that the detecto	er has not been replace	d since the last
<u>su sut</u>	8.4.9.	1.1 If t nu cu se to	the detector has been r umber listed in step 8.4 prent detector. (If the shal number from a DD replace the detector.)	eplaced, verify that the .9.4 matches the seria detector is inaccessi C file copy of the work	e detector serial I number of the ble, obtain the package used
511 GU	8.4:9.2	At the KEL	IC, record the radiation	n level.	
•		ALLOWABI	LE ≥1.00 E03 mr/hr	e	
		148/ <u> </u>	3E03	•	
	N A	149/ <u>i.D</u>	2 E 0 3		
<u> 402 94</u>	8.4.9.3	At the KEL	JC, place the LOCAL/R	EMOTE keyswitch in L	OCAL ·
<u>fu 44</u>	. 8.4.9.4	At the KEI function ke for the det	LIC, place the monitor rys *FTN 1 06 1 ENT*. F ector being checked.	r in the Calibrate Mod Record the Ion Chamb Mark the other unit de	de by pressing per current level electors *N/A*.

RU-148 DETECTOR

	DETECTOR SERIAL NUMBER	ALLOWABLE RANGE (IN AMPS)	AS LEFT VALUE
UNIT 1	22706	9.66E-12 TO 1.79E-11	. 1.07 E-11
UNIT 2	22709	7.00E-12 TO 1.30E-11	NA
UNIT 3	62165	1.67E-11 TO 3.09E-11	NA
Ac	ceptance Criteria (Rel. 2.2.)	6.3)	

Figure 7 - Calibration Data for RU-148, Unit 1 (WO # 00591469 Completed on 9/23/93)

• • • • ×. . . .

•

.

•

n ₹ ¹ • • • • •

NUCLEAR ADMINISTRATIVE AND TECHNICAL MANUAL

Page 51 of 58

RADIATION MONITORING CAUBRATION		74ST-9SO23	Revision 0.00
TEAL I OU NEW SOOPE AREA MONITORS	*		

148 149

a a			i	HU-149 DETECTOR		
		DETE SERIAL	CTOR NUMBER	ALLOWABLE RANGE (IN AMPS)	AS LEFT VALUE	
	UNIT 1	22	711	8.12E-12 TO 1.51E-11	1.04 E-11	
	UNIT 2	53	364	1.40E-11 TO 2.60E-11	NA	
	UNIT 3	62163		1.40E-11 TO 2.6CE-11	N/A	
	Acc	eptance Crit	eria (Rel. 2.	2.6.3)		
<u>44 51</u>		8.4.9.5	At the KE function ke	LIC. place the monitor in the No eys "FTN 1 C6 0 ENT".	ormal mode by pressing	
<u> 44 64</u>		8.4.9.6	At the KEL	IC, turn the LOCAL/REMOTE keys	witch to REMOTE.	
	8.4.	10 Restora	ation of Mon	itor		
<u>fit 44</u>		8.4.10.1	Ensure the	e monitor alarms and indications a	ppear normal.	
51A 51A	,	8:4.10.2	Ensure the KERIC has been returned to normal and all alarms a indications appear normal.			
<u>44 44</u>		8.4.10.3	At the min monitor nu	icomputer terminal, place the mon umber then press function key UNI	itor On Line by typing the T ON.	
·		A4 *	RU-148 m RU-149 m	onitor number 42. onitor number 50.		
<u>54 N/A</u>		8.4.10.4	For RU-14	8, proceed to section 8.5, RU-148	Functional Test.	
N/A GLA	•	8.4.10.5	For RU-14	9, proceed to section 8.6, RU-149	Functional Test.	

Figure 8 - Calibration Data for RU-149, Unit 1 (WO # 00591469 Completed on 9/23/93)

, _ _ _ /1

۰. ۲

- - -- -· · .

• • • , ,

ł

NUCLEAR ADMINISTRATIVE AND TECHNICAL MANUAL

Page 50 of 58

RADIATION I TEST FOR N	MONITORING CA	LIBRATION A MONITORS	74ST-9SQ23	Revision 0.00
		·		κ
<u>148 149</u>		1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 - 1997 -	•	
	8.4.9 A	s Left Keep Alive	,	
Jus Jus	8.4.9.1	Using SIMS, verify that the detect calibration.	or has not been replaced	l since the last
Jos quint	8.4.9.	1.1 If the detector has been number listed in step 8.4 current _detector (If _the serial number from a DD to replace the detector.)	replaced, verify that the 1.9.4 matches the serial detector is inaccessib of file copy of the work	detector serial number of the le, .obtain the package used
Jazal	8.4.9.2	At the KELIC, record the radiatio	n level.	'
VV	· ,	ALLOWABLE ≥1.00 E03 mr/hr 148/		
qui sui	8.4.9.3	149/ <u>1.00 E 0 3</u> At the KELIC, place the LOCAL/	REMOTE keyswitch in L	OCAL
for thes	8.4.9 <u>.</u> 4	At the KELIC, place the monitor function keys 'FTN 1 06 1 ENT'. for the detector being checked.	r in the Calibrate Mod Record the Ion Chambe Mark the other unit det	e by pressing er current level ectors 'N/A'.
•		RU-148 DETECTOR		

1	DETECTOR " SERIAL NUMBER	ALLOWABLE RANGE (IN AMPS)	AS LEFT VALUE
UNIT 1	22706	9.66E-12 TO 1.79E-11	N/A
UNIT 2	22709	7.00E-12 TO 1.30E-11	8.15E-12
UNIT 3	62165	1.67E-11 TO 3.09E-11	r/4
• Ac	centance Criteria (Rel. 2.2.	6.3)	

Figure 9 - Calibration Data for RU-148, Unit 2 (WO # 00650676 Completed on 2/24/94)

· · · · n

× ب •

٢

.

• **x** ¹

NUCLEAR ADMINISTRATIVE AND TECHNICAL MANUAL

Page 51 of 58

· · · · · · · · · · · · · · · · · · ·	*	

<u>148 149</u>

				RU-149 DETECTOR	
		DETE SERIAL		ALLOWABLE RANGE (IN AMPS)	AS LEFT VALUE
	UNIT 1	22	711	8.12E-12 TO 1.51E-11	NA
	UNIT 2	53:	364	1.40E-11 TO 2.60E-11	2.16E-11
	UNIT 3	62	163	1.40E-11 TO 2.60E-11	N/W
,	Acc	eptance Crit	eria (Ref. 2.)	2.6.3)	
the free		8.4.9.5	At the KE	LIC, place the monitor in the No eys *FTN 1 06 0 ENT*	rmal mode by pressing
fri fur		8.4.9.6	At the KEL	IC, turn the LOCAL/REMOTE keysv	witch to REMOTE.
V. V.,	8.4.	10 Restora	tion of Moni	itor	
and from	/	8.4.10.1	Ensure the	e monitor alarms and indications ap	pear normal.
Si gu	1	8.4.10.2	Ensure the indications	e KERIC has been returned to no	rmal and all alarms and
Dry An	1	8.4.10.3	At the mini monitor nu	icomputer terminal, place the moni Imber then press function key UNIT	tor On Line by typing the ON.
		<i>:</i> •	RU-148 mo RU-149 mo	onitor number 42. onitor number 50.	
AIN NIA	ł	8.4.10.4	For RU-14	8, proceed to section 8.5, RU-148 F	Functional Test.
N/A Giv		8.4.10.5	For RU-14	9, proceed to section 8.6, RU-149 F	Functional Test.

Figure 10 - Calibration Data for RU-149, Unit 2 (WO # 00650676 Completed on 2/24/94)

* • • • •

x

, ,

, **.**

. `

,

NUCLEAR ADMINISTRATIVE AND TECHNICAL MANUAL

Page 50 of 58

RADIATION MONITORING CALIBRATION TEST FOR NEW SCOPE AREA MONITORS			74ST-9SQ23	Revision 0.00	
<u>148 149</u>	£	······	- <u>, , , , , , , , , , , , , , , , , , ,</u>		
* \$.	8.4.9 A	s Left Keep	Alive	•	>
<u>D IIr</u>	8.4.9.1 Using SIMS, verify that the detector has not been replaced since the la calibration.				ed since the last
Q III	8.4.9.	1.1	If the detector has be number listed in step current detector. (If	een replaced, verify that the p 8.4.9.4 matches the seria the detector is inaccessi	e detector serial I number of the ble, obtain the
<u>D 564</u>	8.4.9.2	t At the KE	ELIC, record the radi	iation level.	
· •	•	ALLOWA 148/ 149/	BLE 21.00 E03 mr/r 2 8-19-92 1.00 - 12 E 05 1-5 1-01 E03	0 603 144 9.19.12 0 EO3	
<u> 414</u>	8.4.9.3	At the KE	ELIC, place the LOC	AUREMOTE keyswitch in I	
<u>12 - 514</u>	8.4.9.4	At the Ki function I for the de	ELIC, place the mo keys 'FTN 1 06 1 EN elector being check	onitor in the Calibrate Moo NT. Record the Ion Chamb ed. Mark the other unit de	de by pressing per current level etectors 'N/A'.
-			RU-148 DETECTO	R	,

۰ ب	DETECTOR	ALLOWABLE RANGE (IN AMPS)	AS LEFT VALUE
UNIT 1	22706	9.66E-12 TO 1.79E-11	NA
UNIT 2	22709	7.00E-12 TO 1.30E-11	NA
UNIT 3	62165	1.67E-11 TO 3.09E-11	3.07E-11
AC	ceptance Chiena (Hel. 2.2.	6.3)	-

Figure 11 - Calibration Data for RU-148, Unit 3 (WO # 00565171 Completed on 8/21/92)

پر ۲

٠

٠

, , •

•

• • •

,

NUCLEAR ADMINISTRATIVE AND TECHNICAL MANUAL

RADIATION MONITORING CALIBRATION	74ST-9SO23	Revision 0.00
	<	

<u>148 149</u>

RU-149 DETECTOR

		DETECTOR SERIAL NUMBER		ALLOWABLE RANGE (IN AMPS)	AS`LEFT VALUE		
	UNIT 1	22711		8.12E-12 TO 1.51E-11	NA		
	UNIT 2	53364		1.40E-11 TO 2.60E-11	NA		
	UNIT 3	62163		1.40E-11 TO 2.60E-11	1.72 E-V		
	Acceptance Criteria (Ref. 2.2.6.3)						
<u>\$14</u>	• =	8.4.9.5	At the KE	LIC, place the monitor in the Nor sys FTN 1 06 0 ENT*.	rmal mode by pressing		
R. SIA		8.4.9.6	At the KELIC, turn the LOCAL/REMOTE keyswitch to REMOTE.				
	8.4.10 Restoration of Monitor						
& GA		8.4.10.1	Ensure the	monitor alarms and indications ap	pear normal.		
£2. 51A		8.4.10.2	Ensure the KERIC has been returned to normal and all alarms and indications appear normal.				
<u>L</u> At	ĸ	8.4.10.3	At the minicomputer terminal, place the monitor On Line by typing the monitor number then press function key UNIT ON.				
	•	; **	RU-148 mc RU-149 mc	onitor number 42. onitor number 50.			
<u>IR N/A</u>	e.	8.4.10.4	For RU-14	8, proceed to section 8.5, RU-148 F	Functional Test.		
N/A 5/1		8.4.10.5	For RU-14	9, proceed to section 8.6, RU-149 F	Functional Test.		
					*		

•Figure 12 - Calibration Data for RU-149, Unit 3 (WO # 00565171 Completed on 8/21/92)

.

•

• • . .

. `

.

.

. . .