

TABLE 2

U S NUCLEAR REGULATORY COMMISSION
 OFFICE OF INSPECTION AND ENFORCEMENT
 CONFIRMATORY MEASUREMENTS PROGRAM
 FACILITY: D. C. COOK
 FOR THE 4 QUARTER OF 1985

SAMPLE	ISOTOPE	-----NRC-----		----LICENSEE----		---LICENSEE:NRC----		
		RESULT	ERROR	RESULT	ERROR	RATIO	RES	T
L WASTE	GROSS-B	8.4E-04	3.0E-05	1.2E-03	0.0E-01	1.4E 00	2.8E 01	D
	H-3	5.6E-03	1.5E-04	3.8E-03	0.0E-01	6.7E-01	3.7E 01	D
	FE-55	2.9E-04	8.0E-06	2.0E-04	0.0E-01	7.1E-01	3.6E 01	D
	SR-89	4.5E-07	1.4E-07	3.5E-09	0.0E-01	7.8E-03	3.2E 00	N
	SR-90	2.3E-07	9.0E-08	8.1E-08	0.0E-01	3.5E-01	2.6E 00	N

T TEST RESULTS:

A=AGREEMENT

D=DISAGREEMENT

*=CRITERIA RELAXED

N=COMPARISON

B604270021 860423
 PDR ADOCK 05000315
 G PDR