

Quality Cardiovascular Care LLC

1389 W Main St

Tower 1, Ste 106

Waterbury CT 06708

November 14, 2017

*Br. 1*

USNRC Region I

*03036684*

2100 Renaissance Blvd Ste 100

King of Prussia PA 19406

Greetings:

Please terminate our radioactive materials license number 06-30966-01.

Attached please find proof of disposition of radioactive materials, closeout area survey and closeout wipe test results.

Sincerely,

*Mark Rosenberg*

Mark Rosenberg, M.D.

REC'D IN LAT 11-14-17

*via fax*

*601739*  
NRC/RSN1 MATERIALS-002



Receiving  
 1800 Facility  
 1800 N. Keystone St.  
 Burbank, CA 91504  
 US

Telephone.....: (661) 309-1010  
 Fax.....: (661) 257-8303  
 Giro.....:  
 Tax exempt number.....: SRAC11-638175  
 Recipient.....:

Shipped From:  
 Quality Cardiovascular Care  
 Mark Rosenberg MD  
 PH (203)591-1998  
 mrosenberg@qccheart.com  
 1389 West Main St Tower One Suite  
 106  
 Waterbury, CT 06787

## Return acknowledgement

Return Number.....: 311534  
 RMA number.....: RMA\_030737  
 Date.....: 11/9/2017  
 Page.....: 1 of 1  
 Customer account.....: PINTEC01  
 Customer reference.....:  
 Contact.....: Wendy Koehnen  
 Customer requisition.....:

Dear Valued Customer,  
 Eckert & Ziegler Isotope Products Laboratories has received your radioactive source(s) and takes responsibility for tracking and storage of the source(s) listed below. If you have any further questions about the source(s), please contact EZIP and reference the return number listed above. Thank you for your business.

Item number	Batch number	Serial number	Quantity returned	Received on	Description
SRV-057-5M		1756-22-16	1.000000	11/8/2017	Vial
	Nuclide Info: Co-57	10.000000 mCi		Ref. Date: 10/1/2015	
RV-137-250U		1093-82-9	1.000000	11/8/2017	Vial
	Nuclide Info: Cs-137	250.000000 uCi		Ref. Date: 4/1/2005	



Receiving  
1800 Facility  
1800 N. Keystone St.  
Burbank, CA 91504  
US

Telephone.....: (661) 308-1010  
Fax.....: (661) 257-8303  
Giro.....  
Tax exempt number.....: SRAC11-638175  
Recipient.....:

Shipped From:  
Quality Cardiovascular Care  
Mark Rosenberg MD  
PH (203)591-1898  
mrosenberg@qccheart.com  
1389 West Main St Tower One Suite  
106  
Waterbury, CT 06787

### Return acknowledgement

Return Number: .....: 311534  
RMA number.....: RMA\_030736  
Date .....: 11/9/2017  
Page .....: 1 of 1  
Customer account.....: PINTEC01  
Customer reference.....:  
Contact .....: Wendy Koehnen  
Customer requisition .....

Dear Valued Customer,  
Eckert & Ziegler Isotope Products Laboratories has received your radioactive source(s) and takes responsibility for tracking and storage of the source(s) listed below. If you have any further questions about the source(s), please contact EZIP and reference the return number listed above. Thank you for your business.

Item number	Batch number	Serial number	Quantity returned	Received on	Description
3744.AD.010M.N		1832-064	1.000000	11/8/2017	MED3744

Nuclide Info: Co-57      10.000000 mCi      Ref. Date: 10/1/2015



### AREA WIPE SURVEY

**Location:** Quality Cardiovascular Care  
1389 West Main St  
Waterbury CT Nuclear Medicine  
06798

**Test Date/Time:** 11/14/2017 16:21

**Next Due:** 11/21/17

**Instrument:** well counter  
Serial#: 171143  
Background: 266 Cpm

**Make:** capintec **Model:** crc-15w  
**ID:** Inst. check results: PASS  
**Wipe area sample size:** 100 cm<sup>2</sup>  
**Last calibration:** 10/21/17 13:51:58

Area	Result	Count	Rate	Std Dev	Limit
1 Hot Lab L-Shield	Not Restricted	262	29.09	55	2000
2 Injection Area	Not Restricted	262	29.09	55	2000
3 Tread Mill	Not Restricted	262	29.09	55	2000
4 Camera	Not Restricted	262	29.09	55	2000
5 Processing Station	Not Restricted	262	29.09	55	2000
6 Bathroom	Not Restricted	262	29.09	55	2000

**Comments and Recommendations:**

**Performed by:** JB

**Radiation safety officer:**



### AREA METER SURVEY

**Location:** Quality Cardiovascular Care  
 1389 West Main St  
 Waterbury CT 06798  
 Nuclear Medicine

**Test Date/Time:** 11/14/2017 16:29  
**Next Due:** 11/15/17

**Instrument:** Ludlum-729      **Make:** Ludlum      **Model:** 14c  
**Serial#:** 210729      **ID:** 1  
**Background:** 0.03 mR/Hr      **Instrument check results:** 0.55mR/hr  
**Test initials:** JB      **Last calibration:** 11/11/16

Area	Restriction	Background	Survey	Level	Count/Min
1 Hot Lab L-Shield	Restricted	0.03	0		2
2 Injection Area	Non-Restricted	0.03	0		0.5
3 Tread Mill	Non-Restricted	0.03	0		0.5
4 Camera	Non-Restricted	0.03	0		0.5
5 Processing Station	Non-Restricted	0.03	0		0.5
6 Bathroom	Non-Restricted	0.03	0		0.5

**Comments and Recommendations:**

**Performed by:** Stuart R. Korchin, P.E., FACR

**Radiation safety officer:**



ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE

Name and Address of Applicant and/or Licensee

Quality Cardiovascular Care, LLC  
Attn: Mark Rosenberg M.D.  
1389 West Main Street  
Tower One, Suite 106  
Waterbury, CT 06708

Date

11/21/2017

License Number(s)

06-30966-01

Mail Control Number(s)

601739

Licensing and/or Technical Reviewer or Branch

Medical Branch

This is to acknowledge receipt of your:  Letter and/or  Application Dated: 11/14/2017

The initial processing, which included an administrative review, has been performed.

Amendment  Termination  New License  Renewal

There were no administrative omissions identified during our initial review.

This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

Your application for a new NRC license did not include your taxpayer identification number. Please complete and submit NRC Form 531, Request for Taxpayer Identification Number, located at the following link: <http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>  
Follow the instructions on the form for submission.

The following administrative omissions have been identified:

[Empty box for administrative omissions]

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region I  
U. S. Nuclear Regulatory Commission  
Division of Nuclear Materials Safety  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406-2713  
(610) 337-5260, (610) 337-5313,  
(610) 337-5398, or (610) 337-5239