

Dykert, Jason

From: Brandon Smid <brandons@helmsengineering.com>
Sent: Wednesday, November 22, 2017 10:06 AM
To: Dykert, Jason
Subject: [External_Sender] RE: NRC inspection report 2017-001

Mr. Dykert,

The two Troxler 3430 gauges, serial numbers 61152 and 27243, that were not leak tested have been put into our secure storage facility and have been tagged out until a leak test can be performed. This leak test is set for December 14, 2017. The two gauges will not be used or transported until after the leak test results have been obtained. If you have any further questions, please let me know. Thank you for your assistance in this matter.

Brandon D. Smid, P.E.



221 Brown County Highway 19
PO Box 111
Aberdeen, South Dakota 57401
Phone: 605.225.1212
Cell: 605.216.8097
Fax: 605.225.3189
Email: brandons@helmsengineering.com

From: Dykert, Jason [mailto:Jason.Dykert@nrc.gov]
Sent: Wednesday, November 01, 2017 2:22 PM
To: Brandon Smid <brandons@helmsengineering.com>
Subject: NRC inspection report 2017-001

Mr. Smid,

I have attached a copy of the inspection report you will be receiving in the mail shortly. You can find it in our web based database ADAMS under reference number ML17279B127 as well.

Please let me know if you have any questions,

Jason C Dykert
Health Physicist, Inspection Branch
US NRC, Region IV
(817) 200 - 1427

