

STAN A. HUBER CONSULTANTS, INC.

200 NORTH CEDAR ROAD – NEW LENOX IL 60451-1751 – PHONE 815-485-6161

SEALED GAMMA/BETA SOURCES - LEAK TEST CERTIFICATE

REPORT DATE: 6/21/2017
 FACILITY: Cabell Huntington Hospital
 Huntington WV
 NRC LICENSE NUMBER:
 STATE LICENSE NUMBER: 47-00404-02
 SOURCE IDENTIFICATION:

47-00404-02
 03003370

RADIONUCLIDE: Am-241
 ACTIVITY: 0.0113 Ci
 CALIBR. DATE: 3/17/1978
 MANUFACTURER: Amersham
 MODEL NUMBER: Not Provided
 SERIAL NUMBER: Not Provided
 OTHER DESCRIPTION: Ward Cable

ASSAY RESULTS:

COUNTING EFFICIENCY: 0.645
 BACKGROUND AT TIME OF ASSAY (CPM): 48
 LLD (uCi): 0.00002

	MAX CPM	NET CPM	DPM	ACTIVITY uCi
WET WIPE	58	10	<LLD	<LLD
DRY WIPE	51	3	<LLD	<LLD

ANALYSIS PERFORMED USING:

Packard Auto-Gamma Scintillation Spectrometer Model #D5003 Serial #406282 AND/OR
 Packard 1900CA Tri-Carb Liquid Scintillation Analyzer Model #A1900 Serial #101464

ANALYSIS OF RESULTS.

Sources are not leaking at this time
 Removable contamination is less than 5E-3 uCi. (0.005)

NEXT LEAK TEST DUE BY: December 13, 2017

LEAK TEST PERFORMED BY: Tina Shoemaker DATE: 6/13/2017

ANALYSIS PERFORMED BY: Richard Devore  DATE: 6/21/2017

RADIATION SAFETY OFFICER SIGNATURE: _____

601765
 NMSS/RGN1 MATERIALS-002