

NOV 18 1986

Docket Nos. 50-220
50-410

Niagara Mohawk Power Corporation
ATTN: Mr. C. V. Mangan
Senior Vice President
300 Erie Boulevard West
Syracuse, New York 13202

Gentlemen:

Subject: Combined Inspection Report Nos. 50-220/86-20; 50-410/86-55

This refers to the routine safety inspection conducted by Mr. H. Zibulsky of this office on October 7-9, 1986 of activities authorized by NRC License Nos. DPR-63 and CPPR-112 and to the discussions of our findings held by Mr. Zibulsky with Mr. J. Aldrich and other members of your staff at the conclusion of the inspection.

Areas examined during this inspection are described in the NRC Region I Inspection Report which is enclosed with this letter. Within these areas, the inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector.

Within the scope of this inspection, no violations were observed.

No reply to this letter is required. Your cooperation with us in this matter is appreciated.

Sincerely,

Original Signed By:

for *Thomas T. Martin*
Thomas T. Martin, Director
Division of Radiation Safety
and Safeguards

Enclosure: Combined NRC Region I Inspection Report Numbers 50-220/86-20 and 50-410/86-55

8611250312 861118
PDR ADOCK 05000220
G PDR

OFFICIAL RECORD COPY

CIR NMP 86-20/55 - 0001.0.0
11/12/86

11
JED



cc w/encl:

T. E. Lempges, Vice President, Nuclear Generation
 J. A. Perry, Vice President, Quality Assurance
 T. Perkins, General Superintendent, Nuclear Generation
 W. Hansen, Manager of Quality Assurance
 T. Roman, Station Superintendent
 J. Aldrich, Supervisor, Operations
 W. Drews, Technical Superintendent
 D. Quamme, Manager of Quality Assurance
 C. Beckham, NMPC QA Manager
 R. B. Abbott, Station Superintendent
 Troy B. Conner, Jr. Esquire
 John W. Keib, Esquire
 Director, Power Division
 Public Document Room (PDR)
 Local Public Document Room (LPDR)
 Nuclear Safety Information Center (NSIC)
 NRC Resident Inspector
 State of New York

bcc w/encl:

Region I Docket Room (with concurrences)
 Management Assistant, DRMA (w/o encl)
 DRP Section Chief
 L. J. Cunningham, IE
 Robert J. Bores, DRSS
 Region I SLO

RI:DRSS *[initials]*
 Zibulsky/rw
 11/13/86

RI:DRSS *[initials]*
 Pascjak
 11/17/86

RI:DRSS *[initials]*
 Bellamy
 11/17/86

RI:DRSS *[initials]*
 T. Martin
 11/17/86

OFFICIAL RECORD COPY

CIR NMP 86-20/55 - 0001.1.0
 11/12/86

10/10/10



U.S. NUCLEAR REGULATORY COMMISSION
REGION I

Report No. 50-220/86-20
50-410/86-55

Docket No. 50-220
50-410

License No. DPR-63
CPPR-112

Priority -

Category C
B

Licensee: Niagara Mohawk Power Corporation

300 Erie Boulevard West

Syracuse, New York 13202

Facility Name: Nine Mile Point, Units 1 and 2

Inspection At: Scriba, New York

Inspection Conducted: October 7-9, 1986

Inspectors:

Harvey Zibulsky
H. Zibulsky, Chemist

11-13-86
date

Approved by:

W. J. Pasciak
W. J. Pasciak, Chief, Effluents Radiation
Protection Section, DRSS

11-17-86
date

Inspection Summary: Inspection on October 7-9, 1986 (Combined Report Nos.
50-220/86-20; 50-410/86-55)

Areas Inspected: Routine, announced inspection of the nonradiological chemistry program. Areas reviewed included measurement control and analytical procedure evaluations.

Results: No violations were identified.





DETAILS

1.0 Individuals Contacted

- * J. Aldrich, Operations Superintendent, Unit 1
- * E. Leach, Superintendent, Chemistry and Radiation Management
- * J. Duell, Supervisor, Chemistry and Radiation Protection
- * J. Blasiak, Supervisor, Chemistry, Unit 1
- * A. Ross, Supervisor, Chemistry, Unit 2
- * J. Coates, Chief Technician A
- T. Roman, Station Superintendent, Unit 1
- J. Moser, Chemistry Technician, Unit 1

* denotes those present at the exit interview.

The inspector also interviewed other licensee employees including members of the chemistry staff.

2.0 Measurement Control Evaluation

This is a verification of the licensee's measurement capabilities on actual plant water samples.


Samples from the Effluent and Influent were taken and spiked with an anion standard and a metal standard. Each sample, with the spikes, were divided into two. The licensee will analyze one sample and the other was sent to Brookhaven National Laboratory (BNL) for independent verification. On completion of the analyses by both laboratories, a statistical evaluation will be made (Inspector Follow-up Item 50-220/86-20-01).

The inspector observed that the licensee used one standard stock solution for calibration purposes. A separate control solution is needed to provide an analytical cross check on the continuing quality of the stock solutions. The licensee agreed to maintain two standard stock solutions.

Some of the calibration curves were not statistically fit to the data points but were graphically approximated. This could produce as much as 15 percent error as may have been the case in the licensee's analysis of the NRC blind standard for chloride. The licensee agreed to use a statistical method to draw the calibration curves.

3.0 Analytical Procedures Evaluation

During the inspection, standard chemical solutions were submitted by the inspector to the licensee for analysis. The standard solutions were prepared by BNL for NRC Region 1, and were analyzed by the licensee using normal methods and equipment. The analysis of standards is used to verify the various plant systems with respect to Technical Specification and other regulatory requirements. In addition, the analysis of standards is



1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
66
67
68
69
70
71
72
73
74
75
76
77
78
79
80
81
82
83
84
85
86
87
88
89
90
91
92
93
94
95
96
97
98
99
100

1
2

3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
66
67
68
69
70
71
72
73
74
75
76
77
78
79
80
81
82
83
84
85
86
87
88
89
90
91
92
93
94
95
96
97
98
99
100

101
102
103
104
105
106
107
108
109
110
111
112
113
114
115
116
117
118
119
120
121
122
123
124
125
126
127
128
129
130
131
132
133
134
135
136
137
138
139
140
141
142
143
144
145
146
147
148
149
150
151
152
153
154
155
156
157
158
159
160
161
162
163
164
165
166
167
168
169
170
171
172
173
174
175
176
177
178
179
180
181
182
183
184
185
186
187
188
189
190
191
192
193
194
195
196
197
198
199
200

201
202
203
204
205
206
207
208
209
210
211
212
213
214
215
216
217
218
219
220
221
222
223
224
225
226
227
228
229
230
231
232
233
234
235
236
237
238
239
240
241
242
243
244
245
246
247
248
249
250
251
252
253
254
255
256
257
258
259
260
261
262
263
264
265
266
267
268
269
270
271
272
273
274
275
276
277
278
279
280
281
282
283
284
285
286
287
288
289
290
291
292
293
294
295
296
297
298
299
300

301
302
303
304
305
306
307
308
309
310
311
312
313
314
315
316
317
318
319
320
321
322
323
324
325
326
327
328
329
330
331
332
333
334
335
336
337
338
339
340
341
342
343
344
345
346
347
348
349
350
351
352
353
354
355
356
357
358
359
360
361
362
363
364
365
366
367
368
369
370
371
372
373
374
375
376
377
378
379
380
381
382
383
384
385
386
387
388
389
390
391
392
393
394
395
396
397
398
399
400

401
402
403
404
405
406
407
408
409
410
411
412
413
414
415
416
417
418
419
420
421
422
423
424
425
426
427
428
429
430
431
432
433
434
435
436
437
438
439
440
441
442
443
444
445
446
447
448
449
450
451
452
453
454
455
456
457
458
459
460
461
462
463
464
465
466
467
468
469
470
471
472
473
474
475
476
477
478
479
480
481
482
483
484
485
486
487
488
489
490
491
492
493
494
495
496
497
498
499
500



used to evaluate the licensee's analytical procedures with respect to accuracy and precision.

The results of the standard measurements comparison indicated that five out of thirty-three comparisons were in disagreement under the criteria used for comparing results (see Attachment 1). The results of the comparisons are listed in Table 1.

The NRC boron standard solutions were analyzed by the licensee but were not reported because the concentrations were not in the range that is usually determined.

The specific ion electrode chloride disagreement was due to the licensee graphically approximating the data points on the calibration curve instead of statistically fitting the curve. The spectrophotometric chloride disagreement was due to the insensitivity of the procedure at levels less than 50 ppb chloride.

The silica disagreement was due to the improper size cuvette that was used. A 1 cm cell was used instead of a 5 cm cell, as Unit #2 used. For any silica concentration less than 30 ppb, a larger cuvette size should be used to increase sensitivity.

The chromium disagreement seemed to be due to a sampling error.

The sodium disagreement and determination of the 3 NRC standards showed a positive bias. This may be due to contamination from the potassium nitrate that was added to the sample before analysis. The licensee is going to investigate the cause of this systematic bias.

The inspector advised the licensee that they should consider purchasing an ion chromatograph. This instrument will enable the licensee to analyze sulfate concentrations as low as 20 ppb, which is the EPRI Guideline parameter. The licensee's turbidimetric procedure is not sensitive enough to meet the EPRI requirement for sulfate.

4. Exit Interview

The inspector met with the licensee representatives (denoted in paragraph 1) at the conclusion of the inspection on October 9, 1986, and summarized the scope and findings of the inspection. At no time during this inspection was written material provided to the licensee by the inspector.



TABLE 1
CAPABILITY TEST RESULTS
NINE MILE POINT, UNITS 1 AND 2

<u>Chemical Parameter</u>	<u>NRC Value</u>	<u>Lic. Value</u>	<u>Ratio</u> ^(Lic) <u>(NRC)</u>	<u>Comparison</u>
Results in parts per billion (ppb)				
Fluoride (Sp. Ion Electrode) #2	23.1±0.5	22.3±0.6	0.97±0.03	Agreement
	43.5±1.9	41.7±0.6	0.96±0.04	Agreement
	80.5±2.2	80.3±0.6	1.0	Agreement
Chloride (Sp. Ion Electrode) #1	24.1±3.1	19±1	0.79±0.11	Disagreement
	37.4±1.2	36.7±0.6	0.98±0.04	Agreement
	80.5±2.2	81.0±1.7	1.0	Agreement
Chloride (Spectrophotometry) #1	24.1±3.1	37.0±1.7	1.53±0.21	Disagreement
	37.4±1.2	37.0±13.1	1.0	Agreement
	80.5±2.2	73.3±4.6	0.91±0.06	Agreement
Chloride (Spectrophotometry) #2	24.1±3.1	24.0±6.0	1.0	Agreement
	37.4±1.2	38.3±2.5	1.02±0.07	Agreement
	80.5±2.2	74.0±5.2	0.92±0.07	Agreement
Silica (Spectrophotometry) #1	10.9±1.1	12.0±0	1.10±0.11	Agreement
	21.8±1.4	18.0±0	0.83±0.05	Disagreement
	32.0±1.0	29.7±2.5	0.93±0.08	Agreement
Silica (Spectrophotometry) #2	10.9±1.1	12.0±0	1.10±0.11	Agreement
	21.8±1.4	23.3±1.5	1.07±0.10	Agreement
	32.0±1.0	32.0±2.0	1.0	Agreement
Results in parts per million (ppm)				
Iron (Atomic Absorption)	0.587±0.04	0.613±0.01	1.04±0.07	Agreement
	1.15±0.04	1.21±0.02	1.05±0.04	Agreement
	1.76±0.05	1.73±0.02	0.98±0.03	Agreement
Copper (Atomic Absorption)	0.562±0.03	0.606±0.01	1.08±0.06	Agreement
	1.16±0.06	1.18±0.01	1.02±0.05	Agreement
	1.74±0.07	1.76±0.01	1.01±0.04	Agreement

100



Table 1

<u>Chemical Parameter</u>	<u>NRC Value</u>	<u>Lic. Value</u>	<u>Ratio (NRC)</u>	<u>Comparison</u>
Nickel (Atomic Absorption)	0.611±0.03	0.621±0.01	1.02±0.05	Agreement
	1.22±0.04	1.20±0	0.98±0.03	Agreement
	1.84±0.05	1.73±0.03	0.94±0.03	Agreement
Chromium (Atomic Absorption)	0.612±0.04	0.638±0.04	1.04±0.09	Agreement
	1.13±0.04	1.29±0.04	1.14±0.05	Disagreement
	1.72±0.10	1.82±0.01	1.06±0.06	Agreement
Sodium (Atomic Absorption)	0.183±0.02	0.240±0	1.31±0.14	Disagreement
	0.369±0.03	0.431±0	1.17±0.10	Agreement
	0.576±0.03	0.625±0	1.09±0.06	Agreement



ATTACHMENT 1

CRITERIA FOR COMPARING ANALYTICAL MEASUREMENTS

This attachment provides criteria for comparing results of capability tests. In these criteria, the judgement limits are based on the uncertainty of the ratio of the licensee's value to the NRC value. The following steps are performed:

- (1) the ratio of the licensee's value to the NRC value is computed

$$\left(\text{ratio} = \frac{\text{Licensee Value}}{\text{NRC Value}}\right);$$

- (2) the uncertainty of the ratio is propagated.¹

If the absolute value of one minus the ratio is less than or equal to twice the ratio uncertainty, the results are in agreement.

$$(|1 - \text{ratio}| \leq 2 \text{ uncertainty})$$

$$^1 Z = \frac{x}{y}, \text{ then } \frac{S_z^2}{Z^2} + \frac{S_x^2}{x^2} + \frac{S_y^2}{y^2}$$

¹(From: Bevington, P. R., Data Reduction and Error Analysis for the Physical Sciences, McGraw-Hill, New York, 1969)

1990

