



EMERY SAPP & SONS, INC

O: 573-445-8331
F: 573-445-0266

2301 I-70 Drive NW
Columbia, MO 65202

WWW.EMERYSAPP.COM

February 13, 2018
From: Kyle Frye
2301 I-70 Drive NW
Columbia, Mo 65202
816-262-0170

To: Laura Cender
U.S NRC Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4352

Reference: Additional information to control number 602281

Laura:

I would like to make note that nuclear material is no longer being stored at 2600 N. Stadium Blvd. Columbia, Mo 65202. I have attached a copy of the most current leak tests reports.

Please feel free to contact me if you have any questions. Thank you.

Sincerely,

A handwritten signature in black ink that reads 'Kyle Frye'. The signature is written in a cursive, flowing style.

Kyle Frye
Quality Control Manager



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 876-9537 Fax: (866) 391-2759

License: NC 032-0182-1

KYLE FRYE
EMERY SAPP & SONS INC.
2301 NW I-70 DRIVE
COLUMBIA, MO 65202

Cust ID: 13132

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440P **Serial No:** 62372

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-9005	03/11/2009	Cs-137	0.296	8
78-5687	02/26/2009	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 09/15/2017
Sample analyzed on: 09/18/2017 12:52:51 P **Position:** 12
Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.24E+01	1.97E+01
Background measurement (cpm)	0	24
Sample measurement (cpm)	0	13
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	3.7E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

KYLE FRYE
EMERY SAPP & SONS INC.
2301 NW I-70 DRIVE
COLUMBIA, MO 65202

Cust ID: 13132

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 **Serial No:** 68788

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
78-9172	02/20/2014	Am-241:Be	1.48	40
77-13290	01/03/2014	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 09/13/2017
Sample analyzed on: 09/18/2017 12:54:04 P **Position:** 13
Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.24E+01	1.97E+01
Background measurement (cpm)	0	24
Sample measurement (cpm)	0	26
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	3.7E-01	1.3E+00

This certifies that the leak test results are:

✓ **Less than 185 Bq (0.005 uCi)**

Greater than 185 Bq (0.005 uCi)