



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION IV  
1600 E. LAMAR BLVD  
ARLINGTON TX 76011-4511

February 26, 2018

Mr. Thomas Greening  
Radiology Department  
Guam Regional Medical City  
Radiation Oncology  
133 Route 3  
Dededo, Guam 96929

SUBJECT: TERMINATION OF NRC RADIOACTIVE MATERIALS LICENSE NO. 56-35371-01

By letter dated January 26, 2018, Mr. Alan Funtanilla, Vice President, Deputy Chief Operating Officer, informed the U.S. Nuclear Regulatory Commission (NRC) that Guam Regional Medical City wished to terminate NRC radioactive materials license number 56-35371-01. The NRC Form 314 dated February 14, 2018, stated that all radioactive material had been returned to the manufacturer and/or distributor and the supporting documentation included a copy of the most recent leak test results. The leak test results show that there was no evidence of leakage of radioactive material from the sealed sources authorized under the NRC license.

Based on the documentation submitted and a review of your historical file, the staff concluded that all licensable radioactive material has been removed from your facility. Therefore, your facility located at 133 Route 3, Dededo, Guam, is suitable for unrestricted use in accordance with 10 CFR 20.1402, "Radiological Criteria for Unrestricted Use." Based on these conclusions, no further remediation or actions with respect to NRC regulated material is required.

Please find enclosed Amendment No. 2 to license number 56-35371-01 authorizing your license termination. An environmental assessment for this action is not required since this action is categorically excluded under 10 CFR 51.22(c)(20)(ii).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

You can contact me at 817-200-1189 if you have questions about this matter.

Sincerely,

/RA/

Roberto J. Torres, M.S., Senior Health Physicist  
Materials Licensing and Inspection Branch

Docket: 030-38990  
License: 56-35371-01  
Control: 602434

Enclosure: As stated