

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

<p>1. LICENSEE LOCATION INSPECTED</p> <p>Southgate Radiology 15300 Trenton Road Southgate, MI 48195</p> <p>REPORT NUMBER(S) 2017001</p>	<p>2. NRC REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2445 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>
-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

<p>3. DOCKET NUMBER(S)</p> <p>030-17665</p>	<p>4. LICENSE NUMBER(S)</p> <p>21-20030-01</p>	<p>5. DATE(S) OF INSPECTION</p> <p>December 18, 2017</p>
---------------------------------------------	------------------------------------------------	----------------------------------------------------------

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, monitored, corrected, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
 (Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	David Tehran	<i>David Tehran</i>	12/18
BRANCH CHIEF	Aaron McCraw	<i>[Signature]</i>	02/28/2018

Docket File Information
SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Southgate Radiology 15300 Trenton Road Southgate, MI 48195 REPORT NUMBER(S) 2017001	2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
-----------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------------

3. DOCKET NUMBER(S) 030-17665	4. LICENSE NUMBER(S) 21-20030-01	5. DATE(S) OF INSPECTION December 18, 2017
--------------------------------------	-----------------------------------------	---------------------------------------------------

6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS All
--------------------------------------------	--------------------------------------

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT Dennis Vollman, DO - RSO	4. TELEPHONE NUMBER (313) 281-6600
---------------------------------	----------------------	-----------------------------------------------------	-------------------------------------------

Main Office Inspection Next Inspection Date: 12/18/2022

Field Office Inspection _____

Temporary Job Site Inspection _____

PROGRAM SCOPE

This was an unannounced, routine inspection of a private medical practice authorized to use byproduct material for diagnostic medical purposes at its clinic in Southgate, Michigan. At the time of the inspection, the licensee employed one full-time technologist who administered general nuclear medicine studies to approximately 15 patients per month, Monday through Friday. The licensee received unit doses only from a local radiopharmacy, and retained the services of a consulting physicist who audited the radiation safety program semiannually.

PERFORMANCE OBSERVATIONS

This inspection consisted interviews with licensee personnel, a review of selected records, and a tour of the nuclear medicine department in Southgate. The inspector was unable to observe the administration of any radiopharmaceuticals. However, the staff was able to demonstrate the implementation of licensee procedures for receipt of packages containing licensed material and spill response. The inspector also reviewed a selection of records, including dose receipt, dose administration and contamination control protocols.

No violations of NRC requirements were identified as a result of this inspection.