

From: [Gaskins, Farrah](mailto:Gaskins.Farrah)
To: jnorweck@radiology-inc.com
Subject: Request for additional information for License No. 47-31291-01
Date: Thursday, March 01, 2018 9:12:00 AM

PLEASE CONFIRM RECEIPT OF THIS REQUEST FOR ADDITIONAL INFORMATION.

Licensee: Boone Memorial Hospital
License No.: 47-31291-01
Docket No.: 03037592
Control No.: 602389

Dear Mr. Norweck:

This is in reference to the application dated February 2, 2018, requesting to renew Nuclear Regulatory Commission License No. 47-31291-01. In order to continue our review, we need the following additional information:

1. Item No. 2 of your application lists licensee's name as Boone Memorial Hospital, however, Item No. 1 on your current license reads Boone Memorial Hospital, Inc. Please confirm the correct licensee name.
2. Although you, James Norweck, are currently listed on the license as the Radiation Safety Officer (RSO), since you are a consultant outside of the licensee's organization, please respond to the following:
 - a. Describe the control over the radiation safety program that will be delegated so that the consultant-RSO will be able to exercise authority over authorized users when confronted with radiation safety problems that require implementation of corrective actions.
 - b. Describe the relationship that will exist between the consultant-RSO and your institutional management regarding expenditure of funds to facilitate the objectives of your radiation safety program and related regulatory requirements.
 - c. Identify other commitments of the consultant-RSO for other NRC or Agreement State licensed facilities, along with a description of how the consultant-RSO will allocate time to permit the performance of the duties of the RSO as described in the regulations. State the consultant-RSO's minimum amount of on-site time (hours per week).
 - d. Appoint an in-house representative who will serve as the point of contact during the RSO's absence. This person may be allowed to assist the consultant RSO with limited authority.
 - e. Describe the overall availability of the consultant-RSO to respond to questions or operational issues that arise during the conduct of your radiation safety program and related regulatory requirements. Specify the maximum amount of time it will take the RSO to arrive at the facility in the event of an

emergency that requires his presence.

We will continue our review upon receipt of the requested information. You may respond to my attention in writing by letter, email (if letter is signed by senior management and scanned into a pdf format), or fax (610-337-5269), referencing mail control number 602389. If you have any questions regarding this deficiency letter, please call me at (610) 337-5143.

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

Sincerely,

Farrak C. Gaskins

Health Physicist, Medical and Licensing Assistance Branch
U.S. NRC, Region I
2100 Renaissance Blvd, Suite 100
King of Prussia, PA 19406
(610) 337-5143 office
(610) 337-5269 fax