



O: 573-445-8331 2301 I-70 Drive NW
F: 573-445-0266 Columbia, MO 65202

WWW.FMERYSAPP.COM

EMERY SAPP & SONS, INC

April 3, 2018
From: Kyle Frye
2301 I-70 Drive NW
Columbia, Mo 65202
816-262-0170

To: Materials Licensing Branch
U.S NRC Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4352

Please find the attached form NRC 314 for the termination of Swift Construction Company's license number 24-35005-01. Emery Sapp & Sons Inc. purchased Swift. The facility where the gauge was stored will continue to be a storage location and is now listed as a storage location on Emery Sapp & Sons Inc. license number 24-32751-01. I have also attached the most recent leak test showing that the gauge is acceptable for continuous use.

Please feel free to contact me if you have any questions. Thank you.

Sincerely,

Kyle Frye
Quality Control Manager

RECEIVED APR 03 2018

40 YEARS STRONG

CERTIFICATE OF DISPOSITION OF MATERIALS

PLEASE READ THESE INSTRUCTIONS BEFORE COMPLETING NRC FORM 314.

Subpart E of 10 CFR Part 20 establishes the radiological criteria for license terminations/decommissioning of facilities licensed under 10 CFR Parts 30, 40, 50, 60, 61, 70, and 72, as well as other facilities subject to the Commission's jurisdiction under the Atomic Energy Act of 1954, as amended, and the Energy Reorganization Act of 1974, as amended.

INSTRUCTIONS

Section B, Item 2.

Licensees should describe the specific radioactive material transfer actions. If radioactive wastes were generated in terminating this license, the licensee should describe the disposal actions taken, including the disposition of low-level radioactive waste, mixed waste, greater-than-Class-C waste, and sealed sources.

Section B, Item 2.a.

The information provided concerning the transfer of radioactive material to another licensee should specify the date of the transfer, the name of the licensee recipient, an individual contact name and telephone number for the licensee recipient, and the recipient's NRC or Agreement State license number.

Section B, Item 2.b.

For disposal of radioactive materials, licensees should describe the specific disposal method or procedure (e.g., decay-in-storage). For those cases when radioactive materials are disposed of by a licensed disposal site or by a waste contractor, the licensee should specify the name, address, and telephone number of the licensed disposal site operator or waste contractor.

Section B, Item 2.c.

"Residual radioactivity," as defined in 10 CFR 20.1003, means radioactivity in 'areas' (structures, materials, soils, etc.) remaining as a result of activities (licensed and unlicensed) under the licensee's control from sources used by the licensee, excluding background radiation. ALARA is defined in 10 CFR 20.1003.

FILE CERTIFICATES AS FOLLOWS:

IF YOU ARE LOCATED IN:

ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA, KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND CERTIFICATES TO:

LICENSING ASSISTANT SECTION
NUCLEAR MATERIALS SAFETY BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
2100 RENAISSANCE BOULEVARD, SUITE 100
KING OF PRUSSIA, PA 19406-2713



ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND CERTIFICATES TO:

MATERIALS LICENSING SECTION
U.S. NUCLEAR REGULATORY COMMISSION, REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, IL 60532-4352

IF YOU ARE LOCATED IN:

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MISSISSIPPI, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND CERTIFICATES TO:

MATERIAL RADIATION PROTECTION SECTION
U. S. NUCLEAR REGULATORY COMMISSION, REGION IV
1600 E. LAMAR BOULEVARD
ARLINGTON, TX 76011-4511

| | | | |
|--|--|---|---|
| NRC FORM 314 (02-2017) 10 CFR 30.35(j)(1); 40.42(j)(1); 70.38(j)(1); and 72.54(k)(5)(i)(1) |  <p>U.S. NUCLEAR REGULATORY COMMISSION</p> <p>CERTIFICATE OF DISPOSITION OF MATERIALS</p> | APPROVED BY OMB: NO. 3150-0029 EXPIRES: 02/29/2020 | Estimated burden per response to comply with this mandatory collection request 30 minutes. This submission is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-0 F33), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to Information.Resource@nrc.gov , and to the Desk Officer, Office of Information and Regulatory Affairs, NRC-10202, (3150-0029), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection. |
| LICENSEE NAME AND ADDRESS 1. Swift Construction Company 2. P.O. Box 788 Neosho, Mo 64805 | LICENSE NUMBER 24-35005-01 | DOCKET NUMBER 030-38574 | |
| | | LICENSE EXPIRATION DATE December 31, 2022 | |
| A. LICENSE STATUS (Check the appropriate box) | | | |
| <input type="checkbox"/> This license has expired. <input checked="" type="checkbox"/> This license has not yet expired; please terminate it. | | | |
| B. DISPOSAL OF RADIOACTIVE MATERIAL <i>(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)</i> | | | |
| The licensee, or any individual executing this certificate on behalf of the licensee, certifies that: | | | |
| <input type="checkbox"/> 1. No radioactive materials have ever been procured or possessed by the licensee under this license. | | | |
| <input checked="" type="checkbox"/> 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner. | | | |
| <input checked="" type="checkbox"/> a. Transfer of radioactive materials to the licensee listed below: Emery Sapp & Sons Inc. License Number: 24-32751-01 | | | |
| <input type="checkbox"/> b. Disposal of radioactive materials: | | | |
| <input type="checkbox"/> 1. Directly by the licensee: | | | |
| <input type="checkbox"/> 2. By licensed disposal site: | | | |
| <input type="checkbox"/> 3. By waste contractor: | | | |
| <input type="checkbox"/> c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA. | | | |
| C. SURVEYS PERFORMED AND REPORTED | | | |
| <input type="checkbox"/> 1. A radiation survey was conducted by the licensee. The survey confirms: | | | |
| <input type="checkbox"/> a. the absence of licensed radioactive materials | | | |
| <input type="checkbox"/> b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA. | | | |
| <input type="checkbox"/> 2. A copy of the radiation survey results: | | | |
| <input type="checkbox"/> a. is attached; or <input type="checkbox"/> b. is not attached (Provide explanation); or <input type="checkbox"/> c. was forwarded to NRC on: _____ Date | | | |
| <input checked="" type="checkbox"/> 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and | | | |
| <input checked="" type="checkbox"/> a. The results of the latest leak test are attached; and/or <input type="checkbox"/> b. No leaking sources have ever been identified. | | | |
| The person to be contacted regarding the information provided on this form: | | | |
| NAME Kyle Frye | TITLE Radiation Safety Officer | TELEPHONE (Include Area Code) (816) 262-0170 | E-MAIL ADDRESS kyle.frye@emerysapp.com |
| Mail all future correspondence regarding this license to: Emery Sapp and Sons, Inc. 2301 I-70 Drive NW Columbia, Mo 65202 | | | |
| C. CERTIFYING OFFICIAL I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT | | | |
| PRINTED NAME AND TITLE Kyle Frye Radiation Safety Officer | SIGNATURE  | DATE 03/29/2018 | |
| WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION. | | | |

"Specializing In Your Radiation Safety Needs"


R. M. WESTER and ASSOCIATES, INC

 215 Indacom Drive
 St. Peters, MO 63376
 (636) 928-9628
 www.rmwest.com

RADIOACTIVE SEALED SOURCE LEAK TEST REPORT

Test Date: December 20, 2017Analytical Date: December 20, 2017

Source Identification:

Manufacturer: TroxlerModel No.: 3450Mach S/N: 60Radionuclide: Cs-137Activity: 8.0 mCiSource S/N: 750-45130Radionuclide: Am-241/BeActivity: 40.0 mCiSource S/N: 47-24401Sample Submitted By: Aidyl Galafa-ColonFacility: Emery Sapp & Sons, Inc.Address: 2301 I-70 Drive NWColumbia, MO 65202

The identified sealed source listed above has been tested for leakage of radioactive materials as required by the United States Nuclear Regulatory Commission. The analysis of the wipe material used in testing the sealed source reveals the presence of $\leq 9.14 \times 10^{-5}$ μCi of loose contamination for Cs-137 and $\leq 9.01 \times 10^{-5}$ μCi of loose contamination for Am-241.

- (X) This source is acceptable for continued use.
- () This source has been found to have a level of loose contamination greater than 0.005 μCi of removable radioactive materials, and should be removed from service immediately.
- (N/A) Operational and performance check of shutter mechanism satisfactory.

Next Leak Test Date: June 20, 2018Analysis By: Aidyl Galafa-ColonReviewed By: [Signature]