



April 16, 2018

U. S. Nuclear Regulatory Commission
Materials Licensing Section
ATTN: Daniel Strohmeyer
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

Mr. Strohmeyer:

CardioSpecialists Group, Ltd. would like to provide the attached sealed source leak test form in response to your request for additional information.

If there are any questions concerning this license amendment, please contact Mr. Bryce Caudle at 317-443-9035 or by email at bcaudle@mpcphysics.com.

Sincerely,

A handwritten signature in black ink that reads "Ken Ramsey".

Dr. Kenneth Ramsey
President and Managing Partner
CardioSpecialists Group, Ltd.
2510 Gettler, Suite 420
Dyer, IN 46311

DYER
24 Joliet Street
Dyer, IN 46311
PH: 219 865 2141 (IN)
PH: 708 895 1650 (IL)

HAMMOND
5454 Hohman Avenue
Hammond, IN 46320
PH: 219 932 2300 (IN)
PH: 708 891 9305 (IL)

MUNSTER
701 Superior Avenue
Munster, IN 46321
PH: 219 922 4200

Standard Nuclear Consultants, Inc.
Elburn, IL 60119
630/365-1200
IEMA License No. IL-01300-01

LEAK TEST REPORT FOR RADIOACTIVE SEALED SOURCES

Licensee: Cardiospecialists Group, Ltd., Munster, IN

Cs-137	Co-57
Manufacturer: N.A. Scientific	Manufacturer: RADQUAL
Model No. MED 3550 Vial	Model No. BM05-15
Serial No. 16686	Serial No. BM051517136101
Cal. Date: 3/1/2002	Cal. Date: 5/15/2017
Activity: 193.8 uCi	Activity: 15 mCi
Results	Results
Gross CPM 376	Gross CPM 397
Bkgd. CPM 389	Bkgd. CPM 389
Net CPM 000	Net CPM 008
<0.005 uCi removable activity	<0.005 uCi removable activity

Leak test samples acquired by Margie Biltgen on December 19, 2017

Leak test samples analyzed by Jim Mikowski on December 24, 2017 / *gm*

Reviewed by: *[Signature]*
Radiation Safety Officer

NEXT LEAK TEST DUE: June 2018

Song, Taehoon

From: Strohmeyer, Daniel
Sent: Thursday, April 19, 2018 9:53 AM
To: Pavon, Sandy; Sandrik, Lauren; Song, Taehoon; Tomczak, Tammy
Subject: FW: Re: Re: [ACTION] CN 602706 CardioSpecialist
Attachments: Scanned from a Franciscan multifunction device.pdf; Scanned From a Franciscan MFD.PDF

Categories: Material Licensing

Good Morning,

Please add the attach to ADAMS for CN 602706.

Thanks,
Daniel