



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION IV  
1600 E. LAMAR BLVD  
ARLINGTON TX 76011-4511

April 6, 2018

Mr. Dale Schippers  
Radiation Safety Officer  
The Queen's Medical Center  
1301 Punchbowl Street  
Honolulu, HI 96813

SUBJECT: TERMINATION OF NRC RADIOACTIVE MATERIALS LICENSE  
NUMBER 53-29099-01

Dear Mr. Schippers:

By letter dated November 30, 2017, Ms. Cindy Kamikawa, RN, MS informed the U.S. Nuclear Regulatory Commission (NRC) that North Hawaii Community Hospital wished to terminate their NRC radioactive materials license. The NRC Form 314 dated April 3, 2018, stated specifically that all radioactive material and in an amendment request from The Queen's Medical Center, the location of use, and several authorized users had been transferred and added to The Queen's Medical Center, License Number 53-16533-02, as appropriate.

The supporting documentation to the NRC Form 314 also provided a copy of the latest leak test results. The leak test results document that there was no evidence of leakage of radioactive material from the sealed source. In addition, the supporting documentation to the NRC Form 314 stated that the licensee only possessed radioactive materials with radioactive decay half-lives of less than 120 days. A review of the historical file supported that only short lived radioactive materials were possessed by the licensee. The NRC staff has reviewed the supporting documentation to your NRC Form 314. Based on these conclusions, no further remediation or actions with respect to NRC regulated material is required.

Please find enclosed Amendment Number 16 to License Number 53-29099-01 authorizing North Hawaii Community Hospital's license termination. An environmental assessment for this action is not required since this action is categorically excluded under 10 CFR 51.22(c)(20)(i) and (ii).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

You can contact me at 817-200-1132 if you have questions or require clarification on any of the information stated above.

Sincerely,

*/RA/*

Jacqueline D. Cook, Senior Health Physicist  
Materials Licensing and Inspection Branch

Docket: 030-34081  
License: 53-29099-01  
Control: 602215

Enclosure: As stated