

From: [Lanzisera, Penny](#)
To: [Dan Berkley](#)
Subject: Request for additional information
Date: Thursday, April 12, 2018 12:12:00 PM

Licensee – Camden Clark Memorial Hospital Corporation
License No. 47-09772-02
Docket No. 03003390
Mail Control 602493

Mr. Berkley, please submit the following additional information to support your amendment request dated February 12, 2018:

1. Please provide an inventory list of sources possessed prior to the waste shipment and an inventory list of sources disposed to Ecology Services. In addition, please provide confirmation of receipt of these same sources by Ecology Services. For instance, you were licensed for several sources totaling 600 millicuries and appear to have disposed of 455 millicuries. It is unclear if the difference is associated with decay alone.
2. Please provide the last leak test records for the sources transferred.
3. With regards to your new treatment room, please provide a description of:
 - a. Warning systems and restricted area controls (e.g., locks, signs, warning lights and alarms, interlock systems);
 - b. Viewing and intercom systems;
 - c. Steps that will be taken to ensure that no two units can be operated simultaneously; and
 - d. Emergency response equipment stored in treatment room.

You may submit the information in a letter with attachments either via signed pdf sent via email or fax to 610-337-5269. Please reference Mail Control No. 602493 in your response. Thank you for your assistance,

Penny Lanzisera
Senior Health Physicist
U.S. NRC, Region 1