



UNITED STATES
NUCLEAR REGULATORY COMMISSION
 REGION IV
 1800 EAST LAMAR BLVD
 ARLINGTON, TEXAS 76011-4511

EMAIL



Name: Jeremy Axbom, Regulatory Compliance Manager License: 43-35027-01
Organization: Construction Materials Technology
 d/b/a CMT Engineering Laboratories
 d/b/a Precision Calibration
Phone: 801-386-6594 Docket: 030-38599
E-mail Address: jeremy.axbom@cmtlaboratories.com Control: 602446
From: Jacqueline D. Cook
Date: May 9, 2018
Subject: Letter dated February 15, 2018 for License Amendment
Pages: 2

Mr. Axbom:

Per your letter dated February 15, 2018, the items on the next page are deficiencies which require your response. **Please respond to this e-mail by Friday, May 18, 2018.** Our fax number is (817) 200-1263. **Please provide a response in a signed and dated letter in pdf format when responding via email.** My email address is Jackie.Cook@nrc.gov. When responding to this e-mail, please include the license, docket, and control numbers located at the top of this page.

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

/RA/

Jacqueline D. Cook
 Senior Health Physicist

PUBLIC
 Immediate Release
 Normal Release

NON-PUBLIC
 A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other:

Reviewer: JDC Date: 5/9/18

1. Please note NUREG-1556, Vol. 1, Rev. 2, *Consolidated Guidance About Materials Licenses: Program-Specific Guidance About Portable Gauge Licenses* (<http://www.nrc.gov/reading-rm/doc-collections/nuregs/staff/sr1556/v1/>), 8.9, Item 9 request the following:

Provide a facility diagram for each permanent portable gauge storage location (the new storage location requested in Item 2 of your letter dated February 15, 2018). Please include on the diagram the use of adjacent areas (including above and below), and information relevant to public dose and security as discussed in Sections 8.10.5, "Public Dose," and 8.10.6, "Operating, Emergency, and Security Procedures," respectively, in NUREG-1556, Vol. 1, Rev. 2.

2. In your amendment application dated February 15, 2018, you requested to remove a location of use and storage on your license for unrestricted use, please submit a copy(ies) of the latest leak test result(s) and a statement that "no leaking sources have ever been identified" while stored at this storage location.
3. Please specify the quantity of Instrotek Model 3500 series & Troxler Electronic Laboratories Model 4640 gauges you propose to have in your possession.
4. Please specify the zip code of Nellis AFB, Nevada.