

VIRGINIA ELECTRIC AND POWER COMPANY

RICHMOND, VIRGINIA 23261

August 4, 1980

Mr. James P. O'Reilly, Director
Office of Inspection and Enforcement
U. S. Nuclear Regulatory Commission
Region II
101 Marietta Street, Suite 3100
Atlanta, Georgia 30303

Serial No. 623
NO/RMT:ms
Docket Nos. 50-280
50-281
License Nos. DPR-32
DPR-37

Dear Mr. O'Reilly:

We have reviewed your letter of July 7, 1980 in reference to the inspection conducted at Surry Power Station on June 3-6, 1980, and reported in IE Inspection Report Nos. 50-280/80-21 and 50-281/80-22. Our response to the specific infraction is attached.

We have determined that no proprietary information is contained in the report. Accordingly, the Virginia Electric and Power Company has no objection to these inspection reports being made a matter of public disclosure.

Very truly yours,



B. R. Sylvia
Manager - Nuclear
Operations and Maintenance

Attachment

cc: Mr. Steven A. Varga

A001
5
1/1

8008060320

Q

RESPONSE TO VIOLATION LISTED IN
IE INSPECTION REPORT NOS.
50-280/80-21 AND 50-281/80-22

NRC Comments:

As required by Technical Specification 6.4.A, detailed written procedures with appropriate check off lists shall be provided for release of radioactive effluents.

Contrary to the above, there were no written procedures for preparation of efficiency calibration standard for the Ge(Li) detectors which are used to assure compliance with radioactive effluent release limits.

Response:

The information is correct as stated. A procedure did exist for calibration of the Ge(Li) detector, however the procedure did not contain detailed instructions for source preparation.

1. Corrective steps which have been taken:

As a result of the infraction involving insufficient details for the preparation of source standards, Health Physics Procedure - 3.4.2 has been written which more clearly details the preparation method.

2. Corrective steps taken to avoid further non-compliance: Health Physics Procedure 3.4.2 will be used in preparation of source standards.

3. Date when full compliance will be achieved:

Full compliance was achieved July 17, 1980.

RESPONSE TO VIOLATION LISTED IN
IE INSPECTION REPORT NOS.
50-280/80-21 AND 50-281/80-22

NRC Comments:

As required by Technical Specification 6.4.A, detailed written procedures with appropriate check off lists shall be provided for release of radioactive effluents.

Contrary to the above, there were no written procedures for preparation of efficiency calibration standard for the Ge(Li) detectors which are used to assure compliance with radioactive effluent release limits.

Response:

The information is correct as stated. A procedure did exist for calibration of the Ge(Li) detector, however the procedure did not contain detailed instructions for source preparation.

1. Corrective steps which have been taken:

As a result of the infraction involving insufficient details for the preparation of source standards, Health Physics Procedure - 3.4.2 has been written which more clearly details the preparation method.

2. Corrective steps taken to avoid further non-compliance: Health Physics Procedure 3.4.2 will be used in preparation of source standards.

3. Date when full compliance will be achieved:

Full compliance was achieved July 17, 1980.