



Region I Office
 Division of Nuclear Materials Safety
 2100 Renaissance Boulevard, Suite 100
 King of Prussia, PA 19406-2713
 (610)337-5000

Telephone Conversation Record

Date: 5/11/18

License No. 44-10187-03

Docket No. (no hyphens): 03003289

Mail Control/Report No. 602772

Licensee Name: The University of Vermont Medical Center

Participant(s) Name/Title: Brian Oyudomari, RSO

Work Telephone No. 802-847-4897

Business Cellphone No.

NRC Representative Name/Title: Tara Weidner, Sr HP

Subject: RAI (add AMP, remove LOU, add LOU)

Discussion:

In the letter dated March 14, 2018, UVM requested that

- (1) Emilie Soisson, Ph.D. be added as an AMP for the Varian HDR. Dr. Soisson completed her training in the US, but has been a medical physicist at McGill University Health Center in Quebec, Canada from 2009-2016. The documentation provided by UVM did not appear to adequately show her experience at McGill. Therefore, I requested that UVM provide a NRC 313A (AMP) to show Dr. Soisson's training and experience.
- (2) The Molecular Imaging Room 1243 be added as a location of use. This area was previously on the license, but was removed for renovations and the replacement of the PET/CT camera. According to the RSO, the shielding previously in the room was not touched during the install of the new camera. I requested that he provide a revised floor plan for the area
- (3) The Alliance imaging mobile van be removed as a LOU. During the renovation of the above PET/CT area, Alliance Imaging provided a mobile van for UVMs exclusive use. As of 12/20/17, use of RAM ceased in the mobile van. A final survey was completed on 12/27/17. According to the RSO, Alliance has taken possession of the van and the sealed sources contained therein. I requested documentation of the transfer of the van and sealed sources.

Action Required: Wait for additional information

SUNSI REVIEW	
Document Availability:	<input checked="" type="checkbox"/> Public or <input type="checkbox"/> Non-Public
Document Sensitivity:	
<i>(select "1" value to the right)</i>	
<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/> MD 3.4 Non-Public B.1 (Non-Sensitive)
<input type="checkbox"/> Non-Sensitive Copyright	<input type="checkbox"/> MD 3.4 Non-Public A.3 (Sensitive Security Related)
<input type="checkbox"/> Sensitive – Proprietary	<input type="checkbox"/> MD 3.4 Non-Public A.4 (Sensitive Proprietary)
<input type="checkbox"/> Sensitive – Privacy Act (includes Personally Identifiable Information (PII))	<input type="checkbox"/> MD 3.4 Non-Public A.5 (Sensitive-Privacy Act)
<input type="checkbox"/> Sensitive – Internal, Periodic Review required (All Other Sensitive Internal Info.)	<input type="checkbox"/> MD 3.4 Non-Public A.6 (Sensitive-Federal, State, etc)
<input type="checkbox"/> Sensitive – Security-Related-Periodic Review Required	<input type="checkbox"/> MD 3.4 Non-Public A.7 (Internal)
SUNSI Review Completed by: TLWeidner	