



<b>NRC FORM 314</b> (02-2017) 10 CFR 30.360(N)(1) 40 CFR 61.101, 70.380(N)(1) and 72.540(N)(1)(1)	<b>U.S. NUCLEAR REGULATORY COMMISSION</b>  <div style="text-align: center;">   <b>CERTIFICATE OF DISPOSITION OF MATERIALS</b> </div>	APPROVED BY OMB: NO. 3160-0028  EXPIRES: 02/28/2020  <small>Estimated license fee response to comply with the mandatory collection request 30 months. This subtotal is used by NRC as part of the basis for the determination that the facility is licensed for radioactive use. Send comments regarding license activities to the PDA, Safety, and Information Collection Branch (7-1183), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by email to: <a href="mailto:licensing@nrc.gov">licensing@nrc.gov</a>, and to the Desk Officer, Office of Information and Regulatory Affairs, NRCB-5202, (3190-0038), Office of Management and Budget, Washington, DC 20503. If a license is not to appear in information collection does not display a currently valid OMB control number, the NRC may not conduct or license, and a person is not required to respond to, the information collection.</small>			
LICENSEE NAME AND ADDRESS Vermont Testing & Consulting Corporation PO Box 420 Waterbury, VT 05676	<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:50%;">LICENSE NUMBER 44-28180-02</td> <td style="width:50%;">DOCKET NUMBER 030-36475</td> </tr> <tr> <td colspan="2">LICENSE EXPIRATION DATE April 30, 2024</td> </tr> </table>	LICENSE NUMBER 44-28180-02	DOCKET NUMBER 030-36475	LICENSE EXPIRATION DATE April 30, 2024	
LICENSE NUMBER 44-28180-02	DOCKET NUMBER 030-36475				
LICENSE EXPIRATION DATE April 30, 2024					
<b>A. LICENSE STATUS (Check the appropriate box)</b> <input type="checkbox"/> This license has expired. <input checked="" type="checkbox"/> This license has not yet expired; please terminate it.					
<b>B. DISPOSAL OF RADIOACTIVE MATERIAL</b> <i>(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)</i> The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:					
<input type="checkbox"/> 1. No radioactive materials have ever been procured or possessed by the licensee under this license. <input type="checkbox"/> 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner: <input checked="" type="checkbox"/> a. Transfer of radioactive materials to the licensee listed below: North East Technical Services, Inc  <input type="checkbox"/> b. Disposal of radioactive materials: <input type="checkbox"/> 1. Directly by the licensee:  <input type="checkbox"/> 2. By licensed disposal site:  <input type="checkbox"/> 3. By waste contractor:  <input type="checkbox"/> c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.					
<b>C. SURVEYS PERFORMED AND REPORTED</b>					
<input type="checkbox"/> 1. A radiation survey was conducted by the licensee. The survey confirms: <input type="checkbox"/> a. the absence of licensed radioactive materials <input type="checkbox"/> b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA. <input type="checkbox"/> 2. A copy of the radiation survey results: <input type="checkbox"/> a. is attached; or <input type="checkbox"/> b. is not attached (Provide explanation); or <input type="checkbox"/> c. was forwarded to NRC on: _____ Date _____ <input checked="" type="checkbox"/> 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and <input checked="" type="checkbox"/> a. The results of the latest leak test are attached; and/or <input type="checkbox"/> b. No leaking sources have ever been identified.					
The person to be contacted regarding the information provided on this form:					
NAME Jacques B. Bouramia	TITLE President/RSO	TELEPHONE (include Area Code) 802-229-8685			
E-MAIL ADDRESS Jacquesbouramia@msn.com					
Mail all future correspondence regarding this license to: Vermont Testing & Consulting Corp PO Box 420					
<b>C. CERTIFYING OFFICIAL</b> I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT					
PRINTED NAME AND TITLE Jacques B. Bouramia, President/RSO	SIGNATURE 	DATE 05/31/18			
<small>WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.</small>					

**REC'D IN LAT** 06/04/2018

608749  
NMSS/RGNI MATERIALS-002



**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057  
Research Triangle Park, NC 27709  
Tel: (877) 876-9537 Fax: (866) 391-2759  
License: NC 032-0182-1

Cust ID: 13195

JACQUES BOURAMIA  
VERMONT TESTING & CONSULTING  
45 BOURMIA DRIVE  
PO BOX 420  
WATERBURY, VT 05676

**LEAK TEST CERTIFICATE**

**DEVICE:**

Model: 3430      Serial No: 19154

**SEALED SOURCES:**

Serial No.	Measure Date	Nuclide	GBq	mCi
50-8797	04/10/1990	Cs-137	0.296	8
47-14620	05/25/1990	Am-241:Be	1.48	40

**LEAK TEST ANALYSIS:**

Sample collected on: 02/12/2018  
Sample analyzed on: 03/07/2018 12:21:24 P      Position: 30  
Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	24
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.  
 3008 Cornwallis Rd., P.O. Box 12057  
 Research Triangle Park, NC 27709  
 Tel: (877) 876-9637 Fax: (866) 391-2759  
 License: NC 032-0182-1

Cust ID: 13195

JACQUES BOURAMIA  
 VERMONT TESTING & CONSULTING  
 45 BOURMIA DRIVE  
 PO BOX 420  
 WATERBURY, VT 05676

**LEAK TEST CERTIFICATE**

**DEVICE:**

Model: 3440      Serial No: 27533

**SEALED SOURCES:**

Serial No.	Measure Date	Nuclide	GBq	mCi
750-1332	04/24/1997	Cs-137	0.296	8
47-24064	11/27/1996	Am-241:Be	1.48	40

**LEAK TEST ANALYSIS:**

Sample collected on: 02/12/2018  
 Sample analyzed on: 03/07/2018 12:20:11 P      Position: 29  
 Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	31
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)

*Handwritten signature*



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12067  
Research Triangle Park, NC 27709  
Tel: (877) 876-9537 Fax: (866) 391-2759  
License: NC 032-0182-1

Cust ID: 13195

JACQUES BOURAMIA  
VERMONT TESTING & CONSULTING  
45 BOURMIA DRIVE  
PO BOX 420  
WATERBURY, VT 05676

LEAK TEST CERTIFICATE

DEVICE:

Model: 3430 Serial No: 35824

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-2904	06/11/2004	Cs-137	0.296	8
78-616	07/26/2004	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 02/12/2018  
Sample analyzed on: 03/07/2018 12:18:59 P Position: 28  
Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	29
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)



**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057  
Research Triangle Park, NC 27709  
Tel: (877) 876-9537 Fax: (866) 391-2759  
License: NC 032-0182-1

Cust ID: 13195

JACQUES BOURAMIA  
VERMONT TESTING & CONSULTING  
45 BOURMIA DRIVE  
PO BOX 420  
WATERBURY, VT 05676

**LEAK TEST CERTIFICATE**

**DEVICE:**

Model: 3440      Serial No: 30464

**SEALED SOURCES:**

Serial No.	Measure Date	Nuclide	GBq	mCi
750-4936	04/27/1999	Cs-137	0.296	8
47-27529	07/28/1998	Am-241:Be	1.48	40

**LEAK TEST ANALYSIS:**

Sample collected on: 02/12/2018  
Sample analyzed on: 03/07/2016 12:17:46 P      Position: 27  
Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	25
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)



**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057  
Research Triangle Park, NC 27709  
Tel: (877) 876-9537 Fax: (866) 391-2759  
License: NC 032-0182-1

JACQUES BOURAMIA  
VERMONT TESTING & CONSULTING.  
45 BOURMIA DRIVE  
PO BOX 420  
WATERBURY, VT 05678

Cust ID: 13195

**LEAK TEST CERTIFICATE**

**DEVICE:**

**Model:** 3430      **Serial No:** 35866

**SEALED SOURCES:**

Serial No.	Measure Date	Nuclide	GBq	mCi
77-2976	09/16/2004	Cs-137	0.296	8
78-843	07/26/2004	Am-241:Be	1.48	40

**LEAK TEST ANALYSIS:**

**Sample collected on:** 02/12/2018 *After Six Month Again on 08/12/18*  
**Sample analyzed on:** 03/07/2018 12:16:34 P      **Position:** 26  
**Analyzed by:** HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	21
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)