

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

| | |
|--|---|
| <p>1. LICENSEE/LOCATION INSPECTED:</p> <p>American Health Network of Indiana, LLC 10689 North Pennsylvania Street Indianapolis, IN 46280</p> <p>REPORT NUMBER(S) 2018001</p> | <p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p> |
|--|---|

| | | |
|---|--|---|
| <p>3. DOCKET NUMBER(S)</p> <p>030-36851</p> | <p>4. LICENSE NUMBER(S)</p> <p>13-32555-01</p> | <p>5. DATE(S) OF INSPECTION</p> <p style="text-align: right;">MAY 22nd, 2018</p> |
|---|--|---|

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

| TITLE | PRINTED NAME | SIGNATURE | DATE |
|---------------------------|--------------|-------------------|------------|
| LICENSEE'S REPRESENTATIVE | | | |
| NRC INSPECTOR | Ryan Craffey | <i>R. Craffey</i> | 5/22/18 |
| BRANCH CHIEF | Aaron McCraw | <i>A. McCraw</i> | 06/01/2018 |

Docket File Information

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

| | |
|--|---|
| <p>1. LICENSEE/LOCATION INSPECTED:</p> <p>American Health Network of Indiana, LLC 10689 North Pennsylvania Street Indianapolis, IN 46280</p> <p>REPORT NUMBER(S) 2018001</p> | <p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p> |
|--|---|

| | | |
|---|--|---|
| <p>3. DOCKET NUMBER(S)</p> <p>030-36851</p> | <p>4. LICENSE NUMBER(S)</p> <p>13-32555-01</p> | <p>5. DATE(S) OF INSPECTION</p> <p>May 22, 2018</p> |
|---|--|---|

| | |
|---|---|
| <p>6. INSPECTION PROCEDURES USED</p> <p>87131</p> | <p>7. INSPECTION FOCUS AREAS</p> <p>All</p> |
|---|---|

SUPPLEMENTAL INSPECTION INFORMATION

| | | | |
|--|-----------------------------|---|--|
| <p>1. PROGRAM CODE(S)</p> <p>02120</p> | <p>2. PRIORITY</p> <p>3</p> | <p>3. LICENSEE CONTACT</p> <p>Edward Wroblewski, MA - RSO</p> | <p>4. TELEPHONE NUMBER</p> <p>(317) 755-9688</p> |
|--|-----------------------------|---|--|

Main Office Inspection Next Inspection Date: 05/22/2021

Field Office Inspection 3631 Morrison Road, Suite 200, Muncie, IN

Temporary Job Site Inspection _____

PROGRAM SCOPE

This was an unannounced routine inspection of a state-wide physician group practice authorized to use byproduct material for diagnostic and therapeutic medical purposes at its family medicine and specialty care clinic in Muncie, Indiana. At the time of the inspection, one technologist performed around five diagnostic administrations per day, Monday through Thursday, using unit doses only. The licensee had not performed any therapeutic administrations since before the last inspection, and had no plans to do so in the foreseeable future. The licensee retained the services of a consulting physicist who audited the radiation safety program quarterly.

PERFORMANCE OBSERVATIONS

The inspector toured the clinic in Muncie to evaluate the licensee's measures for materials security, hazard communication, and exposure control. The inspector conducted independent surveys of the facility, and found no evidence of residual contamination or area exposure rates in excess of regulatory limits to members of the public. The inspector observed the administration of two unit doses for cardiac stress tests, as well as demonstrations by the technologist of the receipt of packages containing radioactive material, the conduct of area surveys, and decay-in-storage radioactive waste handling.

The inspector also reviewed a selection of records, including consultant audits, documentation of various radiation surveys and instrument quality assurance checks, sealed source inventories and leak test results, decay-in-storage waste disposal logs, and personnel dosimetry reports.

No violations of NRC requirements were identified as a result of this inspection.