



CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to InfoCollect.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NE08-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

ENGINEERING & TESTING 2000, INC.
274 ALLEN LANE LEWISBURG WV, 24901

Q5

LICENSE NUMBER

47-19327-01

DOCKET NUMBER

030-19035

LICENSE EXPIRATION DATE

JUNE 30, 2021

A. LICENSE STATUS (Check the appropriate box)

- This license has expired.
- This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:
 - a. Transfer of radioactive materials to the licensee listed below:
TROXLER TRANSFERED 3-27-2018, NORTH CAROLINA RML#032-0182-01 CONTACT :EMILY DAVIS 919-485-2228
 - b. Disposal of radioactive materials:
 - 1. Directly by the licensee:
 - 2. By licensed disposal site:
 - 3. By waste contractor:
 - c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

REC PG 1 05 29 18 AM 05 58

C. SURVEYS PERFORMED AND REPORTED

- 1. A radiation survey was conducted by the licensee. The survey confirms:
 - a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- 2. A copy of the radiation survey results:
 - a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date
- 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
BERNARD HOLLIDAY	VP-RSO	681-318-3527	bholliday@suddenlinkmail.com

Mail all future correspondence regarding this license to:

274 ALLEN LANE LEWISBURG WV 24901

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
BERNARD HOLLIDAY	<i>Bernard Holliday</i>	5-22-18 608961

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.



Troxler Electronic Laboratories, Inc. - Troxler International, Ltd.
3008 Cornwallis Road, P.O. Box 12057, Research Triangle Park, NC 27709 - USA
Telephone: 919/549-8661 Telefax: 919/549/0761

March 29, 2018

Engineering Testing 2000
Attn: Bernard Holliday
274 Allen Way
Lewisburg, WV 24901

Attn: Bernard Holliday

This is to acknowledge receipt and transfer of ownership of the nuclear gauge(s) listed below, and the radioactive sources contained therein, to Troxler under North Carolina Radioactive Material License #032-0182-1. You should retain this letter in your files for future reference.

Date Received: 03-27-18

Job Number	Model	Serial Number
56437	3411	9667
56438	3411	7944
56439	3411	4664
56440	3411	4776

Regards,

Emily Davis
Service/Repair Department
Troxler Electronics



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

BERNARD HOLLIDAY
ENGINEERING & TESTING 2000
PO BOX 1725
LEWISBURG, WV 24901

Cust ID: 13633

LEAK TEST CERTIFICATE

DEVICE:

Model: 3411 Serial No: 4664

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
40-1660	07/18/1977	Cs-137	0.296	8
47-694	07/18/1977	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 03/27/2018

Sample analyzed on: 03/29/2018 4:47:23 PM Position: 14

Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	21
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi)

Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

BERNARD HOLLIDAY
ENGINEERING & TESTING 2000
PO BOX 1725
LEWISBURG, WV 24901

Cust ID: 13633

LEAK TEST CERTIFICATE

DEVICE:

Model: 3411 **Serial No:** 9667

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
46-1029	06/03/1983	Am-241:Be	1.48	40
40-7017	06/03/1983	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 03/27/2018
Sample analyzed on: 03/29/2018 4:48:36 PM **Position:** 15
Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	34
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

BERNARD HOLLIDAY
ENGINEERING & TESTING 2000
PO BOX 1725
LEWISBURG, WV 24901

Cust ID: 13633

LEAK TEST CERTIFICATE

DEVICE:

Model: 3411 Serial No: 4776

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
40-1800	08/26/1977	Cs-137	0.296	8
47-838	08/26/1977	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 03/27/2018
Sample analyzed on: 03/29/2018 4:49:48 PM Position: 16
Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	29
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

- Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.
 3008 Cornwallis Rd., P.O. Box 12057
 Research Triangle Park, NC 27709
 Tel: (877) 876-9537 Fax: (866) 391-2759
 License: NC 032-0182-1

BERNARD HOLLIDAY
 ENGINEERING & TESTING 2000
 PO BOX 1725
 LEWISBURG, WV 24901

Cust ID: 13633

LEAK TEST CERTIFICATE

DEVICE:

Model: 3411 Serial No: 7944

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
40-5175	06/03/1981	Cs-137	0.296	8
47-4266	06/03/1981	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 03/27/2018
 Sample analyzed on: 03/29/2018 4:51:01 PM Position: 17
 Analyzed by: HEB

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.21E+01	2.02E+01
Background measurement (cpm)	0	25
Sample measurement (cpm)	0	17
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	4.2E-01	1.3E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)

DIANNE & HOBBS
274 ALLEN HALL
LEWISBURG, WV 24901



U.S. POSTAGE
PAID
LEWISBURG, WV
24901
MAY 22, 1988
AMOUNT
\$1.21
R2305K143053-06

Lnms

LICENSING ASSISTANT SECTION
NUCLEAR MATERIALS SAFETY BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION 1
2100 RENAISSANCE BOULEVARD, SUITE 100
KING OF PRUSSIA, PA 19106-2713



ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE

Name and Address of Applicant and/or Licensee Engineering & Testing 2000, Inc. ATTN: William L. Toney, Jr., President P. O. Box 1149 Lewisburg, WV 24901-4049	Date June 4, 2018
	License Number(s) 47-19327-01
	Mail Control Number(s) 608961
	Licensing and/or Technical Reviewer or Branch Betsy Ullrich

This is to acknowledge receipt of your: Letter and/or Application Dated: 05/22/2018

The initial processing, which included an administrative review, has been performed.
 Amendment Termination New License Renewal

There were no administrative omissions identified during our initial review.

This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

Your application for a new NRC license did not include your taxpayer identification number. Please complete and submit NRC Form 531, Request for Taxpayer Identification Number, located at the following link: <http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>
 Follow the instructions on the form for submission.

The following administrative omissions have been identified:
 [Empty box for listing omissions]

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region I
 U. S. Nuclear Regulatory Commission
 Division of Nuclear Materials Safety
 2100 Renaissance Boulevard, Suite 100
 King of Prussia, PA 19406-2713
 (610) 337-5260, (610) 337-5313,
 (610) 337-5398, or (610) 337-5239