




UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
1600 EAST LAMAR BLVD
ARLINGTON, TEXAS 76011-4511

EMAIL



Name: Troy Curnutt License: 11-34221-01
Organization: Diagnostic Imaging Services of Idaho Docket: 030-36948
Phone: 208-406-2543 Control: 602789
E-mail Address: nukemdude@gmail.com
From: Jacqueline D. Cook
Date: June 2, 2018
Subject: Letter dated March 20, 2018 for License Amendment
Pages: 2

Mr. Curnutt:

Per your letter dated March 20, 2018, the item on the next page is a deficiency which requires your response. **Please respond to this e-mail by Monday, June 18, 2018.** Our fax number is (817) 200-1263. **Please provide a response in a signed and dated letter in pdf format when responding via email.** My email address is Jackie.Cook@nrc.gov. When responding to this e-mail, please include the license, docket, and control numbers located at the top of this page.

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

/RA/

Jacqueline D. Cook
Senior Health Physicist

PUBLIC

- Immediate Release
 Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
 A.7 Sensitive Internal
 Other:

Reviewer: JM Date: 6/2/18

1. In your amendment request dated March 20, 2018, you told us you have moved physical locations and request transfer of 2 calibration reference sources/standards to a new storage location. In order to remove the old location of use and storage on your license for unrestricted use, please submit a copy(ies) of the latest leak test result(s) and a statement, if applicable and appropriate, that "no leaking sources have ever been identified" while stored at this storage location.