



MEMO

C O U N T Y M I C H I G A N

DEPARTMENT OF FACILITIES MANAGEMENT

FACILITIES PLANNING & ENGINEERING

To: Laura Cender, U. S. Nuclear Regulatory Commission (NRC)

From: Carmelo Moyet, Jr., Radiation Safety Officer
Facilities Planning & Engineering
1 Public Works Drive
Waterford, MI 48328

Date: June 21, 2018

Re: U. S. NRC Materials License 21-24953-01

The gauge manufacturer or other person licensed by the NRC or an Agreement State will perform non-routine maintenance or repair operations that require detaching the source or source rod from the gauge.

Thank you.

Carmelo Moyet, Jr., Radiation Safety Officer

Cender, Laura

From: Moyet, Carmelo <moyetc@oakgov.com>
Sent: Thursday, June 21, 2018 3:43 PM
To: Cender, Laura
Cc: Moyet, Carmelo
Subject: [External_Sender] RE: NRC License Amendment Request - Additional Information Required
Attachments: Non Routine Maintenance Statement Letter 062118.pdf

Laura,

Thank you for your time on this matter.

Please find attached Non-Routine Maintenance and Repair statement letter.

Thank you.

Carmelo

From: Cender, Laura [mailto:Laura.Cender@nrc.gov]
Sent: Thursday, June 21, 2018 2:06 PM
To: Moyet, Carmelo
Subject: NRC License Amendment Request - Additional Information Required

Hello Carmelo,

Thank you for taking time out of your day to discuss you recent NRC license amendment request. As we discussed, your current license allows for non-routine maintenance activities to be performed by the former RSO William Jackson. As Mr. Jackson is no longer with your organization we will need to update this section with this amendment. Per our discussion today my understanding is that you are not wanting to perform non-routine maintenance activities at this time. If this is the case, please provide the following commitment to only have non-routine maintenance operations performed by an Agreement State or NRC authorized entity.

"The gauge manufacturer or other person licensed by the NRC or an Agreement State will perform non-routine maintenance or repair operations that require detaching the source or source rod from the gauge."

Please ensure that your statement is both signed and dated, and please feel free to contact me at 630-829-9712 or via email if you have any questions.

Thank you,
Laura

Laura Cender
U.S. Nuclear Regulatory Commission
Materials Licensing Branch
E-mail: Laura.Cender@nrc.gov
Phone: (630) 829-9712