

**From:** [Nguyen, Janice](#)  
**To:** [JRiopel@echm.org](mailto:JRiopel@echm.org)  
**Subject:** NRC Request for Additional Information for Prospect Manchester Hospital and Prospect Rockville Hospital (Mail Control Numbers 608629 and 608639)  
**Date:** Friday, June 15, 2018 3:31:00 PM

---

Prospect Manchester Hospital, Inc.  
d/b/a The Manchester Memorial Hospital  
License No. 06-03413-01  
Docket No. 030-01253  
Mail Control No. 608629

Prospect Rockville Hospital, Inc.  
d/b/a The Rockville General Hospital Incorporated  
License No. 06-13001-02  
Docket No. 030-09938  
Mail Control No. 608639

Dear Mr. Riopel,

**Could you please reply back to this email to confirm receipt?**

This is in reference to your letter dated April 23, 2018, requesting to amend Nuclear Regulatory Commission Licenses No. 06-03413-01 and 06-13001-02. As we discussed by telephone this afternoon, we need the following additional information in order to add Alok Azad Anand, M.D. as an authorized user:

- 1) In order to verify Dr. Anand's preceptors, please provide a copy of the radioactive material licenses that list Peggy Chan, M.D. and Stanley J. Goldsmith, M.D. as authorized users of 10 CFR 35.100, 35.200, and 35.300 materials. If the licenses are broad scope licenses that do not specifically list authorized users, please provide a copy of the license permit for each authorized user, showing what materials they are currently authorized to use.
- 2) We currently do not recognize Dr. Anand's Board Certification for authorization of 10 CFR 35.396, "Parenteral administration of any beta-emitter, or photon-emitting radionuclide with a photon energy less than 150 keV for which a written directive is required." If you wish Dr. Anand to be authorized for 10 CFR 35.396, please resubmit the NRC Form 313A (AUT) using the training and experience pathway, showing he has completed 700 hours of training and experience, including a minimum of 200 hours of classroom and laboratory training, as required by 10 CFR 35.390(b)(1). In addition, he should provide casework and preceptor attestation.

Please note that Yttrium-90 microspheres casework falls under 10 CFR 35.1000, and does not count toward 10 CFR 35.396.

We will continue our review upon receipt of the requested information. You may respond to my attention in writing by letter, email (if letter is signed by senior management and scanned into a pdf format), or fax (610-337-5269), referencing mail control numbers 608629 and 608639. If we do not receive a reply from you within 30 days, we will assume that you do not wish to pursue your application. If you have any technical questions

regarding this deficiency letter, please call me at (610) 337-5006.

Current NRC regulations and guidance are included on the NRC's website at [www.nrc.gov](http://www.nrc.gov); select **Nuclear Materials; Med, Ind, & Academic Uses**; then **Licensee Toolkits, see our toolkit index page**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 8:00 a.m. to 5:30 p.m. EST, Monday through Friday (except Federal holidays).

Thank you in advance for your help!

*Jan*

Janice Nguyen  
Senior Health Physicist  
U.S. Nuclear Regulatory Commission  
2100 Renaissance Boulevard, Suite 100  
King of Prussia, PA 19406  
Office (610) 337-5006  
FAX (610) 337-5269