



CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (1-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to InfoCollect.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOF-10702, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

Power Mountain Contura, LLC
P.O. Box 707
Summersville, WV 26651-0707

Dr. 2

LICENSE NUMBER

47-25566-01

DOCKET NUMBER

030-35744

LICENSE EXPIRATION DATE

October 31, 2021

A. LICENSE STATUS (Check the appropriate box)

- This license has expired.
- This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
 - a. Transfer of radioactive materials to the licensee listed below:
Berthold Technologies USA, LLC; Oak Ridge, TN 37830-0895 (TRMLR-01082-B23)
 - b. Disposal of radioactive materials:
 - 1. Directly by the licensee:
 - 2. By licensed disposal site:
 - 3. By waste contractor:
 - c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- 1. A radiation survey was conducted by the licensee. The survey confirms:
 - a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- 2. A copy of the radiation survey results:
 - a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date
- 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Christopher Ray	Superintendent	304-872-0852 <i>304-872-8123</i>	Christopher.Ray@conturaenergy.com

Mail all future correspondence regarding this license to:

Power Mountain Contura, P.O. Box 707, Summersville, WV 26651-0707

C. CERTIFYING OFFICIAL
I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
Christopher Ray - Superintendent	<i>Christopher S Ray</i>	8-27-2018

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

609852
RADIOACTIVE MATERIALS-002

REC'D IN LAT *09/05/2018*



6312 Clinton Street
 Morton Grove, IL 60053-2723
 847-965-1999

SEALED SOURCE LEAK TEST ANALYSIS REPORT

Facility: Power Mountain Coal
 Address: #4 Jerry Fork Rd.
 Drennen, WV 26667

Person Submitting Samples: John Grizzell
 Date Samples Collected: 7/24/2018
 Send Results To: Power Mountain Coal
ATTN: Chris Ray
 #4 Jerry Fork Rd.
 Drennen, WV 26667

Telephone: 304-872-8123

FAX:

E-Mail: christopher.ray@conturaenergy.com

License No.:

EC: John Grizzell, Chester Douglas

EC E-Mail: john.grizzell@berthold-us.com, chester.douglas@berthold-us.com

RSSI SAMPLE NUMBER	CLIENT REFERENCE	RADIO-NUCLIDE	ACTIVITY (mCi)	SOURCE CALIBRATION DATE	DEVICE MANUFACTURER	DEVICE MODEL NUMBER	DEVICE SERIAL NUMBER	SOURCE SERIAL NUMBER	RESULTS* (μCi)
P182197	0305/07	Cs-137	30		Berthold	LB7440-D-CR	37624-11333	0305/07	<4.2×10 ⁻⁶
P182198	0090/06	Cs-137	100		Berthold	LB7440-D-CR	37624-13538	0090/06	<4.2×10 ⁻⁶
P182199	0307/07	Cs-137	30		Berthold	LB7440-D-CR	37624-11336	0307/07	<4.2×10 ⁻⁶
P182200	0313/07	Cs-137	30		Berthold	LB7440-D-CR	37624-12787	0313/07	<4.2×10 ⁻⁶
P182201	0308/07	Cs-137	30		Berthold	LB7440-D-CR	37624-12743	0308/07	<4.2×10 ⁻⁶
P182202	0310/07	Cs-137	30		Berthold	LB7440-D-CR	37624-12738	0310/07	<4.2×10 ⁻⁶

Analysis authorized by license No. IL-01429-01. Analysis approved by the Canadian Nuclear Safety Commission, meeting the criteria and requirements of R-116. Analysis instrument: Nuclear-Chicago Spectro Shield 1152 α/β Counter, serial number 28458. Calibrated February 23, 2018. Measuring time = 5 minutes. Alpha background = 0.08 cpm. Beta/Gamma background = 6.19 cpm. Analysis location: RSSI.

The analytical results relate only to the specific leak test kit(s) returned by the client, in the condition as received at the laboratory. Calculations of leak test results are based upon the radionuclide(s) as provided by the client. This report shall not be reproduced except in its entirety without the approval of RSSI. Most regulatory agencies consider a source to be leaking if a leak test indicates the presence of more than 0.005 μCi (5×10⁻³ μCi) of removable contamination. Sources for which the results indicate the presence of 5×10⁻⁴ μCi or more of removable activity should be investigated.

* < followed by a number indicates the activity is below the limit of detection.

A highlighted result exceeds regulatory limits and was reported by telephone to the person submitting the sample and/or _____ on _____.

Analyzed by: Aaron Morris
 P182197-202

Receipt Date: July 30, 2018
 Counting Date: July 30, 2018
 Analysis Date: July 30, 2018



Berthold Technologies USA, LLC
 99 Midway Lane
 Oak Ridge, TN 37830-0895 USA
 Phone: 865-483-1488
 Fax: 865-425-4309
www.berthold-us.com

Radioactive Source Transfer of Ownership

Sales Order Number:
PCP004581

Customer Name & Contact Info:
Power Mountain Contura
4 Jerry Fork Rd.
Drennen, WV 26667

This Certificate is to serve notice that Berthold Technologies USA, LLC located in Oak Ridge, TN (TRML R-01082-B23) has completed the inspection on the six (6) devices listed below and has transferred ownership to Power Mountain Contura under a General License. General Licensee training was provided and the transfer of ownership was effective 8/22/18.

Source #	Isotope	Original Activity (mCi)	Device Manufacturer
0313/07	Cs-137	30	Berthold
0310/07	Cs-137	30	Berthold
0308/07	Cs-137	30	Berthold
0307/07	Cs-137	30	Berthold
0305/07	Cs-137	30	Berthold
0090/06	Cs-137	100	Berthold

* See attached breakdown for shield Serial #s and Model #s.

Radiation Safety Officer: C. Douglas Date: 8-27-18

Berthold Technologies
Quarterly Report Of Installation



OR Actual Ship Date Between 1/1/2018 and 8/22/2018 and Ref No = 1483B and License Type = GL

Ref #	RSO Name / Phone	End User Address	OR Actual Ship Date	Isotope	Activity (mCi)	Shield Model / Serial Number	Source Model/Serial
Power Mountain Contura							
1483B	Jeff Henderson 423-292-6044	# 4 Jerry Fork Rd. Drennen WV 26667	08/21/2018	Cs-137	30	LB 7440-D-CR 37624-12787	Cs7.P02 0313/07
1483B	Jeff Henderson 423-292-6044	# 4 Jerry Fork Rd. Drennen WV 26667	08/21/2018	Cs-137	30	LB 7440-D-CR 37624-12738	Cs7.P02 0310/07
1483B	Jeff Henderson 423-292-6044	# 4 Jerry Fork Rd. Drennen WV 26667	08/21/2018	Cs-137	30	LB 7440-D-CR 37624-12743	Cs7.P02 0308/07

Number of Sources: 3

TOTAL Number of Sources: 3

Quarterly Report Of Installation

OR Actual Ship Date Between 1/1/2018 and 8/22/2018 and Ref No = 1482B and License Type = GL

Ref #	RSO Name / Phone	End User Address	OR Actual Ship Date	Isotope	Activity (mCi)	Shield Model / Serial Number	Source Model/Serial
Power Mountain Contura							
1482B	Jeff Henderson 423-292-6044	# 4 Jerry Fork Rd. Drennen WV 26667	08/21/2018	Cs-137	30	LB 7440-D-CR 37624-11336	Cs7.P02 0307/07
1482B	Jeff Henderson 423-292-6044	# 4 Jerry Fork Rd. Drennen WV 26667	08/21/2018	Cs-137	30	LB 7440-D-CR 37624-11333	Cs7.P02 0305/07
1482B	Jeff Henderson 423-292-6044	# 4 Jerry Fork Rd. Drennen WV 26667	08/21/2018	Cs-137	100	LB 7440-D-CR 37624-13538	Cs7.P02 0090/06

Number of Sources: 3


TOTAL Number of Sources: 3



Power Mountain Contura, LLC
P. O. Box 707
Summersville, WV 26651

CHARLESTON WV 250

29 AUG 2018 PM 2 L



Licensing Assistant Section
Nuclear Materials Safety Branch
U.S. Nuclear Regulatory Commission, Region 1
2100 Renaissance Boulevard, Suite 100

King of Prussia, PA 19406-0713



ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE

Name and Address of Applicant and/or Licensee Power Mountain Contura, LLC ATTN: William G. Newsome, RSO P. O. Box 707 Summersville, WV 26651-0707	Date September 10, 2018
	License Number(s) 47-25566-01
	Mail Control Number(s) 609852
	Licensing and/or Technical Reviewer or Branch Commercial, Industrial, R&D, & Academic Branch

This is to acknowledge receipt of your: Letter and/or Application Dated: 08/27/2018

The initial processing, which included an administrative review, has been performed.
 Amendment Termination New License Renewal

There were no administrative omissions identified during our initial review.

This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

Your application for a new NRC license did not include your taxpayer identification number. Please complete and submit NRC Form 531, Request for Taxpayer Identification Number, located at the following link: <http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>
Follow the instructions on the form for submission.

The following administrative omissions have been identified:

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Region I
U. S. Nuclear Regulatory Commission
Division of Nuclear Materials Safety
2100 Renaissance Boulevard, Suite 100
King of Prussia, PA 19406-2713
(610) 337-5260, (610) 337-5313,
(610) 337-5398, or (610) 337-5239