



September 27, 2018

United States Nuclear Regulatory Commission Region III, Materials Licensing Section 2443 Warrenville Road, Ste 210 Lisle, IL 60532-4352

Close-out request for License. Eastpointe Radiologists, # 21-24311-02

A close-out survey was performed on September 10, 2018 of our facility Eastpointe Radiologists., 21003 Mack Avenue, Grosse Pointe Woods, MI 48236-1333 and is enclosed for your review.

The last date of receipt and use of radioactive material at this site was August 29, 2018. Tc-99m and I-123 were the only radionuclides used at this facility. A current copy of the leak test results for the sealed sources used at our facility is enclosed for your review. Also the documentation for the transfer of our sealed sources to another license is enclosed. This facility is exempt from decommissioning due to low volumes and short physical half-lives of materials used.

We did not dispose of radioactive materials by release to sewers or incineration. In addition, there were no spills of any long-lived radioactive material or on-site burials of radioactive materials.

Cordially,

1. Dasm

Raymond Rahi, M.D. Eastpointe Radiologists, P.C. President and R.S.O.

the for the sea here of the	U.S. NUCLEAR REGULATORY COMMISSION	APPROVED BY OMB: NO. 3150-002	8 EXPIRES: 02/29/2020			
(02-2017) 10 CFR 30.36(j)(1); 40.42(j)(1); 70.38(j)(1); and 72.54(k)(5)(1)(1)	IFICATE OF DISPOSITION OF MATERIALS	Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the FOIA, Privacy, and Information Collections Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by e-mail to Infocotlects Resource@inc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.				
LICENSEE NAME AND ADDRESS		LICENSE NUMBER	DOCKET NUMBER			
Eastpointe Radiologists 21003 Mack Ave.		21-24311-02				
Grosse Pointe Woods, MI 48236-1333		LICENSE EXPIRATION DATE				
		2/2	28/21			
	A. LICENSE STATUS (Check the	e appropriate box)				
This license has expired.	his license has not yet expired; please	e terminate it.				
(Check the appropriate	B. DISPOSAL OF RADIOACT boxes and complete as necessary. If a		le attachments)			
The licensee, or any individual executing						
	e ever been procured or possessed by		•			
2. All activities authorized by thi under this license number cite	s license have ceased, and all radioad ed above have been disposed of in the terials to the licensee listed below:	tive materials procured and/or				
	Moross Rd., Detroit, MI 48236					
b. Disposal of radioactive mat						
1. Directly by the licen						
2. By licensed disposal site:						
3. By waste contractor:						
c. All radioactive materials ha Part 20, Subpart E, and is	ive been removed such that any remain ALARA.	ning residual radioactivity is wit	hin the limits of 10 CFR			
· · · · · · · · · · · · · · · · · · ·	C. SURVEYS PERFORMED A					
	ted by the licensee. The survey confir	ms:				
a. the absence of licensed radioactive materials						
b. that any remaining residua	I radioactivity is within the limits of 10	CFR 20, Subpart E, and is ALA	RA.			
2. A copy of the radiation survey results:						
✓ a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on:						
3. A radiation survey is not required as only sealed sources were ever possessed under this license, and						
a. The results of the latest leak test are attached; and/or b. No leaking sources have ever been identified.						
The person to be contacted regarding the information provided on this form:           NAME         TELEPHONE (Include Area Code)         E-MAIL ADDRESS						
	nitle Diagnostic Manager					
		313-881-0411 rb	d@eprpc.com			
Mail all future correspondence regarding this license to: Eastpointe Radiologists, 21003 Mack Ave., Grosse Pointe Woods, MI 48236-1333						
C. CERTIFYING OFFICIAL I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT						
PRINTED NAME AND TITLE SIGNATURE DATE						
Raymond Rahi, M.D. 9127/18						
WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.						

### Date Performed: September 10, 2018

Performed By: Laura Luna, Physics Consultant

Wipe tests analyzed with Capintec 15W. Efficiency for Cobalt-57 is 1.2 dpm/cpm. MDA is 33 dpm. A generic conversion factor of 2 dpm/cpm was used. This efficiency was performed March 23, 2018.

Survey performed with Ludlum survey meter model 14C with end window probe S/N # 167300, calibrated on 3/30/18 with a check source reading of 4.0 mR/hr. Range used was 0.1 mR/hr.

All sealed radioactive sources were transferred to St. John Hospital, 22101 Moross Rd., Detroit, MI 48236. Documentation of receipt is enclosed.

### Visual Check

The area was checked to ensure that all radioactive waste had been removed. No radioactive material was located.


### Radiation Level Survey

No area within the department demonstrated radiation levels in excess of the background reading of 0.02 mR/hr.

Ryl Relie m

		License #	21-24311-02	
		0/40/2048		
a siy	concepts to a second or a concept and a second second of the standard second second	9/10/2018	Performed By:	
10 a 107 7		Capintec 15W	MDA: Check Value:	the complete second state and the production of an end of the second state of the second state of the
u e coĝo	Survey Meter:		Calibration Date:	e de la complete menter par completente d'al contra de la completente de la completente de la completente de la
	Serial #:		Calibration Date:	
T		Wipes	-	Surveys
		dpm	-	<u>mR/h</u> 0.02
$\frac{1}{2}$	Hot Lab Counter	0		0.02
$\frac{2}{2}$	RA Storage Area	0		0.02
3 4	Injection Area Hot Lab Floor	0	-	0.02
4 5	Hot Lab Floor	0		0.02
5 6	Hot Lab Floor	0		0.02
7	Imaging Rm	0		0.02
8	Imaging Rm	0		0.02
9	Imaging Rm	0		0.02
10	Imaging Rm	0		0.02
11	Imaging Rm	0		0.02
12	Imaging Rm	0		0.02
13	Imaging Rm	0		0.02
14	Imaging Rm	0	-	0.02
15	Imaging Rm	0		0.02
6	Imaging Rm	0	-	0.02
17	Imaging Rm	0		0.02
18	Imaging Rm	0		0.02
9	Imaging Rm	0		0.02
20	Imaging Rm	0		0.02
21	Imaging Rm	0		0.02
22	Imaging Rm	0		0.02
23	Imaging Rm	0		0.02
24	Imaging Rm	0		0.02
25	Imaging Rm	0		0.02
26	Imaging Rm	0		0.02
27	Imaging Rm	0		0.02
28	Imaging Rm	0		0.02
29	Imaging Rm	0		0.02
30	Imaging Rm	0		0.02
1	Imaging Rm	0		0.02
32	Imaging Rm	0		0.02
<u>.</u>				
-				······································
_	Background			
	Dackyrounu			
	ACTION LEVELS:	200 dom	Efficiency (dpm/cpm):	2.0
	AVIIVIA LE VELS.	> 0.2 mR/hr	Enciency (aphicpilly.	<ul> <li>An annual company of the contract of the contract</li></ul>
	COMMENTS. No o	vidence of removabl	e contamination	, where any state of a second state of the state of the second sta
	COMMENTS: NO E			n an
	n an an ann an Araba an Araba an an ann an anns an anns an anns an an anns an an anns an an an an an an an an a		ра ] на сворудат с на селедат с средство и салосто с салосто с на селедат с с 	a a constant a faith an
				9/27/18

## Numbered Diagram



.

#### August 29, 2018

Nine sources were added to the inventory at St. John Hospital #21-03210-01, 22101 Moross Rd, Detroit, MI 48236 removed from Eastpointe Radiologists, 21003 Mack Ave, Grosse Pointe, MI 48236 . The reasoning for transfer was the site was closed by management, and the sites are affiliated. The following sources were inventoried, and leak tested on site prior to transfer from East Pointe Radiologists. Then upon arrival at St John Hospital, the sources were inventoried and leak tested prior to placement in our inventory and into storage at our basement. At both testing points, the sources were noted to be NOT leaking – 0dpm, all sources.

		CALIBRATION	CALIBRAT	ION	
NUCLIDE	TYPE	ACTIVITY	DATE	MFGR/MN	SN
Ba-133	Vial	271.3uCi	11/1/2000	IPL	743-3-8
Cs-137	Rod	0.5uCi	11/1/2000	GF-0008	725-42-6
Cs-137	Vial	216uCi	11/1/2000	IPL	743-1-22
Co-57	Spots x2	50uCi	2/1/2013	IPL	1615-84-4
Co-57	Spots x2	50uCi	10/1/2007	IPL	12-66-15-4/5
Co-57	Spots x2	50uCi	4/1/2010	IPL	1407-48/1407-57-20

The sources listed above were leak tested and were found to have less than 0.005  $\mu$ Ci removable contaminations. The following are the Minimum Detectable Activity (MDA) a three-minute assay. The leak test was done using a Biodex Atomiab 950 multichannel analyzer and well counter.

	Nuclide	MDA Background		
	Cs-137	1.1E-4 μCi	75 cpm	
	Co-57	1.5E-5 μCi	47 cpm	
	Ba-133	7.9E-5 μCi	174 cpm	
Analyzed by:	LT	Speer Smith	August 29, 2	2018
			Date	GA
RADIATION S	SAFETY OF	FICER: XIIII	N. Alla	Tel
/ Sincerely,			V	$\mathbf{X}$
Uncerety,		$\lambda / \lambda$	1	
		1 An IX		
Xana	Apri.	TH-		
Laura T. Speer S	Smith, MS			
St. John Hoalth	System			

St. John Health System Radiation Safety Officer



# Sealed Source Leak Test

Licensee: Eastpointe Radiologists Date: 06/07/18						
				Pe	rformed b	oy: <u>Laura Luna</u>
Calibration Calibration						
Nuclide	Туре	Activity	Date	Location	M/N	S/N
Ba-133	Vial	271.3	11/01/00	Hot Lab	IPL	743-3-8
Current Activity: 85.111						
Cs-137	Vial	216.3 uCi	11/01/00	Hot Lab	IPL	743-1-22
Current Activity: 144.02uCi						

Comment: The sources listed above were leak tested using a dry wipe technique and were found to have less than 0.005 uCi removable activity. The leak test wipes were analyzed using instrumentation capable of detecting 185 Bq (0.005 uCi) radioactivity on the wipe.

RADIATION SAFETY OFFICER:

NRLS m

Sealed Source Leak Test Page 1

