

Brachytherapy
Inspection Checklist

Attachment 2.6-6
Vermont Department of Health
Radioactive Materials Program Supervisory Accompaniment
Inspection Checklist
Brachytherapy

Licensee:

License No.:

Licensee Contact:

Inspection Date:

Telephone Number:

Email:

Last Inspection Date:

Priority:

Inspection Objectives:

- To determine if licensed activities are being conducted in a manner that will protect the health and safety of workers, the general public, and patients.
- To determine if licensed activities are being conducted in accordance with Vermont Department of Health regulations.

Focus Elements:

1. Security and Control of Licensed Materials
2. Shielding of Licensed Materials
3. Comprehensive Safety Measures
4. Radiation Dosimetry Programs
5. Radiation Instrumentation and Surveys
6. Radiation Safety Training and Practices
7. Management Oversight and Program Scope
8. Licensed Activities Performed by Contracted Personnel
9. Other Medical Uses of Byproduct Material or Radiation from Byproduct Material

Inspection Site Address (authorized use or storage):

Type of Inspection:

- Announced Unannounced
 Initial Routine
 Other:

**Date of Last
Inspection:**

Amendments and Significant Program Changes (Review from last license renewal)

Amendment #:

Date:

Amendment Item(s):

Note:

Program Inspection History

Is this an initial inspection?

Yes No

Brachytherapy
Inspection Checklist

Scope of Licensee Program	
Locations where licensed materials are being used, possessed, and stored are as described on the license. [L/C, 10 CFR 30.32]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Has the mailing address changed? [10 CFR 30.32]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Has the company ownership changed? [10 CFR 30.34]	<input type="checkbox"/> Yes <input type="checkbox"/> No
If yes, was the Department notified? [10 CFR 30.34]	<input type="checkbox"/> Yes <input type="checkbox"/> No
List location(s) of licensed material and identify the location of this inspection.	<input type="checkbox"/> Yes <input type="checkbox"/> No
Are all authorized users (AUs) listed on the license? [L/C, 10 CFR 35.14]	<input type="checkbox"/> Yes <input type="checkbox"/> No
If no, was the Department notified of changes to the AU list?	<input type="checkbox"/> Yes <input type="checkbox"/> No
Do new AUs meet Department training requirements? [10 CFR 35.57]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Personnel interviewed at licensee address during the inspection (**indicates those individuals in attendance at exit meeting.)	
Individual 1: Individual 2: Individual 3: Individual 4:	
Describe the licensed material program (types and quantities of licensed materials received, transferred, or redistributed; number of facilities/customers served; size of staff; etc.):	
Management Oversight	
Management supports ALARA. [10 CFR 20.1101]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Management supports RSO efforts.	<input type="checkbox"/> Yes <input type="checkbox"/> No
Are radiation protection annual audits being performed? [10 CFR 20.1101]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Who conducts audits? _____	
Scope of audit (areas of the program licensee reviewed) [10 CFR 20.1101]:	<input type="checkbox"/> Yes <input type="checkbox"/> No

Brachytherapy
Inspection Checklist

<p>Audits are conducted at intervals not exceeding 12 months. [10 CFR 20.1101]</p> <p>Audits and review records of the licensee program are being maintained. [10 CFR 20.2102] Note: These records must be kept for three years after they are made.</p> <p>Were deficiencies found in the program following a self-audit? If yes, have the deficiencies been corrected? Note: The inspector should look for repeat deficiencies.</p> <p>Audit records were reviewed by Department inspector.</p> <p>Performance Evaluation Factors (PEF) Note: PEF evaluations are best accomplished by interviewing management, RSO, ANPs, AUs, and other licensee personnel.</p> <p>Senior management is involved with radiation safety program and RSO oversight.</p> <p>The RSO has sufficient time to perform his/her radiation safety duties.</p> <p>The licensee has sufficient staffing to support its activities and radiation protection programs.</p> <p>Adequate audits are being implemented.</p>	<p><input type="checkbox"/> Yes <input type="checkbox"/> No</p>
Liquid Brachytherapy, GliaSite Therapy	
Inspection Procedures Used: 87132, 87134, 83822	
Inspection Site Address:	
<p>Location authorized on license?</p> <p>Use authorized on license?</p> <p> Manufacturer: _____</p> <p>Are authorized users approved on the license or by the RSC? [L/C] List the authorized users.</p> <p>The AU has specific vendor training in the use of the Proxima Therapeutics' GliaSite RTS. [L/C]</p>	<p><input type="checkbox"/> Yes <input type="checkbox"/> No</p>

Brachytherapy
Inspection Checklist

<p>Written Directive:</p> <p>Prior to implantation the written directive includes the treatment site, the radionuclide (including the chemical and physical form), and dose. [10 CFR 35.40]</p> <p>After implantation but prior to completion of the procedure, the written directive includes the radionuclide (including the chemical and physical form), the treatment site, and the total dose. [10 CFR 35.40]</p> <p>Procedures are in place to confirm that the balloon does not leak prior to injection of the radionuclide and implantation in the patient or human research subject. [L/C]</p> <p>(1) Has the licensee had a leaking source?</p> <p>(2) If yes, was the Department notified within five working days?</p> <p>The licensee labels syringes containing I-125 Iotrex™ and syringe radiation shields with the radionuclide and therapeutic procedure (i.e. I-125 Iotrex™ for brain brachytherapy.)</p> <p>Have any medical events occurred associated with the liquid brachytherapy program (e.g., mislabeled syringes, color code syringe/vial errors, picking up the wrong syringe, etc.)?</p> <p>The licensee acquired an SSD certificate prior to medical use of the liquid source.</p> <p>Is the licensee following the SSD limitations/physical conditions of use?</p> <p>If no, was a safety evaluation performed by the broad scope medical use licensee that addressed the conditions of use? [Note: A safety evaluation must be performed by a broad scope medical use licensee if the physical conditions of use, as stated in the SSD certificate, are exceeded.] [L/C]</p> <p>Did the licensee perform a survey after implanting the source? [10 CFR 35.404]</p> <p>Did the licensee perform a survey after removing the source? [10 CFR 35.404]</p>	<p><input type="checkbox"/> Yes <input type="checkbox"/> No</p>
Microsphere Brachytherapy Sources and Devices	
Inspection Site Address:	
Location authorized on license?	<input type="checkbox"/> Yes <input type="checkbox"/> No

Brachytherapy
Inspection Checklist

Use authorized on license? Manufacturer: _____	<input type="checkbox"/> Yes <input type="checkbox"/> No
Are authorized users approved on the license or by the RSC? [L/C] List the authorized users.	<input type="checkbox"/> Yes <input type="checkbox"/> No
The AU has specific vendor training in the use of microspheres and the microsphere delivery system [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
All requirements for brachytherapy sources and manual brachytherapy use are being met. [Note: Leak tests are not required because the activity per microsphere (the sealed source) meets the criteria in 10 CFR 35.67.] [10 CFR 35, Subpart F]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Written Directive: (1) Prior to implantation the written directive includes the treatment site, the radionuclide including the chemical and physical form (i.e., Y-90 microspheres), and dose. [Note: For Y-90 microspheres, "prescribed dose" in the written directive means the total dose.] [10 CFR 35.40]	<input type="checkbox"/> Yes <input type="checkbox"/> No
(2) After implantation but prior to completion of the procedure, the written directive includes the radionuclide (including the chemical and physical form), the treatment site, and the total dose. [10 CFR 35.40]	<input type="checkbox"/> Yes <input type="checkbox"/> No
(3) Is the licensee following the written directive procedures that describe how to quantify the total dose to the treatment site as well as the total dose to other sites upon completion of the administration? [10 CFR 35.41, L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Are quarterly inventories of sealed sources and brachytherapy sources performed?	<input type="checkbox"/> Yes <input type="checkbox"/> No
Do the inventories include the individual aggregates of the microspheres identifying the radioisotope, the container of the aggregate, the total activity of the aggregate, and the container location? [10 CFR 35.2406, L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Are procedures in place to ensure that the bremsstrahlung emissions from each patient permit release? [Note: TEDE < 5 mSv (500 mrem) to others and TEDE < 1 mSv (100 mrem) to breast-feeding infant or child.] [10 CFR 35.75]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Y-90 microspheres placed in vials, syringes, or radiation shields are labeled by the manufacturer or by the licensee	<input type="checkbox"/> Yes <input type="checkbox"/> No

Brachytherapy
Inspection Checklist

<p>The administration syringes and syringe radiation shields are labeled with the radioisotope, form, and therapeutic procedure (i.e., Y-90 microspheres, brachytherapy)</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>The required SSD certificate for the Y-90 microsphere delivery system is available for review.</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>(1) Is the licensee following the SSD limitations/physical conditions of use?</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>(2) If no, was a safety evaluation performed by the broad scope medical use licensee that addressed the conditions of use? [Note: A safety evaluation must be performed by a broad scope medical use licensee if the physical conditions of use, as stated in the SSD certificate, are exceeded.] [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No

Note: The U.S. Food and Drug Administration currently approves the MDS NORDION Y-90 TheraSphere® microspheres under the provisions of a "Humanitarian Device Exemption" (HDE No H9800006). An Institutional Review Board review and approval is required before a humanitarian use device is used at a facility as well as continuing review of its use.

Intervascular Brachytherapy (IVB)	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
--	--

Inspection Site Address:

<p>Location authorized on license?</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Use authorized on license?</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Manufacturer: _____</p>	
<p>Are authorized users and authorized medical physicist approved on the license or by the RSC? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>List the authorized users and authorized medical physicist.</p>	
<p>Have the Authorized User, Interventional Cardiologist/Physician, and Authorized Medical Physicist received the vendor training for use of the device?</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Does the Authorized User consult with the Interventional Cardiologist/Physician and Authorized Medical Physicist (AMP) prior to initiating treatment?</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Are all procedures conducted in the physical presence of the Authorized User or the AMP? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No

Brachytherapy
Inspection Checklist

<p>Are written directives signed and dated prior to all treatments, specifying treatment site, the radionuclide, and dose? [10 CFR 35.40]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>The AMP performs independent measurement of the source output prior to the first patient treatment using a dosimetry system that meets the requirements of 10 CFR 35.630.</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Has the licensee developed, implemented, and maintained written emergency procedures for both stuck and detached sources, including the provisions for appropriate emergency response equipment and appropriate surgical procedures? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Does the licensee survey the patient and the IVB treatment catheter immediately following source retraction or removal confirming complete retraction of the source(s)? [10 CFR 35.404]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Are source trains removed from use after the “use by” date? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Does the licensee perform “source stepping?” [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Is the licensee properly using portable shields? [L/C]</p>	<input type="checkbox"/> N/A <input type="checkbox"/> Yes <input type="checkbox"/> No
<p>In a three-ribbon set containing 6, 10, or 14 seeds per ribbon, no single seed exceeds 35 millicuries, or a total of 1.1 curies per set.</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
Novoste Beta-Cath™ IVB System	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Manufacturer: _____</p>	
<p>Is the licensee authorized for Strontium-90 sealed sources, 5 mCi per source and a total of 800 mCi? [Note: The exposure rate from this system is greater than 1,200 rads/hr.] [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Are Authorized Users and Authorized Medical Physicist approved on the license or by the RSC? [L/C] List the authorized users and authorized medical physicist.</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Have the Authorized User, Interventional Cardiologist/Physician, and Authorized Medical Physicist received the vendor training for use of the device?</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Does the Authorized User consult with the Interventional Cardiologist/Physician and Authorized Medical Physicist (AMP) prior to initiating treatment?</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Are all procedures conducted in the physical presence of the Authorized User or the AMP? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No

Brachytherapy
Inspection Checklist

<p>Are written directives signed and dated prior to all treatments, specifying treatment site, the radionuclide, and dose? [10 CFR 35.40]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Does the licensee perform “source stepping?” [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A
<p style="padding-left: 40px;">If yes, does the licensee have written procedures for “source stepping?” [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Does the licensee survey the patient and the IVB treatment catheter immediately following source retraction or removal confirming complete retraction of the source(s)? [10 CFR 35.404]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Is an introducer sheath used, unless contraindicated? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Is a dual syringe system used? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Has the licensee developed, implemented, and maintained written emergency procedures for both stuck and detached sources, including the provisions for appropriate emergency response equipment and appropriate surgical procedures? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Is the source storage container locked in a secured location? [Note: Because Sr-90 is a pure beta-emitter, shielding calculations are not necessary for areas outside of the treatment room and device storage areas.]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Has the AMP performed an independent measurement of source output prior to first treatment? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Were the independent measurements performed with a dosimetry system that meets the requirements of 10 CFR 35.630? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Is the IVB system inspected and serviced at intervals recommended by the manufacturer? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Is maintenance and repair performed only by the manufacturer or persons specifically licensed by the Department, NRC, or an Agreement State? [L/C]</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No
<p>Guidant Galileo™ IVB System [Note: Guidant Corporation VI Model Galileo™ intravascular brachytherapy high dose rate afterloader devices.]</p>	
<p>Are authorized users and authorized medical physicist approved on the license or by the RSC? [L/C]</p> <p style="padding-left: 40px;">List the authorized users and authorized medical physicist.</p>	<input type="checkbox"/> Yes <input type="checkbox"/> No

Brachytherapy
Inspection Checklist

Have the Authorized User, Interventional Cardiologist/Physician, and Authorized Medical Physicist received the vendor training for use of the device?	<input type="checkbox"/> Yes <input type="checkbox"/> No
Does the Authorized User consult with the Interventional Cardiologist/Physician and Authorized Medical Physicist (AMP) prior to initiating treatment?	<input type="checkbox"/> Yes <input type="checkbox"/> No
Are all procedures conducted in the physical presence of the Authorized User or the AMP? [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Are written directives signed and dated prior to all treatments, specifying treatment site, the radionuclide, and dose? [10 CFR 35.40]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Does the licensee perform “source stepping” or “pullback” procedures? [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
If yes, does the licensee have written procedures? [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Does the licensee survey the patient and the IVB treatment catheter immediately following source retraction or removal confirming complete retraction of the source(s)? [10 CFR 35.404]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Has the licensee developed, implemented, and maintained written emergency procedures for both stuck and detached sources, including the provisions for appropriate emergency response equipment and appropriate surgical procedures? [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Are the delivery device and source assembly secured in locked storage when not in use? [Note: Because P-32 is a pure beta-emitter, shielding calculations are not necessary for areas outside of the treatment room and device storage areas.] [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Does the licensee have a method for key control for the console? [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Has the AMP performed an independent measurement of source output prior to first treatment? [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Were the independent measurements performed with a dosimetry system that meets the requirements of 10 CFR 35.630? [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
Is the IVB system inspected and serviced at intervals recommended by the manufacturer? [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No
The source assembly/cartridge are not used for more than 60 days or 650 cycles, whichever comes first, in accordance with the manufacturer’s guidance. [L/C]	<input type="checkbox"/> Yes <input type="checkbox"/> No

Brachytherapy
Inspection Checklist

<p>HDR units and storage devices are properly labeled.</p> <p>HDR treatment units, the console, and the console key are secured from unauthorized entry and removal when not in use.</p> <p>HDR unit keys are controlled.</p> <p>Operating and emergency procedures are maintained at the treatment console. [10 CFR 35.610]</p> <p>Licensee implements actions to verify that:</p> <p style="padding-left: 40px;">Prior to treatment, the treatment plan is in accordance with the written directive.</p> <p style="padding-left: 40px;">Prior to treatment, the treatment parameters (source positioning, HDR unit settings, applicator type and size, etc.) are in accordance with the written directive.</p>	<p><input type="checkbox"/> Yes <input type="checkbox"/> No</p>
Security Inspection	
<p>On-site Security Review: If this licensee is authorized for possession of material equal to or exceeding the Category 2 threshold, complete an on-site security review per 10 CFR 37.43.</p> <p>If yes,</p> <p>Licensee Contact Name:</p> <p>Contact Telephone Number:</p> <p>Contact Email Address:</p>	<p><input type="checkbox"/> Yes <input type="checkbox"/> No</p> <p style="padding-left: 100px;"><input type="checkbox"/> N/A</p>
Summary of Observations, Findings, and Conclusions	
<p>Business Operations:</p> <p>Facility – Visit all storage and use locations identified on the application.</p> <p>Radiation Safety Operations.</p> <p>Personnel.</p> <p>Overall Assessment</p> <p>Note: If there is not sufficient information to conclude that licensed material will be used as specified on the license, immediately notify Department supervision.</p>	
Exit Meeting at Conclusion of Inspection	
<p>Identify and list the individuals in attendance:</p>	

Brachytherapy
Inspection Checklist

Date Meeting Conducted:
List those issues discussed at the exit meeting:
Summary of Violations and Recommendations