



UNITED STATES  
NUCLEAR REGULATORY COMMISSION  
REGION IV  
1600 E. LAMAR BLVD  
ARLINGTON TX 76011-4511

November 19, 2018

Ms. Robin L. Johnson, MHA  
Barrett Hospital & Healthcare  
600 Highway 91 South  
Dillon, MT 59725

SUBJECT: TERMINATION OF NRC RADIOACTIVE MATERIALS LICENSE NO. 25-29088-01

By letter dated September 25, 2018, you informed the U.S. Nuclear Regulatory Commission (NRC) that you wished to terminate your NRC radioactive materials license. The NRC Form 314 dated July 10, 2016, stated that all radioactive material had been transferred to Front Range Nuclear Services (49-27531-01) and Thomas Gray and Associates (CA 2105-30). Transfer records were reviewed during a routine inspection conducted on May 23, 2018.

The supporting documentation to the NRC Form 314 also provided a copy of the latest leak test results. The supporting documentation to the NRC Form 314 stated that the licensee only possessed radioactive materials with radioactive decay half-lives of less than 120 days. A review of the historical file supported that only short lived radioactive materials were possessed by the licensee. The NRC staff has reviewed the supporting documentation to your NRC Form 314. Based on the documentation submitted and a review of your historical file, the staff concluded that all licensable radioactive material has been removed from the facility and the facility located at 600 Highway 91, South, Dillon Montana, 59725, is suitable for unrestricted use in accordance with 10 CFR 20.1402, "Radiological Criteria for Unrestricted Use." Based on these conclusions, no further remediation or actions with respect to NRC regulated material is required.

Please find enclosed Amendment No. 57 to License No. 25-29088-01 authorizing your license termination. An environmental assessment for this action is not required since this action is categorically excluded under 10 CFR 51.22(c)(20)(i).

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

R. Johnson

- 2 -

You can contact me at 817-200-1590 if you have questions or require clarification on any of the information stated above.

Sincerely,

**/RA/**

Michelle Simmons, Health Physicist  
Materials Licensing and Inspection Branch

Docket: 030-33800  
License: 25-29088-01  
Control: 610078

Enclosure: As stated

- 2 -