

**From:** [Lanzisera, Penny](#)  
**To:** [k.ostrom@ampmedicalphysics.com](mailto:k.ostrom@ampmedicalphysics.com)  
**Subject:** Request for Additional Information for Dr. Terry Termination  
**Date:** Friday, December 14, 2018 12:43:00 PM

---

Licensee: Richard F. Terry, M.D., Inc.  
License No. 47-25637-01  
Docket No. 03036406  
Mail Control 610657

Dear Mr. Ostrom, in order to complete our review of the termination request submitted by Dr. Terry, please submit the following additional information pursuant to 10 CFR Part 30:

1. The NRC Form 314 form submitted indicates that surveys are attached; however they were omitted. Please submit radiation surveys to show the areas of use are free of contamination. Alternatively, you may indicate the last date of use of Technetium-99m if more than a month ago.
2. The NRC Form 314 form submitted indicates that sources were shipped to Eckert and Ziegler. At the time of NRC's inspection in 2015, the facility possessed 6 sources – 3 dose calibrator source vials; 1 flood source; 1 rod source; and 1 survey instrument check source. Please submit documentation showing the confirmation of receipt of these sources by Eckert and Ziegler.
3. Please submit leak test records for the above sources, unless exempt (e.g., rod source and survey instrument check source).
4. Please indicate if the address indicated in the NRC Form 314 to mail all correspondence is a private address or public address.

You may submit the above information either via pdf to my attention or via fax to 610-337-5269. Please refer to Mail Control No. 610657 in your reply.

Thank you for your assistance.

Penny Lanzisera  
Senior HP, US NRC Region I