

San Cristobal Oncology Services, Inc.

Date: February 20, 2019

License No. 52-35511-01
Docket No. 030-39138
Control No. 610361 *my*

Farrah C. Gaskings
Health Physicist
U.S Nuclear Regulatory Commission
Region I, 2100 Renaissance Blvd, Suite 100
King of Prussia, PA 19406-2713

SUBJECT: Additional information Requested for additional information for license
No. 52-35511-01

Dear Mr. Gaskings:

This letter is in reference to the additional information required of your email dated January 31, 2019.

We have taken the decision not to continue with part 10CFR 35.600 in our license application for HDR. This is a temporary action until we get an authorized user that meets these requirements. We will continue with the parts 10 CFR 35.100, 10 CFR 35.200 and 10 CFR 35.400. For this reason, we only answer the question two (2) of your request.

2. In our request for additional information email dated November 7, 2018, we asked that you confirm the sealed source manufacturer and model numbers requested for materials to be used under 10 CFR 35.400 in your original application dated October 29, 2018 because we could not locate a sealed source and device certificate for the AEA Technology QSA Incorporate (formerly Amersham Corporation) Model CDC-T1 source. In your response dated December 28, 2018, you stated "In reference to the question, do not consider the information submitted previously, it must be replaced by the following; we will use the Theragenics I-Seed Model AgX100." Please confirm if this means you do not wish to pursue any of the sealed source manufacturer and model numbers requested in the October 2018 application or just the AEA Technology QSA Incorporate (formerly Amersham Corporation) Model CDC-T1 for 10 CFR 35.400 which was in question.


We are confirming that we will use the Theragenics I-Seed Model AgX100.

Rec'd. in LAT-03/04/2019

611044
RESEARCH MATERIALS

Note: Amersham Corporation was absorbed by Theragenics Company. If in the future, we want to add other source models we will notify you.

Sincerely,

A handwritten signature in black ink, appearing to read 'Pedro L. Benetti Loyola', written over a horizontal line.

Pedro L. Benetti Loyola MHSA
Senior Executive and Vice President
Hospital San Cristobal
Carr. PR 506 Km 1.0
Coto Laurel, PR 00780