

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

<p>1. LICENSEE/LOCATION INSPECTED:</p> <p>N. Gupta, M.D., P.C. 1660 Kingway Court Trenton, MI</p> <p>REPORT NUMBER(S) <b>2019001</b></p>	<p>2. NRC/REGIONAL OFFICE</p> <p>Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352</p>	
<p>3. DOCKET NUMBER(S)</p> <p>030-34742</p>	<p>4. LICENSE NUMBER(S)</p> <p>21-32096-01</p>	<p>5. DATE(S) OF INSPECTION</p> <p>3/1/19, with in-office review through 3/5/19</p>

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert G. Gattone, Jr.	<i>Robert G. Gattone, Jr.</i>	3/14/19
BRANCH CHIEF	Aaron T. McCraw	<i>[Signature]</i>	3/19/19

**Docket File Information**  
**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:  N. Gupta, M.D., P.C. 1660 Kingway Court Trenton, MI  REPORT NUMBER(S) <b>2019001</b>	2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
-----------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------------------------------------------------------------------------------------------------------------

3. DOCKET NUMBER(S)  030-34742	4. LICENSE NUMBER(S)  21-32096-01	5. DATE(S) OF INSPECTION  3/1/19, with in-office review through 3/5/19
--------------------------------------	-----------------------------------------	---------------------------------------------------------------------------------

6. INSPECTION PROCEDURES USED  87130	7. INSPECTION FOCUS AREAS  03.01 through 03.07
--------------------------------------------	------------------------------------------------------

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)  02201	2. PRIORITY  5	3. LICENSEE CONTACT  Jack Juni, M.D., RSO	4. TELEPHONE NUMBER  (248) 640-9849
---------------------------------	----------------------	-------------------------------------------------	-------------------------------------------

Main Office Inspection      Next Inspection Date: 03/01/2024

Field Office Inspection \_\_\_\_\_

Temporary Job Site Inspection \_\_\_\_\_

**PROGRAM SCOPE**

This was an unannounced, routine inspection of the licensee's radiation protection program. The in-office review involved receiving information that was unavailable during the onsite inspection. The licensee's work hours begin at 7:30 am on Mondays, Wednesdays, Thursdays, and Fridays. The licensee had one nuclear medicine technologist (NMT). The RSO was also the authorized user, as required. The licensee conducted technetium-99m Sestamibi scans.

**Performance Observations**

The inspector: (1) observed the the NMT conduct a survey meter constancy check; (2) observed that the licensee had two calibrated survey meters; (3) observed the NMT conduct ambient exposure rate and contamination surveys for received packages containing licensed material; (4) reviewed selected sealed source inventory records; (5) reviewed selected sealed source leak test records; (6) noted that the facility was as per the diagram in Condition 14 of the license; (7) reviewed selected radiation protection program audit records; (8) reviewed dosimetry records for 2014 through 2018, and the doses for whole body and extremity were well below the annual limits; (9) observed the NMT administer technetium-99m Sestamibi to a patient while using a syringe shield, lab coat, gloves, and whole body and extremity dosimeter badges; (10) observed the NMT demonstrate how he would respond to a radioactive spill based on a scenario posed by the inspector; and (11) observed that licensed material was secured as required.

No violations of NRC regulatory requirements were identified as a result of this inspection.