

March 29, 2019

United States
Nuclear Regulatory Commission
Washington, DC 20555-0001

Licensee: Rutland Regional Medical Center
License No.: 44-14121-01
Control No.: 611575


SUBJECT: Submission of Additional Requested Information for License Amendment

To Whom It May Concern:

Attached you will find the requested documentation of the closeout survey documents for the two locations of Contracted Mobile Unit and the prior Hot Lab location.

In addition, the amendment to the spelling of Claudio Fort's last name, please.

Sincerely yours,



Claudio Fort
President and CEO
Rutland Regional Medical Center

Cc: Shari Patch
Director, DI

/pb

Healthy You. Healthy Together.

611575

Rec'd. in LAT-04/02/2019



Rutland Regional Medical Center

Nuclear Medicine Department

160 Allen Street, Rutland, VT 05701

Lic.# 44-14121-01

February 15, 2019

Area Wipe Report

Group Name : CLOSE OUT SURVEYS ECAM

Date Range : 01/04/2019 To 01/04/2019

Date: 01-04-2019 15:13:00	Tech Name: PAUL WAITE	Background CPM: 280.00
Model Name: ATOM LAB 950	Next Calibration Date: 06-20-2019	Background Subtracted: YES
Serial Number: 4259608	Efficiency: 100.00%	Check Source : 184930*
Wipe Area: 100.00		Battery : N/A

Area Name	Gross CPM Value	Gross DPM Value	Net CPM Value	Net DPM Value	Trigger Value
ECAM TABLE	280	280	0	0	2000
ECAM HEAD 1	280	280	0	0	2000
ECAM HEAD 2	280	280	0	0	2000
ECAM FLOOR	280	280	0	0	2000
ECAM COUNTER	280	280	0	0	2000
PROBE/WELL COUNTER ROOM	280	280	0	0	2000

Note:

Areas are over trigger limit.

Test Not performed or Not Required: " _ "

Check source reading was done today but not as part of this survey/wipe test: " * "

Default Trigger Limit : 2000.00 DPM


Technologist Signature: - _____

Rutland Regional Medical Center

Nuclear Medicine Department

Area Wipe Report : Group Name : CLOSE OUT SURVEYS ECAM

Date Range : 01/04/2019 To 01/04/2019

R.S.O. Signature: - 

Report Printed: 2/15/2019 3:17:16PM



Area Monitoring Report

Group Name : CLOSE OUT SURVEYS ECAM


Date Range : 01/04/2019 To 01/04/2019

Meter: LUDLUM 3	Efficiency: N/A	Ck. Src:	Battery Check:	Date/Time: 01-04-2019 15:11
Meter Model: Ludlum		Meter S/N: 29537	Background: mR/hr 0.02	Tech Name: PAUL WAITE
				Background Subtracted: NO

Area Name	Area Description	Net mR/hr	Trigger Value
ECAM TABLE		0.02	0.05
ECAM HEAD 1		0.02	0.05
ECAM HEAD 2		0.02	0.05
ECAM FLOOR		0.02	0.05
ECAM COUNTER		0.02	0.05
PROBE/WELL COUNTER ROOM		0.02	5.00

Note:
 Areas Underlined are over trigger limit.

QC Readings marked with '*' (asterisk)' indicates that QC was not part of the Area Monitor test.
 Test Not performed or Not Required : "--"
 Default Trigger Limit : 5.00 mR/hr

Technologist Signature: - 

R.S.O. Signature: - 

Report Printed: 2/15/2019 3:15:11PM



Rutland Regional Medical Center

Nuclear Medicine Department

160 Allen Street, Rutland, VT 05701

Lic.# 44-14121-01

March 28, 2019

Area Wipe Report

Group Name : RPMC OLD HOT LAB CLOSE OUT

Date Range : 03/28/2019 To 03/28/2019

Date/Time	Tech	In st	Batt	Ck.Src	Bkg	Area 1		Area 2		Area 3		Area 4		Area 5		Area 6	
						CPM	DPM	CPM	DPM	CPM	DPM	CPM	DPM	CPM	DPM	CPM	DPM
03-28-19 08:39	PW	4	--	180657*	310	0	0	0	0	0	0	0	0	0	0	0	0

Note:

Above all values listed are NET values.

Areas Underlined are over trigger limit.

Battery Check: "P" = Pass "F" = Fail

Test Not performed or Not Required: "-"

Check source reading was done today but not as part of this survey/wipe test: "*" "

Wipe Area: 100.00 Sq. CM

Default Trigger Limit : 2000.00 DPM

Current Area Trigger Limits (DPM):

DECAY CABINET	-	2,000.00	LEAD STORAGE	-	2,000.00	COUNTERS	-	2,000.00
CABINETS	-	2,000.00	TABLE	-	2,000.00	FLOOR	-	2,000.00

Instrument Details:

#	Instrument	Manufacturer	Next Cal Date	Serial No	Efficiency
4	ATOM LAB 950	Biodex	06-20-2019	4259608	100.00%

Original area names:

Generic Name	Actual area name	Generic Name	Actual area name	Generic Name	Actual area name
Area 1	FLOOR	Area 2	DECAY CABINET	Area 3	COUNTERS

Rutland Regional Medical Center

Nuclear Medicine Department

Area Wipe Report : Group Name : RRM C OLD HOT LAB CLOSE OUT

Date Range : 03/28/2019 To 03/28/2019


Area 4 LEAD STORAGE Area 5 CABINETS Area 6 TABLE

Tech Information:

Tech Tech Name

PW PAUL WAITE

Technologist Signature: - 

R.S.O. Signature: - 

Report Printed: 3/28/2019 8:40:07AM



Rutland Regional Medical Center

Nuclear Medicine Department

160 Allen Street, Rutland, VT 05701

Lic.# 44-14121-01

March 28, 2019

Area Monitoring Report

Group Name : RRMC OLD HOT LAB CLOSE OUT

Date Range : 03/28/2019 To 03/28/2019

Date/Time	In Tech	st	Batt	Ck.Src mR/hr	Bkg mR/hr	Area 1 mR/hr	Area 2 mR/hr	Area 3 mR/hr	Area 4 mR/hr	Area 5 mR/hr	Area 6 mR/hr
03-28-19 08:38	PW	1	P*	2.50*	0.02	0.02	0.02	0.02	0.02	0.02	0.02

Note:

Areas Underlined are over trigger limit:

QC Readings marked with '*' (asterisk) indicates that QC was not part of the Area Monitor test.

Battery Check: "P" = Pass "F" = Fail

Test Not performed or Not Required : "--"

Default Trigger Limit : 5.00 mR/hr

Current Area Trigger Limits (mR/hr):

FLOOR	-	5.00	DECAY CABINET	-	5.00	COUNTERS	-	5.00
LEAD STORAGE	-	5.00	CABINETS	-	5.00	TABLE	-	5.00

Instrument Information

#	Name	Manufacturer	Next Cal	Serial Number	Efficiency
1	LUDLUM 3	Ludlum	08-27-2019	29537	N/A

Original area names:

Generic Name	Actual Area Name	Generic Name	Actual Area Name	Generic Name	Actual Area Name
Area 1	FLOOR	Area 2	DECAY CABINET	Area 3	COUNTERS
Area 4	LEAD STORAGE	Area 5	CABINETS	Area 6	TABLE

Rutland Regional Medical Center

Nuclear Medicine Department


Area Monitoring Report : Group Name : RPMC OLD HOT LAB CLOSE OUT

Date Range : 03/28/2019 To 03/28/2019

Tech Information:

Tech Complete Tech Name

PW PAUL WAITE

Technologist Signature: - 

R.S.O. Signature: - 

Report Printed: 3/28/2019 8:39:59AM
