



Designer and Manufacturer
of
Scientific and Industrial
Instruments

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.

501 Oak Street
325-235-5494
Sweetwater, TX 79556, U.S.A.

10744 Dutchtown Road
865-392-4601
Knoxville, TN 37932, U.S.A.

CUSTOMER CABRERA SERVICES ORDER NO. 20280903/429686

Mfg. Thermo Model MICRO REM Serial No. 1844

Mfg. _____ Model _____ Serial No. _____

Cal. Date 22-Dec-15 Cal Due Date 22-Dec-16 Cal. Interval 1 Year Meterface 0-200µrem/

Check mark Applies to applicable instr. and/or detector IAW mfg. spec. T. 74 °F RH 21 % Alt 704.0 mm Hg

New Instrument Instrument Received Within Toler. +10% 10-20% Out of Tol. Requiring Repair Other-See comments

Mechanical ck. Meter Zeroed Background Subtract Input Sens. Linearity

F/S Resp. ck. Reset ck. Window Operation Geotropism

Audio ck. Alarm Setting ck. Batt. ck. (Min. Volt) _____ VDC

Calibrated in accordance with LMI SOP 14.8

Calibrated in accordance with LMI SOP 14.9

Instrument Volt Set _____ V Input Sens. _____ mV Det. Oper. _____ V at _____ mV Threshold Dial Ratio _____ = _____ mV

HV Readout (2 points) Ref./Inst. _____ / _____ V Ref./Inst. _____ / _____ V

COMMENTS:

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
x 1000	150 mR/hr	150	150
x 1000	50 mR/hr	50	50
x 100	15 mR/hr	150	150
x 100	5 mR/hr	50	50
x 10	1500 µR/hr	150	150
x 10	500 µR/hr	98	98
x 1	150 µR/hr	155	150
x 1	100 µR/hr	105	100
x0.1	15 µR/hr	150	150
x0.1			

*Uncertainty within ± 10% C.F. within ± 20%

Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	Log Scale	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
Digital Readout						

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. State of Texas Calibration License No. LO-1963

The calibration system conforms to the requirements of ANSI/NCCL Z540-1-1994 and ANSI N323-1978

Reference Instruments and/or Sources: Cs-137 S/N: 059 2171CP 2261CP 720 734 781 1131 1616 1696 1909 1916CP 5105 5717CO 5719CO 60646 70897 73410 E552 G112 M565 S-394 S-1054 T10081 T10082 Neutron Am-241 Be S/N: T-304 Ra-226 S/N: Y982

Alpha S/N _____ Beta S/N _____ Other Cs-137 201uCi

m 500 S/N _____ Oscilloscope S/N _____ Multimeter S/N _____

Calibrated By: [Signature] Date 22-Dec-15

Reviewed By: [Signature] Date 30-Dec-15

AC Inst. Passed Dielectric (Hi-Pot) and Continuity Test Only Failed: _____