



Designer and Manufacturer  
of  
Scientific and Industrial  
Instruments

**LUDLUM MEASUREMENTS, INC.**

501 Oak Street 10744 Dutchtown Road  
325-235-5494 865-392-4601  
Sweetwater, TX 79556, U.S.A. Knoxville, TN 37932, U.S.A.

**CERTIFICATE OF CALIBRATION**

CUSTOMER ERG ORDER NO. 20301042/443117

Mfg. Thermo Model \_\_\_\_\_ MICRO REM Serial No. 2079

Mfg. \_\_\_\_\_ Model \_\_\_\_\_ Serial No. \_\_\_\_\_

Cal. Date 29-Nov-16 Cal Due Date 29-Nov-17 Cal. Interval 1 Year Meterface N/A

Check mark  applies to applicable instr. and/or detector IAW mfg. spec. T. 73 °F RH 29 % Alt 698.0 mm Hg

- New Instrument
- Instrument Received
- Within Toler. +/-10%
- 10-20%
- Out of Tol.
- Requiring Repair
- Other-See comments
- Mechanical ck.
- Meter Zeroed
- Background Subtract
- Input Sens. Linearity
- F/S Resp. ck.
- Reset ck.
- Window Operation
- Geotropism
- Audio ck.
- Alarm Setting ck.
- Batt. ck. (Min. Volt) \_\_\_\_\_ VDC
- Calibrated in accordance with LMI SOP 14.9

Instrument Volt Set \_\_\_\_\_ V Input Sens. \_\_\_\_\_ mV Det. Oper. \_\_\_\_\_ V at \_\_\_\_\_ mV Threshold Dial Ratio \_\_\_\_\_ = \_\_\_\_\_ mV

HV Readout (2 points) Ref./Inst. \_\_\_\_\_ / \_\_\_\_\_ V Ref./Inst. \_\_\_\_\_ / \_\_\_\_\_ V

**COMMENTS:**

Gamma Calibration: GM detectors positioned perpendicular to source except for M 44-9 in which the front of probe faces source.

RANGE/MULTIPLIER	REFERENCE CAL. POINT	INSTRUMENT REC'D "AS FOUND READING"	INSTRUMENT METER READING*
X 1000	150 mR/hr	150	150
X 1000	50 mR/hr	45	45
X 100	15 mR/hr	150	150
X 100	5 mR/hr	45	45
X 10	1500 µR/hr	150	150
X 10	500 µR/hr	50	50
X 1	150 µR/hr	190	150
X 1	100 µR/hr	125	100
X 0.1	15 µR/hr	150	150
X 0.1			

\*Uncertainty within ± 10% C.F. within ± 20%

Range(s) Calibrated Electronically

REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*	Log Scale	REFERENCE CAL. POINT	INSTRUMENT RECEIVED	INSTRUMENT METER READING*
_____	_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____	_____
_____	_____	_____	_____	_____	_____	_____

Ludlum Measurements, Inc. certifies that the above instrument has been calibrated by standards traceable to the National Institute of Standards and Technology, or to the calibration facilities of other International Standards Organization members, or have been derived from accepted values of natural physical constants or have been derived by the ratio type of calibration techniques. The calibration system conforms to the requirements of ANSI/NCSS Z540-1-1994 and ANSI N323-1978 State of Texas Calibration License No. LO-1963

- Reference Instruments and/or Sources: Cs-137 S/N  059  2171CP  2261CP  720  734  781  1131  1616  1696  1909  1916CP  2324/2521
- 5717CO  5719CO  60646  70897  73410  E552  G112  2168CP  S-394  S-1054  T10081  T10082 Neutron Am-241 Be  T-304 Ra-226  Y982
- Alpha S/N \_\_\_\_\_  Beta S/N \_\_\_\_\_  Other Cs137#NES-356 GRN
- m 500 S/N \_\_\_\_\_  Oscilloscope S/N \_\_\_\_\_  Multimeter S/N \_\_\_\_\_

Calibrator James McSorley Title Calibrator Date 29NOV16

QC'd By [Signature] Title Service Dept. [Signature] Date 29NOV16

AC Inst.  Passed Dielectric (Hi-Pot) and Continuity Test  
Only  Failed: \_\_\_\_\_



# Certificate of Calibration


## Air Sampler Calibration Form

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGooffice.com

Sampler: Manufacturer: Eberline Model Number: RAS-1 Serial Number: 408667-8  
Warm up Time 30 min Temperature: 75 °F Relative Humidity 20 % Barometric Pressure: 24.6 in. Hg  
Correction Factor: 0.899227

Sampler Rotometer	Calibrator Flow Meter	Corrected Flow
10	10	8.992275
20	19	17.08532
30	29	26.0776
40	35	31.47296
50	46	41.36446
60	56	50.35674
70	64	57.55056

Calibration Chart



$$y = 8.1252x + 0.7708$$

$$R^2 = 0.9974$$

Correction Factor =  $(A * B)^{0.5}$  Corrected Flow = Calibrated Flow Meter \* Correction Factor

$$A = \frac{\text{Barometric Pressure in inches of Hg.}}{29.92} \quad B = \frac{529.67}{459.67 + ^\circ\text{F}} * \frac{181.87}{\mu_{air}}$$

$$\mu_{air} = \frac{14.58 \left( \frac{459.67 + ^\circ\text{F}}{1.8} \right)^{3/2}}{110.4 + \left( \frac{459.67 + ^\circ\text{F}}{1.8} \right)}$$

Comments:

### Reference Instrument:

Air Flow Calibrator:  AFC-85L sn: 6042

HFC-SIDE-60C sn: 12723

Calibrated By: 

Calibration Date: 2/14/17

Calibration Due: 2/14/18

Reviewed By: 

Review Date: 14 Feb 2017



# Certificate of Calibration


## Air Sampler Calibration Form

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGoffice.com

Sampler: Manufacturer: Eberline Model Number: RAS-1 Serial Number: 405685-3  
Warm up Time 20 min Temperature: 73 °F Relative Humidity 20 % Barometric Pressure: 24.63 in. Hg  
Correction Factor: 0.902767

Sampler Rotometer	Calibrator Flow Meter	Corrected Flow
10	10	9.027671
20	19	17.15257
30	29	26.18024
40	34	30.69408
50	47	42.43005
60	52	46.94389
70	58	52.36049

Calibration Chart



$$y = 7.3511x + 2.7083$$

$$R^2 = 0.9887$$

Correction Factor =  $(A * B)^{0.5}$  Corrected Flow = Calibrated Flow Meter \* Correction Factor


$$A = \frac{\text{Barometric Pressure in inches of Hg.}}{29.92} \quad B = \frac{529.67}{459.67 + ^\circ\text{F}} * \frac{181.87}{\mu_{\text{air}}}$$

$$\mu_{\text{air}} = \frac{14.58 \left( \frac{459.67 + ^\circ\text{F}}{1.8} \right)^{3/2}}{110.4 + \left( \frac{459.67 + ^\circ\text{F}}{1.8} \right)}$$

Comments:

### Reference Instrument:

Air Flow Calibrator:  AFC-85L sn: 6042  HFC-SIDE-60C sn: 12723

Calibrated By:  Calibration Date: 14 Feb 2017 Calibration Due: 14 Feb 2017

Reviewed By:  Review Date: 2/14/17



# Certificate of Calibration

## Calibration and Efficiency Determination

Environmental Restoration Group, Inc.  
 8809 Washington St NE, Suite 150  
 Albuquerque, NM 87113  
 (505) 298-4224  
 www.ERGOoffice.com

Meter: Manufacturer: Ludlum Model Number: 12 Serial Number: 276863  
 Detector: Manufacturer: Ludlum Model Number: 44-9 Serial Number: PR147787

Mechanical Check     THR/WIN Operation    HV Check (+/- 2.5%):  500 V     1000 V     1500 V  
 F/S Response Check     Reset Check    Cable Length:  39-inch     72-inch     Other:  
 Geotropism     Audio Check  
 Meter Zeroed     Battery Check (Min 4.4 VDC)  
 Source Distance:  Contact     6 inches     Other:    Threshold: 40 mV    Barometric Pressure: 24.42 inches Hg  
 Source Geometry:  Side     Below     Other:    Window:    Temperature: 75 °F  
 Relative Humidity: 20 %

Instrument found within tolerance:  Yes     No



Range/Multiplier	Reference Setting	"As Found Reading"	Meter Reading	Integrated 1-Min. Count	Log Scale Count
x 1000	400	400 kcpm	400 kcpm	400	kcpm
x 1000	100	100 kcpm	100 kcpm	100	kcpm
x 100	400	400 kcpm	400 kcpm	400	kcpm
x 100	100	100 kcpm	100 kcpm	100	kcpm
x 10	400	400 kcpm	400 kcpm	400	kcpm
x 10	100	100 kcpm	100 kcpm	100	kcpm
x 1	400	400 cpm	400 cpm	400	cpm
x 1	100	100 cpm	100 cpm	100	cpm

Gross Tc-99 counts (cpm): 2500    Gross Sr/Y-90 counts (cpm):  
 Background counts (cpm): 70    Background counts (cpm):  
 Net Tc-99 Counts (cpm): 2430    Net Sr/Y-90 counts (cpm):

Comments:

### Reference Instruments and/or Sources:

Ludlum pulser serial number:  97743     201932    Fluke multimeter serial number:  87490128  
 Alpha Source: Th-230 @ 12,800 dpm (1/4/12) sn: 4098-03     Gamma Source Cs-137 @ 5.2 uCi (1/4/12) sn: 4097-03  
 Beta Source: Tc-99 @ 17,700 dpm (1/4/12) sn: 4099-03     Other Source:

Calibrated By:  Calibration Date: 7 Feb 2017    Calibration Due: 7 Feb 2018  
 Reviewed By:  Review Date: 2/7/17



# Certificate of Calibration

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGoffice.com

## Calibration and Voltage Plateau

Meter/Detector: Manufacturer: Ludlum Model Number: 2929 & 43-10-1 Serial Number: 200051 & PR215948

- Mechanical Check
- F/S Response Check
- Geotropism
- Meter Zeroed
- THR/WIN Operation
- Reset Check
- Audio Check
- Battery Check (Min 4.4 VDC)

HV Check (+/- 2.5%):  500 V  1000 V  1500 V  
Cable Length:  39-inch  72-inch  Other:

Source Distance:  Contact  6 inches  Other: In planchet  
Source Geometry:  Side  Below  Other: In planchet

Alpha Threshold: 170 mV Barometric Pressure: 24.42 inches Hg  
Beta Threshold: 4 mV Temperature: 74 °F  
Beta Window: 46 mV Relative Humidity: 20 %

Instrument found within tolerance:  Yes  No

Range/Multiplier	Reference Setting	Integrated 1-Min. Count "As Found"		Integrated 1-Min. Count "Reading"	
		$\alpha$	$\beta$	$\alpha$	$\beta$
x 1000	400 Kcpm	399766	399773	399766	399773
x 100	40 Kcpm	39980	39980	39980	39980
x 10	4 Kcpm	3998	3998	3998	3998
x 1	400 cpm	400	399	400	399

High Voltage	Pot. Setting	Alpha Source		Beta Source		Background	
		$\alpha$	$\beta$	$\alpha$	$\beta$	$\alpha$	$\beta$
500	2	3426	306	3	784	0	13
550	2.24	4521	263	6	2322	0	49
600	2.48	4708	436	4	3560	1	53
650	2.68	4650	863	4	4631	1	73

Comments: HV Plateau Scaler Count Time = 1-min. Recommended HV = 600, Pot. setting = 2.48

### Reference Instruments and/or Sources:

Ludlum pulser serial number:  97743  201932


Fluke multimeter serial number  87490128

Alpha Source: Th-230 @ 12,800 dpm (1/4/12) sn: 4098-03


Gamma Source Cs-137 @ 5.2 uCi (1/4/12) sn: 4097-03

Beta Source: Tc-99 @ 17,700 dpm (1/4/12) sn: 4099-03

Other Source:

Calibrated By: 

Calibration Date: 6 Dec 16 Calibration Due: 6 Dec 17

Reviewed By: 

Date: 12/6/16



# Certificate of Calibration

## Calibration and Voltage Plateau

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGOoffice.com


Meter: Manufacturer: Ludlum Model Number: 2360 Serial Number: 184909  
Detector: Manufacturer: Ludlum Model Number: 43-93 Serial Number: PR298426

- Mechanical Check
- F/S Response Check
- Geotropism
- Meter Zeroed
- Source Distance:  Contact  6 inches  Other:
- Source Geometry:  Side  Below  Other:
- THR/WIN Operation
- Reset Check
- Audio Check
- Battery Check (Min 4.4 VDC)
- Alpha Threshold: 120 mV
- Beta Threshold: 4
- Beta Window: 30 mV
- HV Check (+/- 2.5%):  500 V  1000 V  1500 V
- Cable Length:  39-inch  72-inch  Other:
- Barometric Pressure: 24.51 inches Hg
- Temperature: 76 °F
- Relative Humidity: 20 %

Instrument found within tolerance:  Yes  No

Range/Multiplier	Reference Setting	"As Found Reading"	Meter Reading	Integrated 1-Min. Count	
				$\alpha$	$\beta$
x 1000	400 Kcpm	400	400	399826	399835
x 1000	100 Kcpm	100	100		
x 100	40 Kcpm	400	400	39985	39986
x 100	10 Kcpm	100	100		
x 10	4 Kcpm	400	400	3998	3997
x 10	1 Kcpm	100	100		
x 1	400 cpm	400	400	400	399
x 1	100 cpm	100	100		

High Voltage	Alpha Source		Beta Source		Background	
	$\alpha$	$\beta$	$\alpha$	$\beta$	$\alpha$	$\beta$
675	1676	352	7	1153	3	82
700	2255	366	8	1843	8	158
725	2437	495	7	2317	1	226
750	2646	579	4	2871	5	378



Comments: HV Plateau Scaler Count Time = 1 min. Recommended HV = 725

### Reference Instruments and/or Sources:

Ludlum pulser serial number:  97743  201932

Fluke multimeter serial number  87490128

Alpha Source: Th-230 (s/n 4098-03) 12,800 dpm on 1/4/12

Gamma Source Cs-137 @ 5.2 uCi (1/4/12) sn: 4097-03

Beta Source: Tc-99 (sn: 4099-03) 17,700 dpm on 1/4/12

Other Source:

Calibrated By:

Calibration Date: 1/16/17

Calibration Due: 1/16/18

Reviewed By:

Date: 1/16/17

ERG Form ITC. 101.C



# Certificate of Calibration

Calibration and Voltage Plateau

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGoffice.com


Meter: Manufacturer: Ludlum Model Number: 2360 Serial Number: 202403  
 Detector: Manufacturer: Ludlum Model Number: 43-37 Serial Number: PR178371

Mechanical Check       THR/WIN Operation      HV Check (+/- 2.5%):  500 V     1000 V     1500 V  
 F/S Response Check     Reset Check                      Cable Length:  39-inch     72-inch     Other: 60"  
 Geotropism                       Audio Check  
 Meter Zeroed                       Battery Check (Min 4.4 VDC)      Alpha Threshold: 100 mV    Barometric Pressure: 24.86 inches Hg  
 Source Distance:  Contact     6 inches     Other:                      Beta Threshold: 4                      Temperature: 75 °F  
 Source Geometry:  Side     Below     Other:                      Beta Window: 40 mV                      Relative Humidity: 20 %

Instrument found within tolerance:  Yes     No

Range/Multiplier	Reference Setting	"As Found Reading"	Meter Reading	Integrated 1-Min. Count	
				$\alpha$	$\beta$
x 1000	400 Kcpm	400	400	399634	399649
x 1000	100 Kcpm	100	100		
x 100	40 Kcpm	400	400	39964	39965
x 100	10 Kcpm	100	100		
x 10	4 Kcpm	400	400	3993	3993
x 10	1 Kcpm	100	100		
x 1	400 cpm	400	400	399	399
x 1	100 cpm	100	100		

High Voltage	Alpha Source		Beta Source		Background	
	$\alpha$	$\beta$	$\alpha$	$\beta$	$\alpha$	$\beta$
1550	1780	1293	1	3838	2	480
1575	2088	1506	4	4504	7	626
1600	2053	1661	6	5040	7	805
1625	2308	1873	5	5320	6	1028



Comments: HV Plateau Scaler Count Time = 1 min. Recommended HV = 1600

### Reference Instruments and/or Sources:

Ludlum pulser serial number:  97743     201932                      Fluke multimeter serial number  87490128  
 Alpha Source: Th-230 (sn 4098-03) 12,800 dpm on 1/4/12                       Gamma Source Cs-137 @ 5.2 uCi (1/4/12) sn: 4097-03  
 Beta Source: Tc-99 (sn 4099-03) 17,700 dpm on 1/4/12                       Other Source:

Calibrated By: [Signature]                      Calibration Date: 11-4-16                      Calibration Due 11-4-17  
 Reviewed By: [Signature]                      Date: 11/04/16



# Certificate of Calibration

Calibration and Voltage Plateau

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGoffice.com

Meter: Manufacturer: Ludlum Model Number: 2360 Serial Number: 193638

Detector: Manufacturer: Ludlum Model Number: 43-93 Serial Number: PR199836

Mechanical Check  THR/WIN Operation HV Check (+/- 2.5%):  500 V  1000 V  1500 V

F/S Response Check  Reset Check Cable Length:  39-inch  72-inch  Other:

Geotropism  Audio Check

Meter Zeroed  Battery Check (Min 4.4 VDC)

Source Distance:  Contact  6 inches  Other:

Alpha Threshold: 120 mV Barometric Pressure: 24.39 inches Hg

Beta Threshold: 4 Temperature: 73 °F


Source Geometry:  Side  Below  Other:

Beta Window: 30 mV Relative Humidity 20 %

Instrument found within tolerance:  Yes  No

Range/Multiplier	Reference Setting	"As Found Reading"	Meter Reading	Integrated 1-Min. Count	
				$\alpha$	$\beta$
x 1000	400 Kcpm	400	400	397888	397923
x 1000	100 Kcpm	100	100		
x 100	40 Kcpm	400	400	39787	39799
x 100	10 Kcpm	100	100		
x 10	4 Kcpm	400	400	3981	3980
x 10	1 Kcpm	100	100		
x 1	400 cpm	400	400	398	398
x 1	100 cpm	100	100		

High Voltage	Alpha Source		Beta Source		Background	
	$\alpha$	$\beta$	$\alpha$	$\beta$	$\alpha$	$\beta$
650	2429	320	4	1255	5	135
675	2687	391	12	2131	3	191
700	2817	539	10	2836	1	270
725	2824	741	7	3269	8	356



Comments: HV Plateau Scaler Count Time = 1 min. Recommended HV = 700

### Reference Instruments and/or Sources:

Ludlum pulser serial number:  97743  201932

Fluke multimeter serial number  87490128

Alpha Source: Th-230 (s/n 4098-03) 12,800 dpm on 1/4/12

Gamma Source Cs-137 @ 5.2 uCi (1/4/12) sn: 4097-03

Beta Source: Tc-99 (sn: 4099-03) 17,700 dpm on 1/4/12

Other Source:

Calibrated By:

Calibration Date: 2/7/17

Calibration Due: 2/7/18

Reviewed By:

Date: 2 Feb 2017

ERG Form ITC. 101.C

This calibration conforms to the requirements and acceptable calibration conditions of ANSI N323.4 - 1997





# Certificate of Calibration

## MSA Lapel Air Sampler Calibration Form

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGoffice.com

Manufacturer: MSA

Model Number: Escort ELF Lapel Air Sampler

Serial Number: A3-48588

Temperature: 72 °F

Relative Humidity: 20 %

Barometric Pressure: 24.92 in. Hg

Single point calibration at two different flow rates. Actual flow rate is expressed as average of six tests. Sampler flow rate adjusted to match average.

Digital Flow Run Number	Test 1 (2000 cc/min)	Test 2 (3000 cc/min)
1	2066	2810
2	2071	2957
3	1916	2847
4	2071	2969
5	2083	2901
6	1967	3051

Average: 2029.0

2922.5

Adjustment Set:

Flow Fault Check:

Final Flow Setting: 2.5 LPM

Comments:

### Reference Instrument:

Air Flow Calibrator:  MSA DIGICAL Calibrator sn: 020158

A.P. Buck mini-Buck Calibrator sn: 053058

Calibrated By:

Calibration Date:

12/29/16

Reviewed By:

Review Date:

12/29/16



# Certificate of Calibration

Calibration and Voltage Plateau

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGoffice.com

Meter: Manufacturer: Ludlum Model Number: 2221r Serial Number: 190205  
Detector: Manufacturer: Ludlum Model Number: FIDLER Serial Number: 010807D

- Mechanical Check
- F/S Response Check
- Geotropism
- Meter Zeroed
- THR/WIN Operation
- Reset Check
- Audio Check
- Battery Check (Min 4.4 VDC)
- Source Distance:  Contact  6 inches  Other: 1/2 "
- Source Geometry:  Side  Below  Other:

HV Check (+/- 2.5%):  500 V  1000 V  1500 V  
Cable Length:  39-inch  72-inch  Other: 60"

Barometric Pressure: 24.42 inches Hg  
Temperature: 74 °F  
Relative Humidity: 20 %

Threshold: 10 mV  
Window:


Instrument found within tolerance:  Yes  No

Range/Multiplier	Reference Setting	"As Found Reading"	Meter Reading	Integrated 1-Min. Count	Log Scale Count
x 1000	400	400	400	399325	400
x 1000	100	100	100		100
x 100	400	400	400	39933	400
x 100	100	100	100		100
x 10	400	400	400	3993	400
x 10	100	100	100		100
x 1	400	400	400	400	400
x 1	100	100	100		100

High Voltage	Source Counts
700	4140
800	82919
900	136675
950	157963
1000	161842
1050	162184
1100	163399
1150	163508
1200	163966

Background  
5070

Voltage Plateau



Comments: HV Plateau Scaler Count Time = 0.5-min. Recommended HV = 1050

### Reference Instruments and/or Sources:

- Ludlum pulser serial number:  97743  201932
- Alpha Source: Th-230 @ 12,800 dpm (1/4/12) sn: 4098-03
- Beta Source: Tc-99 @ 17,700 dpm (1/4/12) sn: 4099-03

- Fluke multimeter serial number:  87490128
- Gamma Source Cs-137 @ 5.2 uCi (1/4/12) sn: 4097-03
- Other Source: Am-241 @ 1uCi

Calibrated By:

Calibration Date: 7 Feb 2017 Calibration Due: 7 Feb 2018

Reviewed By:

Date: 2/7/17

ERG Form ITC. 101.A

This calibration conforms to the requirements and acceptable calibration conditions of ANSI N323.A - 1997



# Certificate of Calibration

Environmental Restoration Group, Inc.  
8809 Washington St NE, Suite 150  
Albuquerque, NM 87113  
(505) 298-4224  
www.ERGoffice.com

## Calibration and Voltage Plateau

Meter: Manufacturer: Ludlum Model Number: 2360 Serial Number: 220242  
Detector: Manufacturer: Ludlum Model Number: Serial Number:

- Mechanical Check
  - F/S Response Check
  - Geotropism
  - Meter Zeroed
  - Source Distance:  Contact  6 inches  Other:
  - Source Geometry:  Side  Below  Other:
  - THR/WIN Operation
  - Reset Check
  - Audio Check
  - Battery Check (Min 4.4 VDC)
  - HV Check (+/- 2.5%):  500 V  1000 V  1500 V
  - Cable Length:  39-inch  72-inch | | Other:
  - Alpha Threshold: 120 mV
  - Beta Threshold: 4
  - Beta Window: 30 mV
  - Barometric Pressure: 24.6 inches Hg
  - Temperature: 74 °F
  - Relative Humidity: 20 %
- Instrument found within tolerance:  Yes  No

Range/Multiplier	Reference Setting		"As Found Reading"	Meter Reading		Integrated 1-Min. Count	
	$\alpha$	$\beta$		$\alpha$	$\beta$	$\alpha$	$\beta$
x 1000	400 Kcpm		400	400	400430	400469	
x 1000	100 Kcpm		100	100			
x 100	40 Kcpm		400	400	40079	40083	
x 100	10 Kcpm		100	100			
x 10	4 Kcpm		400	400	4005	4004	
x 10	1 Kcpm		100	100			
x 1	400 cpm		400	400	400	400	
x 1	100 cpm		100	100			



  

High Voltage	Alpha Source	Beta Source	Background	Voltage Plateau
	$\alpha$ $\beta$	$\alpha$ $\beta$	$\alpha$ $\beta$	

Comments: HV Plateau Scaler Count Time = 1 min. HV = 500

### Reference Instruments and/or Sources:

- Ludlum pulser serial number:  97743  201932  
 Fluke multimeter serial number  87490128  
 Alpha Source: Th-230 (s/n 4098-03) 12,800 dpm on 1/4/12  
 Beta Source: Tc-99 (sn: 4099-03) 17,700 dpm on 1/4/12  
 Gamma Source Cs-137 @ 5.2 uCi (1/4/12) sn: 4097-03  
 Other Source:

Calibrated By:  Calibration Date: 2/13/17 Calibration Due: 2/13/18  
Reviewed By:  Date: 2/13/17

ERG Form ITC. 101.C

This calibration conforms to the requirements and acceptable calibration conditions of ANSI N323.4 - 1997