

From: [Lanzisera, Penny](#)
To: [Bohan, Michael](#)
Subject: Request for Additional Information for Bridgeport Hospital
Date: Thursday, April 04, 2019 5:46:00 PM

Licensee: Bridgeport Hospital
License No. 06-01060-01
Docket No. 03001247
Mail Control No. 611557

Mr. Bohan, to continue our review of the request dated February 27, 2019 to add Dr. Rotem to the above license, we require the following additional information:

1. The license attached to the letter appears to be a permit issued by the Georgia licensee. Please submit a copy of the license issued by the State of Georgia to Augusta Medical Center. In addition, the permit includes Dr. Rotem for microsphere use; but does not appear to include microspheres under possession limits. Please submit a complete permit that includes the possession for microspheres in addition to the authorized users.
2. Please provide evidence from Sirtex of the completion of in-vivo casework by Dr. Rotem (i.e., 3 cases).
3. The Sirtex Sealed Source and Device Registry has been updated to include up to 296 mCi/vial and a new manufacturer in Singapore. In addition, the manufacturer has also suggested that the total possession limit be increased to 1 Ci. Please advise if these revisions should be made to your license.

You may submit the additional information to my attention either via signed pdf sent to my email or via fax to 610-337-5269. If we do not receive a reply within 30 days of this email, we will consider that you no longer require the requested amendment, and void your request. Thank you for your assistance,

Penny Lanzisera
Senior Health Physicist
U.S. NRC, Region I