

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: ARS International, LLC 2609 North River Road Port Allen, LA 70767 REPORT NO.: 2019-001		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region IV, 1600 East Lamar Blvd Arlington, Texas 76011-4511	
3. DOCKET NUMBER	4. LICENSE NUMBER	5. DATE OF INSPECTION	
030-38546	17-29441-01	April 10, 2019	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy to exercise discretion, were satisfied.
- Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

License Condition 24 A requires, in part, that the licensee conduct its program in accordance with the license application.

The license application states that the Radiation Safety Committee shall meet on an annual basis.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Kelly Ausbrooks	<i>[Signature]</i>	4-10-19
NRC INSPECTOR	Jason Dykert	<i>[Signature]</i>	4-10-19
BRANCH CHIEF	Patricia A. Silva	<i>[Signature]</i>	4/30/19

Non-Public
 Sensitive - Security-Related
 Public
 Non-Sensitive

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: ARS International, LLC 2609 North River Road Port Allen, LA 70767 REPORT NO.: 2019-001		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region IV, 1600 East Lamar Blvd Arlington, Texas 76011-4511	
3. DOCKET NUMBER 030-38546	4. LICENSE NUMBER 17-29441-01	5. DATE OF INSPECTION April 10, 2019	

CONTINUED FROM PART 1

Contrary to the license application, from 2017 to 2019, the Radiation Safety Committee did not meet on an annual basis.

The licensee will re-establish the Radiation Safety Committee meetings and meet as often as needed to ensure the radiation safety program is operating in compliance with the license, at a minimum on an annual basis.

License Condition 23 A requires, in part, that proposed revisions to procedures in the license application are documented, reviewed, and approved by the Radiation Safety Committee prior to implementation.

Contrary to the above, on September 25, 2018, Procedure RSP-001 "Laboratory Radiation Safety Plan, Revision 15.3" was revised without review and approval by the Radiation Safety Committee.

The licensee will have the procedure revision reviewed and approved within 30 days by the Radiation Safety Committee. When procedure revisions are needed, the licensee will have them reviewed and approved by the RSC.