

UNITED STATES NUCLEAR REGULATORY COMMISSION WASHINGTON, D.C. 20555-0001

August 1, 2019

David Matos, Program Manager Georgia Environmental Protection Division Radioactive Materials Program 4244 International Parkway, Suite 120 Atlanta, GA 30354

Dear Mr. Matos:

We have reviewed the proposed revisions to the Georgia regulations Chapter 391-3-17, received by our office on June 12, 2019. These regulations were reviewed by comparison to the equivalent U.S. Nuclear Regulatory Commission (NRC) rules and Regulation Amendment Tracking System Identification Numbers (RATS ID) 2013-2, 2018-1 and 2018-2 as identified in the enclosed State Regulation Status (SRS) Data Sheet. We discussed our review of the regulations with you on July 12, 2019.

As a result of our review, we have 27 comments related to compatibility, and one editorial comment that are identified in the enclosure. Please note that we have limited our review to regulations required for compatibility and/or health and safety. Under our current procedure, a finding that the Georgia regulations meet the compatibility and health and safety categories of the equivalent NRC regulation may only be made based on a review of the final Georgia regulations. However, we have determined that if your proposed regulations were adopted, incorporating our comments and without other significant change, they would meet the compatibility and health and safety categories established in the Office of Nuclear Material Safety and Safeguards (NMSS) Procedure SA-200, "Compatibility Categories and Health and Safety Identification for NRC Regulations and Other Program Elements."

We request that when the proposed regulations are adopted and published as final regulations, a copy of the "as published" regulations be provided to us for review. As requested in NMSS Procedure SA-201, "Review of State Regulatory Requirements," please highlight the final changes, and provide a copy to Division of Materials Safety, Security, State, and Tribal Programs, NMSS.

The SRS Data Sheet summarizes our knowledge of the status of other Georgia regulations, as indicated. Please let us know if you note any inaccuracies, or have any comments on the information contained in the SRS Data Sheet. This letter, including the SRS Data Sheet, is posted on the NMSS State Communication Portal: <u>https://scp.nrc.gov/rulemaking.html</u>.

If you have any questions regarding the comments, the compatibility and health and safety categories, or any of the NRC regulations used in the review, please contact Michelle Beardsley, State Regulation Review Coordinator, at (301) 415-0275 (<u>Michelle.Beardsley@nrc.gov</u>) or Monica Ford at (610) 337-5214 (<u>Monica.Ford@nrc.gov</u>).

Sincerely,

/RA Paul Michalak for/

Kevin Williams, Deputy Director Division of Materials Safety, Security, State and Tribal Programs Office of Nuclear Material Safety and Safeguards

Enclosures:

- 1. Compatibility and Editorial Comments
- 2. Georgia SRS Data Sheet

SUBJECT: GEORGIA PROPOSED REGULATIONS TO RATS IDS 2018-1 AND 2018-2 AND REVISED FINAL REGULATIONS

DISTRIBUTION: SP[05] DIR RF (19-32) MFord, RSAO [GA] File

| OFFICE | RI | SALPB | OGC | SALPB:BC | MSST:DD |
|----------------------|-----------|-------------|----------------|-----------|----------------|
| NAME | MFord via | MBeardsley | IIrvin NLO via | PMichalak | PMichalak for: |
| | email | | email | | KWilliams |
| DATE | 6/26/19 | 6/26/19 | 7/12/19 | 08/01/19 | 08/01/19 |
| ML19178A323 (Letter) | | er) OFFICIA | L RECORD COPY | Package | e ML19164A100 |

COMPATIBILITY COMMENTS ON GEORGIA PROPOSED REGULATIONS

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|------------------------|-------------|---------|----------|--|
| 1 | 391-3-17- .02(5)(f) | 40.22(d) | 2013-2 | В | Small quantities of source material |
| | | | | | GA needs to change .02(5)(f) to read: "Any person who receives, possesses, uses, or transfers source material in accordance with the general license granted in subparagraph (b) of this section is exempt from the provisions of <u>391-3-</u> <u>1703 and 391-3-1707</u> to the extent that such receipt, possession, use, and transfer are within the terms of this general license, except that such person shall comply with the provisions of <u>391-3-1703(7)(b) and</u> <u>391-3-1703(13)(a)</u> to the extent necessary to meet the provisions of subparagraphs (c)(<u>2)</u> and (e) of this section. However, this exemption does not apply to any person who also holds a specific license issued <u>under this chapter.</u> " |
| | | | | | Category B designation assigned to 10 CFR 40.22(d). |
| 2 | 391-3-17- .02(5)(m) | 40.55(c) | 2013-2 | В | Conditions of licenses to initially transfer source material for use under the 'small quantities of source material' general license: quality control, labeling, safety instructions, and records and reports |
| | | | | | GA needs to rewrite $.02(5)(m)$ to read "Each person licensed under <u>391-3-</u> <u>1702(5)(h)</u> shall provide the information specified in this paragraph to each person to whom source material is transferred for use under <u>391-3-1702(5)(b)</u> . This information must be transferred before the source material is transferred for the first time in each calendar year to the particular |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|-------------------------------------|-------------|---------|----------|---|
| | | | | | recipient. The required information includes: (1) A copy of <u>391-3-1702(5)(b) and</u> <u>.02(19)</u> , or <u>relevant equivalent</u> <u>regulations of the NRC or an</u> <u>Agreement State</u> . (2) Appropriate radiation safety precautions and instructions relating to handling, use, storage, and disposal of the material." GA needs to make the above change in order to meet the Compatibility Category B designation assigned to 10 CFR 40.55(c). |
| 3 | 391-3-17- .02(5)(k) | 40.55(d) | 2013-2 | В | Conditions of licenses to initially transfer source material for use under the 'small quantities of source material' general license: quality control, labeling, safety instructions, and records and reports GA needs to rewrite the last sentence of .02(5)(k)3. to read "If no transfers have been made to general licensees in a particular Agreement State or falling under the jurisdiction of the NRC, during the reporting period, this information shall be reported to the NRC or responsible Agreement State agency upon request of the agency or the NRC." GA needs to make the above change in order to meet the Compatibility Category B designation assigned to 10 CFR 40.55(d). |
| 4 | 391-3-17- .05(2)(xx) and (yy) | 35.2 | 2018-1 | В | Definitions GA has an incorrect reference in the definition of Associate Radiation Safety Officer. GA needs to revise 391-3-1705(2)(xx)1. to read: "Meets the requirements in Rule .05(22) and .05(27); and." |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|----------------------|-------------|---------|----------|---|
| | | | | | GA has an incorrect reference in the definition of Ophthalmic Physicist. GA needs to revise (yy)1. to read: "Meets the requirements in 391-3-17- .05(27) and 391-3-1705(64)(c)(2); and." |
| | | | | | GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.2, Definitions: Associate Radiation Safety Officer and Ophthalmic Physicist. |
| 5 | 391-3-17- | 35.40 | 2018-1 | H&S | Written directives |
| | .05(19) | | | | GA omits an equivalent regulation to 10 CFR 35.40(b)(6) from 391-3-1705(19). |
| | | | | | GA needs to add language equivalent to 10 CFR 35.40(b)(6) to .05(19) in order to meet the Compatibility Category H&S designation assigned to 10 CFR 35.40. |
| 6 | 391-3-17- .05(22) | 35.50 | 2018-1 | В | Training for Radiation Safety Officer and Associate Radiation Safety Officer |
| | | | | | GA's regulation 391-3-1705(22) is not essentially identical to 10 CFR 35.50. Specifically GA needs to restructure its regulation to match the new structure of 10 CFR 35.50 and change the references to match. As an example .05(22)(a) should not point to the requirement for a written attestation but instead point to the training described in .05(22)(e). The written attestation required by paragraph .05(22)(d) is only required for individuals qualifying under .05(22)(b). |
| | | | | | GA needs to rewrite 391-3-1705(22) as indicated above in order to meet the Compatibility Category B designation assigned to 10 CFR |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|----------------------|-------------|---------|----------|--|
| | | | | | 35.50. |
| 7 | 391-3-17- .05(23) | 35.51 | 2018-1 | В | Training for an authorized medical physicist. |
| | | | | | GA's regulation 391-3-1705(23) is not essentially identical to 10 CFR 35.51. |
| | | | | | Specifically GA needs to change the reference in .05(23)(a) to only point to the requirement for the training described in .05(23)(c). The written attestation required by paragraph .05(23)(b)(2) is only required for individuals qualifying under .05(23)(b). |
| | | | | | GA also needs to remove the references to .05(23)(a)1. and 2. from .05(23)(b)2. |
| | | | | | GA needs to rewrite 391-3-1705(23) as indicated above in order to meet the Compatibility Category B designation assigned to 10 CFR 35.51. |
| 8 | 391-3-17- .05(24) | 35.55 | 2018-1 | В | Training for an authorized nuclear pharmacist. |
| | | | | | GA needs to remove "and who meets the requirements in .05(24)(b)2." from .05(24)(a). |
| | | | | | GA needs to remove ".05(24(a)1., .05(24)(a)2., and 05(24)(a)3." from .05(24)(b)(2). |
| | | | | | GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.55. |
| 9 | 391-3-17- .05(26) | 35.57 | 2018-1 | В | Training for experienced radiation safety officer, teletherapy or medical physicist, authorized medical physicist, authorized user, |

| STA | | NPC SECTION | | CATEGORY | |
|-----|----------------------|-------------|---------|----------|--|
| 514 | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS nuclear pharmacist, and authorized nuclear pharmacist. GA needs to combine 391-3-17- .05(26)(a)(1) and (2). Additionally, |
| | | | | | GA omits "except the Radiation Safety Officers and authorized medical physicists identified in this paragraph must meet the training requirements in § 35.50(d) or § 35.51(c), as appropriate, for any material or uses for which they were not authorized prior to this date." from the end of 391-3-1705(26)(a). |
| | | | | | GA needs to rewrite 391-3-17- .05(26)(b) to incorporate 10 CFR 35.57(b)(1) and incorporate the language in 35.57(b)(2)(i)-(iv). |
| | | | | | GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.57. |
| 10 | 391-3-17- .05(43) | 35.190 | 2018-1 | В | Training for uptake, dilution, and excretion studies |
| | | | | | GA needs to remove the reference to ".05(43)(c)(2)" from .05(43)(a). Written attestations are no longer required for individuals who are certified by a specialty board. |
| | | | | | GA needs to remove the reference to ".05(43)(a)1." from .05(43)(c)2. Additionally GA needs to change the language after the reference to .05(43)(c)1. to say "and is able to independently fulfill the radiation safety-related duties as an authorized user for the medical uses authorized under Rule 005(41)." |
| | | | | | GA needs to change the reference in .05(43)(c)2.(b) from "training and experience specified in paragraph (a)" |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|----------------------|-------------|---------|----------|--|
| | | | | | to "training and experience specified in paragraph (c)1." GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.190. |
| 11 | 391-3-17- .05(45) | 35.204 | 2018-1 | H&S | Permissible molybdenum-99, strontium-82, and strontium-85 concentrations GA needs to change .05(45)(b) to read as follows: "(b) To demonstrate compliance with Rule .05(45)(a), the licensee preparing radioactive drugs from radionuclide generators shall 1. Measure the molybdenum-99 concentration in each eluate from a generator to demonstrate compliance with paragraph (a) of this section. 2. before the first patient use of the day, measure the concentration of radionuclides strontium-82 and strontium-85 to demonstrate compliance with paragraph (a) of this section." GA needs to make the above change in order to meet the Compatibility Category H&S designation assigned to 10 CFR 35.204. |
| 12 | 391-3-17- .05(47) | 35.290 | 2018-1 | В | Training for imaging and localization studies GA needs to remove the reference to ".05(47)(c)(2)" from .05(47)(a). Written attestations are no longer required for individuals who are certified by a specialty board. GA needs to remove the reference to ".05(47)(a)1." from .05(47)(c)2. Additionally GA needs to change the language after the reference to .05(47)(c)1. to say "and is able to independently fulfill the radiation safety-related duties as an authorized |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|----------------------|-------------|---------|----------|--|
| | | | | | user for the medical uses authorized under Rule .05(41) and (44)." GA needs to change the reference in .05(47)(c)2.(b) from "training and experience specified in paragraph (a)" to "training and experience specified in paragraph (c)1." GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.290. |
| 13 | 391-3-17- .05(48) | 35.300 | 2018-1 | В | Use of unsealed byproduct material for which a written directive is required GA needs to change the reference in the intro paragraph to .05(48) to read, "identified in paragraph (52)(b)1.(ii)(VII)." GA needs to make the above change in order to meet the Compatibility Category B designation assigned to 10 CFR 35.300. |
| 14 | 391-3-17- .05(52) | 35.390 | 2018-1 | В | Training for use of unsealed byproduct material for which a written directive is required GA needs to remove the reference to ".05(52)(b)(2)" from .05(52)(a). Written attestations are no longer required for individuals who are certified by a specialty board. GA needs to remove the reference to ".05(52)(a)1. and .05(52)(b)1.(ii)(VII)" from .05(52)(b)2. Additionally GA needs to change the language after the reference to .05(52)(b)1. to say "and is able to independently fulfill the radiation safety-related duties as an authorized user for the medical uses authorized under Rule 005(48)." |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|----------------------|-------------|---------|----------|---|
| | | | | | GA needs to change the reference in .05(52)(b)2.(b) from "training and experience specified in paragraph (a)" to "training and experience specified in paragraph (b)1." GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.390. |
| 15 | 391-3-17- .05(53) | 35.392 | 2018-1 | В | Training for the oral administration of sodium iodide I-131 requiring a written directive in quantities less than or equal to 1.22 gigabecquerels (33 millicuries) GA needs to remove the reference to ".05(53)(c)(3)" from .05(53)(a). Written attestations are no longer required for individuals who are certified by a specialty board. GA needs to change the language after the reference to .05(53)(c)2. in Section .05(53)(c)3. to say, "and is able to independently fulfill the radiation safety-related duties as an authorized user for the medical uses authorized under Rule 005(48)." GA needs to change the reference in .05(53)(c)3.(b) from "training and experience specified in paragraph (a)" to "training and experience specified in paragraph (c)1. and 2." GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.392. |
| 16 | 391-3-17- .05(54) | 35.394 | 2018-1 | В | Training for the oral administration of sodium iodide I–131 requiring a written directive in quantities greater than 1.22 gigabecquerels (33 millicuries) |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|------------------------|-------------|---------|----------|--|
| | | | | | GA needs to remove the reference to ".05(54)(c)(3)" from .05(54)(a). Written attestations are no longer required for individuals who are certified by a specialty board. GA needs to change the language after the reference to .05(54)(c)2. in Section .05(54)(c)3. to say, "and is able to independently fulfill the radiation safety-related duties as an authorized user for the medical uses authorized under Rule 005(48)." GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.394. |
| 17 | 391-3-17- .05(54.1) | 35.396 | 2018-1 | В | Training for the parenteral administration of unsealed byproduct material requiring a written directive GA needs to add ".05(54.1)" after "A supervising authorized user who meets the requirements in Rule .05(52)" in .05(54.1)(d)2. GA needs to change .05(54.1(d)3. to say, "Has obtained written attestation that the individual has satisfactorily completed the requirements in .05(54.1)(d)1. and (d)2., and is able to independently fulfill the radiation safety-related duties as an authorized user for the parenteral administration of unsealed byproduct material requiring a written directive. The attestation must be obtained from either:". GA needs to change the references in the beginning of .05(54.1)(d)3.(a) to read, "A preceptor authorized user, who meets the requirements in Rules .05(26), (52), (54.1), or equivalent." |

| етл | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|--|-------------|---------|----------|---|
| | TE SECTION | INCOSECTION | | | GA needs to change the reference at the beginning of .05(54.1)(d)3.(b) to read, " at least one faculty member is an authorized user who meets the requirements in Rules .05(26), (52), (54.1), or equivalent." GA needs to change the reference at the end of .05(54.1)(d)3.(b) to read, "training and experience specified in paragraph (d)(1) and (2) of this section." GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.396. |
| 18 | 391-3-17- .05(55) | 35.400 | 2018-1 | C | Use of sources for manual brachytherapy GA needs to change .05(55)(b) to read, "In research to deliver therapeutic doses for medical use in accordance with an active Investigational Device Exemption (IDE) application accepted by the U.S. Food and Drug Administration provided the requirements of .05(21)(a) are met." GA needs to make the above change in order to meet the Compatibility Category C designation assigned to 10 CFR 35.400. |
| 19 | 391-3-17- .05(64)(c), (d), & (e) | 35.433 | 2018-1 | B/H&S | Strontium-90 sources for ophthalmic treatments GA needs to change .05(64)(c) to read, "as specified in paragraph (d) of this section are performed" GA needs to change the reference in .05(64)(c)2.(iv)(C) to read, "as detailed in Rule .05(60)." GA needs to change the section that begins with, "The individuals who are |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|----------------------|-------------|---------|----------|--|
| STA | TESECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTSidentified in" to .05(64)(d) (instead of a second paragraph (b)).GA needs to change what should be .05(64)(d) [currently shown as (b)] to read, "The individuals who are identified in paragraph (c) of this section must:".GA needs to change the reference in what should be .05(64)(d)(2) to read, " These procedures must include the frequencies that the individual meeting the requirements in paragraph (c) of this section".GA needs to change the section that begins with, "Licensees must retain a record of the activity of each" to .05(64)(e) [currently shown as a second paragraph (c). |
| | | | | | GA needs to make the above changes in order to meet the Compatibility Categories B and H&S assigned to 10 CFR 35.433. |
| 20 | 391-3-17- .05(63) | 35.490 | 2018-1 | В | Training for use of manual brachytherapy sources GA needs to remove "and who meets the requirements in .05(63)(b)3." from .05(63)(a). GA needs to change the reference in .05(63)(b)1.(ii) to read, "authorized to use byproduct material under Rule .05(55), involving". |
| | | | | | GA needs to change .05(63)(b)3. to read, "Has obtained written attestation that the individual has satisfactorily completed the requirements in paragraphs .05(63)(b)1. and .05(63(b)2. and is able to independently fulfill the radiation safety-related duties as an authorized user of manual brachytherapy |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|----------------------------------|-------------|---------|----------|---|
| | | | | | sources for the medical uses authorized in Rule .05(55)." GA needs to change the references in .05(63)(b)3.(a) and (b) to read, "who meets the requirements in Rules .05(26), (63), or equivalent Agreement State". GA needs to delete, "and has experience in administering dosages in the same dosage category or categories (i.e., in Rule .05(52)(b)1.(ii)(VII)) as the individual requesting authorized user status" from .05(63)(b)3.(a) and (b). GA needs to change the reference at the end of .05(63)(b)3.(b) to read, "must include training and experience specified in paragraph .05(63)(b)1. and 2." GA needs to make the above changes in order to meet the Compatibility Category B designation |
| 21 | 391-3-17- .05(64)(a) & (b) | 35.491 | 2018-1 | В | assigned to 10 CFR 35.490.Training for ophthalmic use of strontium-90GA needs to revise .05(64)(b)3. to read as follows, "has satisfactorily completed the requirements in Rule .05(64)(b)1. and 2., and is able to independently fulfill the radiation safety-related duties as an authorized user of strontium-90 for ophthalmic use."GA needs to make the above changes in order to meet the Compatibility Category B designation assigned to 10 CFR 35.491. |
| 22 | 391-3-17- .05(66) | 35.590 | 2018-1 | В | Training for use of sealed sources and medical devices for diagnosis |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|----------------------|-------------|---------|----------|---|
| | | | | | GA needs to add, "or Nuclear Regulatory Commission" between "Agreement State" and "requirements" in .05(66)(b). GA needs to make the above change in order to meet the Compatibility Category B designation assigned to |
| 23 | 391-3-17- .05(67) | 35.600 | 2018-1 | С | 10 CFR 35.590. Use of a sealed source in a remote afterloader unit, teletherapy unit, or gamma stereotactic radiosurgery unit GA omits, "of Rule .05(21)(1) are met" from .05(67)(a)(2). |
| | | | | | GA needs to make the above change in order to meet the Compatibility Category C designation assigned to 10 CFR 35.600. |
| 24 | 391-3-17- .05(81) | 35.655(a) | 2018-1 | H&S | Full-inspection servicing for teletherapy and gamma stereotactic radiosurgery units |
| | | | | | GA needs to add, "and other safety components" to the end of .05(81)(a). |
| | | | | | GA needs to make the above change in order to meet the Compatibility Category H&S designation assigned to 10 CFR 35.655(a). |
| 25 | 391-3-17- .05(84) | 35.690 | 2018-1 | В | Training for use of remote afterloader units, teletherapy units, and gamma stereotactic radiosurgery units |
| | | | | | GA needs to remove, ".05(84)(b)3." from .05(84)(a). |
| | | | | | GA needs to add, "that is authorized to use radioactive materials in Rule .05(67)" after "medical institution" in .05(84(b)1.(ii). |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|-----------------------|-------------|---------|----------|---|
| | TESECTION | NRC SECTION | RATSID | CATEGORY | GA needs to remove the reference to .05(84)(a)1. from .05(84)(b)3. and add the language, "is able to independently fulfill the radiation safety-related duties as an authorized user of each type of therapeutic medical unit for which the individual is requesting authorized user status." in place of, "and has achieved a level of competency sufficient to function independently as an authorized user of each type of therapeutic medical unit for which the individual is requesting authorized user status." GA needs to change the references in .05(84)(b)3.(a) and (b) to read, "who meets the requirements in Rules .05(26), (84), or equivalent Agreement State…". In section .05(84)(b)3.(a) and (b), GA needs to delete, "and has experience in administering dosages in the same dosage category or categories (i.e., in Rule .05(52)(b)1.(ii)(VII)) as the individual requesting authorized user status", and replace it with, "for the type(s) of therapeutic medical unit for which the individual is requesting authorized user status." GA needs to change the reference at the end of .05(84)(b)3.(b) to read, "must include training and experience specified in paragraph .05(63)(b)1. and 2." GA needs to make the above changes in order to meet the |
| | | | | | Compatibility Category B designation assigned to 10 CFR 35.690. |
| 26 | 391-3-17- .05(115) | 35.3045 | 2018-1 | С | Report and notification of a medical event |
| | | | | | GA needs to add, "except permanent implant brachytherapy" between "material" and "results" in .05(115)(a). |

| STA | TE SECTION | NRC SECTION | RATS ID | CATEGORY | SUBJECT and COMMENTS |
|-----|-----------------------|-------------|---------|----------|---|
| | | | | | GA needs to add, "or the wrong radionuclide for brachytherapy procedures" to .05(115)(a)2.(i). GA needs to rewrite .05(115) (a)3. to contain language equivalent to 10 CFR 35.3045(a)(1)(iii)(A) and (B). GA needs to make the above changes in order to meet the Compatibility Category C designation assigned to 10 CFR 35.3045. |
| 27 | 391-3-17- .05(120) | 35.3204 | 2018-1 | C | Report and notification for an eluate exceeding permissible molybdenum-99, strontium-82, and strontium-85 concentrations GA needs to remove references to the NRC and the NRC Operations Center from .05(120)(a) and (b) and replace them with Georgia Department of Natural Resources, Environmental Protection Division. GA needs to remove the reference to 10 CFR 35.204(a) in .05(120)(a) and replace it with 391-3-1705(45)(a). GA needs to remove the reference to 10 CFR 30.6 and 30.6(a) in .05(120)(b) and replace it with 391-3-1701(13). GA needs to make the above changes in order to meet the Compatibility Category C designation assigned to 10 CFR 35.3204. |

EDITORIAL COMMENT

| STATE SECTION | NRC SECTION | COMMENT |
|---------------|-------------|--------------------------------|
| All | N/A | GA should remove the term |
| | | "licensing state" throughout |
| | | its' regulations, as this term |
| | | no longer exists. |

STATE REGULATION STATUS

State: Georgia

Tracking Ticket Number: 19-32 Date: 08/01/2019

[# amendment(s) reviewed identified by a * at the beginning of the equivalent NRC requirement.]

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|---|--------------------------------|----------------------|--|--|
| 1991-1 | Safety Requirements for Radiographic Equipment Part 34 55 FR 843 (Superceded by 1997-5) | 01/10/1994 | Final | 05/22/1991 | Georgia has adopted Final Regulations equivalent to RATS ID: 1997-5. |
| 1991-2 | ASNT Certification of Radiographers Part 34 56 FR 11504 (Superceded by 1997-5) | none | Not Required | Not Required | Georgia has adopted Final Regulations equivalent to RATS ID: 1997-5. |
| 1991-3 | Standards for Protection Against Radiation Part 20 56 FR 23360; 56 FR 61352; 57 FR 38588; 57 FR 57877; 58 FR 67657; 59 FR 41641; 60 FR 20183 | 01/01/1994 | Final | No Comments 01/27/1998 | |
| 1991-4 | Notification of Incidents Parts 20, 30, 31, 34, 39, 40 and 70 56 FR 64980 | 10/15/1994 | Final ML13364A243 | No Comments 02/20/2014 ML13364A230 | |
| 1992-1 | Quality Management Program and Misadministrations Part 35 56 FR 34104 (Superceded by 2002-2) | 01/27/1995 | Final | 03/16/1994 | Georgia has adopted Final Regulations equivalent to RATS ID: 2002-2. |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|---|--------------------------------|-----------------------------|---------------------------|--|
| 1992-2 | Eliminating the Recordkeeping Requirements for Departures from Manufacturer's Instructions Parts 30 and 35 57 FR 45566 | none | Not Required | Not Required | These regulation changes are not required to be adopted for purposes of Compatibility. |
| 1993-1 | Decommissioning Recordkeeping and License Termination: Documentation Additions [Restricted areas and spill sites] Parts 30 and 40 58 FR 39628 | 10/25/1996 | Final | 03/16/1994 | |
| 1993-2 | Licensing and Radiation Safety Requirements for Irradiators Part 36 58 FR 7715 | 07/01/1996 | Final | 03/16/1994 | |
| 1993-3 | Definition of Land Disposal and Waste Site QA Program Part 61 58 FR 33886 | 07/22/1996 | Not Applicable ⁱ | Not Applicable | Georgia does not have any licensees subject to these regulations (See SECY-95-112) |
| 1994-1 | Self-Guarantee as an Additional Financial Mechanism Parts 30, 40 and 70 58 FR 68726; 59 FR 1618 | none | Final | No Comments 09/25/1998 | These regulation changes are not required to be adopted for purposes of Compatibility. |
| 1994-2 | Uranium Mill Tailings Regulations: Conforming NRC Requirements to EPA Standards Part 40 59 FR 28220 | 07/01/1997 | Not Applicable | Not Applicable | Georgia does not have authority to regulate this material under its Agreement. |
| 1994-3 | Timeliness in Decommissioning Material Facilities Parts 30, 40 and 70 59 FR 36026 | 08/15/1997 | Final | No Comments 09/25/1998 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|---|--------------------------------|--------------------|---------------------------|-------|
| 1995-1 | Preparation, Transfer for Commercial Distribution, and Use of Byproduct Material for Medical Use Parts 30, 32 and 35 59 FR 61767; 59 FR 65243; 60 FR 322 | 01/01/1998 | Final | No Comments 09/25/1998 | |
| 1995-2 | Frequency of Medical Examinations for Use of Respiratory Protection Equipment Part 20 60 FR 7900 | 03/13/1998 | Final | No Comments 09/25/1998 | |
| 1995-3 | Low-Level Waste Shipment Manifest Information and Reporting Parts 20 and 61 60 FR 15649; 60 FR 25983 | 03/01/1998 | Final | No Comments 09/25/1998 | |
| 1995-4 | Performance Requirements for Radiography Equipment Part 34 60 FR 28323 (Superceded by 1997-5) | 06/30/1998 | Final | No Comments 03/16/1994 | |
| 1995-5 | Radiation Protection Requirements: Amended Definitions and Criteria Parts 19 and 20 60 FR 36038 | 08/14/1998 | Final | No Comments 09/25/1998 | |
| 1995-6 | Clarification of Decommissioning Funding Requirements Parts 30, 40 and 70 60 FR 38235 | 11/24/1998 | Final | No Comments 09/25/1998 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|---|--------------------------------|----------------------|--|--|
| 1995-7 | Medical Administration of Radiation and Radioactive Materials Parts 20 and 35 60 FR 48623 (Superceded by 2002-2 and 2005-2) | 10/20/1998 | Final | No Comments 09/25/1998 | Georgia has adopted Final Regulations equivalent to RATS IDs: 2002-2 and 2005-2. |
| 1996-1 | Compatibility with the International Atomic Energy Agency Part 71 60 FR 50248; 61 FR 28724 (Superceded by 2004-1) | 04/01/1999 | Final | No Comments 09/25/1998 | Georgia has adopted Final Regulations equivalent to RATS ID: 2004-1. |
| 1996-2 | One Time Extension of Certain Byproduct, Source and Special Nuclear Materials Licenses Parts 30, 40 and 70 61 FR 1109 | 02/15/1999 | Not Required | Not Required | These regulation changes are not required to be adopted for purposes of Compatibility. |
| 1996-3 | Termination or Transfer of Licensed Activities: Record keeping Requirements Parts 20, 30, 40, 61 and 70 61 FR 24669 | 06/17/1999 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 1997-1 | Resolution of Dual Regulation of Airborne Effluents of Radioactive Materials; Clean Air Act Part 20 61 FR 65120 | 01/9/2000 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 1997-2 | Recognition of Agreement State Licenses in Areas Under Exclusive Federal Jurisdiction Within an Agreement State Part 150 62 FR 1662 | 02/27/2000 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|--|--------------------------------|----------------------|--|---|
| 1997-3 | Criteria for the Release of Individuals Administered Radioactive Material Parts 20 and 35 62 FR 4120 | 05/29/2000 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 1997-4 | Fissile Material Shipments and Exemptions Part 71 62 FR 5907 (Superceded by 2004-1) | 02/10/2000 | Not Required | Not Required | These regulation changes are not required to be adopted for purposes of Compatibility. (See STP-97-078) |
| 1997-5 | Licenses for Industrial Radiography and Radiation Safety Requirements for Industrial Radiography Operations Parts 30, 34, 71 and 150 62 FR 28947 | 06/27/2000 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 1997-6 | Radiological Criteria for License Termination Parts 20, 30, 40 and 70 62 FR 39057 | 08/20/2000 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 1997-7 | Exempt Distribution of a Radioactive Drug Containing One Micro curie of Carbon-14 Urea Part 30 62 FR 63634 | 01/02/2001 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 1998-1 | Deliberate Misconduct by Unlicensed Persons Parts 30, 40, 61, 70, 71 and 150 63 FR 1890; 63 FR 13773 | 02/12/2001 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|--|--------------------------------|--------------------------------|--|---|
| 1998-2 | Self-Guarantee of Decommissioning Funding by Nonprofit and Non-Bond-Issuing Licensees Parts 30, 40 and 70 63 FR 29535 | 07/01/2001 | Not Required | Not Required | These regulation changes are not required to be adopted for purposes of Compatibility. |
| 1998-3 | License Term for Medical Use Licenses Part 35 63 FR 31604 (Superceded by 2002-2) | 07/10/2001 | Not Required | Not Required | These regulation changes are not required to be adopted for purposes of Compatibility. (See STP-98-074) Georgia has adopted Final Regulations equivalent to RATS ID: 2002-2. |
| 1998-4 | Licenses for Industrial Radiography and Radiation Safety Requirements for Industrial Radiographic Operations Part 34 63 FR 37059 | 07/09/2001 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 1998-5 | Minor Corrections, Clarifying Changes, and a Minor Policy Change Parts 20, 32, 35, 36 and 39 63 FR 39477; 63 FR 45393 | 10/26/2001 | R ⁱⁱ ML14190A251 | Comments 08/18/2014 ML14188B167 | |
| 1998-6 | Transfer for Disposal and Manifests: Minor Technical Conforming Amendment Part 20 63 FR 50127 | 11/20/2001 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 1999-1 | Radiological Criteria for License Termination of Uranium Recovery Facilities Part 40 64 FR 17506 | 06/11/2002 | Not Applicable | Not Applicable | Georgia does not have the authority to regulate this material under its Agreement. |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|--|--------------------------------|--|---|--|
| 1999-2 | Requirements for Those Who Possess Certain Industrial Devices Containing Byproduct Material to Provide Requested Information Part 31 64 FR 42269 | 10/04/2002 | Not Required | Not Required | These regulation changes are not required to be adopted for purposes of Compatibility. |
| 1999-3 | Respiratory Protection and Controls to Restrict Internal Exposure Part 20 64 FR 54543; 64 FR 55524 | 02/02/2003 | R ML14190A251 | No Comments 08/18/2014 ML14188B167 | |
| 2000-1 | Energy Compensation Sources for Well Logging and Other Regulatory Clarifications Part 39 65 FR 20337 | 05/17/2003 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 2000-2 | New Dosimetry Technology Parts 34, 36 and 39 65 FR 63750 | 01/08/2004 | Final ML021540172 | No Comments 06/26/2002 ML021790295 | |
| 2001-1 | Requirements for Certain Generally Licensed Industrial Devices Containing Byproduct Material Parts 30, 31 and 32 65 FR 79162 | 02/16/2004 | R ML14190A251 Final ML15225A080 | No Comments 08/18/2014 ML14188B167 No Comments 10/21/2015 | |
| 2002-1 | Revision of the Skin Dose Limit Part 20 67 FR 16298 | 04/05/2005 | Final ML090280212 | ML15225A068 No Comments 02/20/2009 ML090340114 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|---|--------------------------------|-------------------------------------|--|-------|
| 2002-2 | Medical Use of Byproduct Material Parts 20, 32 and 35 67 FR 20249 | 10/24/2005 | Final ML032100723 | No Comments 07/31/2003 ML032160015 | |
| 2003-1 | Financial Assurance for Materials Licensees Parts 30, 40 and 70 68 FR 57327 | 12/03/2006 | Final ML090280212 | No Comments 02/20/2009 ML090340114 | |
| 2004-1 | Compatibility With IAEA Transportation Safety Standards and Other Transportation Safety Amendments Part 71 69 FR 3697 | 10/01/2007 | Final ML090280212 | No Comments 02/20/2009 ML090340114 | |
| 2005-1 | Security Requirements for Portable Gauges Containing Byproduct Material Part 30 70 FR 2001 | 07/11/2008 | Final ML090280212 | No Comments 02/20/2009 ML090340114 | |
| 2005-2 | Medical Use of Byproduct Material - Recognition of Specialty Boards Part 35 70 FR 16336; 71 FR 1926 | 04/29/2008 | Final ML090280212 | No Comments 02/20/2009 ML090340114 | |
| 2005-3 | Increased Controls for Risk-Significant Radioactive Sources (NRC Order EA-05-090) 70 FR 72128 | 12/01/2005 | License Condition ML14139A035 | No Comments 0721/2014 ML14139A033 | |
| 2006-1 | Minor Amendments Parts 20, 30, 32, 35, 40 and 70 71 FR 15005 | 03/27/2009 | Final ML090280212 | No Comments 02/20/2009 ML090340114 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|--|--------------------------------|------------------------------|--|-------|
| 2006-2 | National Source Tracking System - Serialization Requirements Part 32 with reference to Part 20 Appendix E 71 FR 65685 | 02/06/2007 | Final ML090280212 | No Comments 02/20/2009 ML090340114 | |
| 2006-3 | National Source Tracking System Part 20 71 FR 65685, 72 FR 59162 | 01/31/2009 | Final ML090280212 | No Comments 02/20/2009 ML090340114 | |
| 2007-1 | Medical Use of Byproduct Material - Minor Corrections and Clarifications Parts 32 and 35 72 FR 45147, 54207 | 10/29/2010 | Final ML090280212 | No Comments 02/20/2009 ML090340114 | |
| 2007-2 | Exemptions From Licensing, General Licenses, and Distribution of Byproduct Material: Licensing and Reporting Requirements Parts 30, 31, 32 and 150 72 FR 58473 | 12/17/2010 | Final ML13364A243 | No Comments 02/20/2014 ML13364A230 | |
| 2007-3 | Requirements for Expanded Definition of Byproduct Material Parts 20, 30, 31, 32, 33, 35, 61 and 150 72 FR 55864 | 11/30/2010 | R ML14190A251 | No Comments 08/18/2014 ML14188B167 | |
| | 72 FR 33604 | | Final ML15225A080 | Comments 10/21/2015 ML15225A068 | |
| | | | Revised Final ML18005B166 | No Comments 03/07/2018 ML18005B161 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|--|--------------------------------|---|---|-------|
| 2007-4 | Order Imposing Fingerprinting Requirements and Criminal History Records Check Requirements for Unescorted Access to Certain Radioactive Material NRC Order EA-07-305 72 FR 70901 | 06/05/2008 | License Condition ML14139A035 | No Comments 07/21/2014 ML14139A033 | |
| 2008-1 | Occupational Dose Records, Labeling Containers, and Total Effective Dose Equivalent Parts 19 and 20 72 FR 68043 | 02/15/2011 | Final ML13364A243 | No Comments 02/20/2014 ML13364A230 | |
| 2009-1 | Medical Use of Byproduct Material – Authorized User Clarification Part 35 74 FR 33901 | 09/28/2012 | R ML14190A251 Final ML15225A080 | Comments 08/18/2014 ML14188B167 No Comments 10/21/2015 ML15225A068 | |
| 2011-1 | Decommissioning Planning Parts 20, 30, 40 and 70 76 FR 35512 | 12/17/2015 | Proposed ML15224B578 Final ML15336A023 | Comments 09/17/2015 ML15224B576 No Comments 12/11/2015 ML15336A022 | |
| 2011-2 | Licenses, Certifications, and Approvals for Materials Licensees Parts 30, 36, 39, 40, 70 and 150 76 FR 56951 | 11/14/2014 | Proposed ML14190A251 | No Comments 08/18/2014 ML14188B167 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|---|--------------------------------|---|--|-------|
| 2012-1 | Change of Compatibility of Parts 31.5 and 31.6 (See RATS ID: 2001-1 for Rule text) 77 FR 3640 | 01/25/2015 | Proposed ML14190A251 | No Comments 08/18/2014 ML14188B167 | |
| 2012-2 | Advance Notification to Native American Tribes of Transportation of Certain Types of Nuclear Waste Part 71 77 FR 34194 | 08/10/2015 | Proposed ML14190A251 Final ML15225A080 Revised Final ML18005B166 | Comments 08/18/2014 ML14188B167 Comments 10/21/2015 ML15225A068 No Comments 03/07/2018 ML18005B161 | |
| 2012-3 | Technical Corrections Parts 30, 34, 40 and 71 77 FR 39899 | 08/06/2015 | Proposed ML14190A251 Final ML15225A080 | Comments 08/18/2014 ML14188B167 No Comments 10/21/2015 ML15225A068 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|--|--------------------------------|-------------------------------------|--|---------------------------------|
| 2012-4 | Requirements for Distribution of Byproduct Material Parts 30, 31, 32, 40 and 70 77 FR 43666 | 10/23/2015 | Proposed ML14190A251 | Comments 08/18/2014 ML14188B167 | |
| | | | Final ML15225A080 | Comments 10/21/2015 ML15225A068 | |
| | | | Revised Final ML18005B166 | Comment 03/07/2018 ML18005B161 | |
| | | | Revised Final ML19065A113 | No Comments 04/04/2019 ML19065A108 | |
| 2013-1 | Physical Protection of Byproduct Material, Parts 20, 30, 32, 33, 34, 35, 36, 37, 39 and 71 78 FR 16922 | 03/19/2016 | License Condition ML15020A328 | Comment 02/11/2015 ML15020A322 | Part 37; Comment was re: 30.13. |
| | | | Proposed ML15224B578 | No Comments 09/17/2015 ML15224B576 | Georgia submitted 30.13 only |
| | | | Final ML16272A083 | No Comments 10/19/2016 ML16272A065 | Georgia submitted 30.13 only |
| | | | Proposed ML17117A338 | Comment 06/15/2017 ML17117A337 | All Parts |
| | | | | | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|--|--------------------------------|------------------------------|--|-------|
| *2013-2 | Distribution of Source Material to Exempt Persons and to General Licensees and Revision of General License and Exemptions, | 08/27/2016 | Proposed ML16267A235 | Comments 10/25/2016 ML16267A230 | |
| | Parts 30, 40 and 70 78 FR 32310 | | Final ML17177A087 | Comments 10/25/2016 ML16267A230 Comments 07/13/2017 ML17177A086 Comments 04/04/2019 ML19065A108 Comments 08/01/2019 ML19164A100 No Comments 06/15/2017 ML17117A337 No Comments 03/07/2018 ML18005B161 No Comments 04/04/2019 ML19065A108 | |
| | | | Revised Final ML19065A113 | 04/04/2019 | |
| | | | Revised Final ML19175A257 | 08/01/2019 | |
| 2015-1 | Domestic Licensing of Special Nuclear Material – Written Reports and Clarifying Amendments Part 70 | 01/26/2018 | Proposed ML17117A338 | 06/15/2017 | |
| | 79 FR 57721, 80 FR 143 | | Final ML18005B166 | 03/07/2018 | |
| | | | Revised Final ML19065A113 | 04/04/2019 ML19065A108 | |
| 2015-2 | Safeguards Information - Modified Handling Categorization, Change for Materials Facilities Parts 30, 37, 73 and 150 | 01/28/2018 | Proposed ML17117A338 | No Comments 06/15/2017 ML17117A337 | |
| | 79 FR 58664, 80 FR 3865 | | Final ML18005B166 | No Comments 03/07/2018 ML18005B161 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|---|--|------------------------------|---|-------|
| 2015-3 | Revisions to Transportation Safety Requirements and Harmonization with International Atomic Energy Agency Transportation Requirements | 07/13/2018 *extended to 08/15/2020 | Proposed ML17117A338 | No Comments 06/15/2017 ML17117A337 Comments 03/07/2018 ML18005B161 No Comments 04/04/2019 ML19065A108 No Comments 06/15/2017 ML17117A337 No Comments 06/15/2017 ML17117A337 No Comments 03/07/2018 ML18005B161 No Comments 03/07/2018 ML18005B161 No Comments 03/07/2018 ML18005B161 No Comments 06/15/2017 ML17117A337 Comment | |
| | Part 71 80 FR 33987 | See STC 17-060 | Final ML18005B166 | 03/07/2018 | |
| | | | Revised Final ML19065A113 | 04/04/2019 | |
| 2015-4 | Miscellaneous Corrections, Parts 37 and 40 80 FR 45841 | 09/02/2018 | Proposed ML17117A338 | 06/15/2017 | |
| | | | Final ML18005B166 | 03/07/2018 ML18005B161 | |
| 2015-5 | Miscellaneous Corrections, Parts 19, 20, 30, 32, 37, 40, 61, 70, 71 and 150 80 FR 74974 | 12/31/2018 | Proposed ML17117A338 | 06/15/2017 | |
| | 0011(14014 | | Final ML18005B166 | Comment 03/07/2018 ML18005B161 | |
| | | | Revised Final ML19065A113 | No Comments 04/04/2019 ML19065A108 | |
| *2018-1 | Medical Use of Byproduct Material – Medical Event Definitions, Training and Experience, and Clarifying Amendments, 10 CFR Parts 30, 32 and 35 | 01/14/2022 | Proposed ML19175A257 | Comments 08/01/2019 ML19164A100 | |

| RATS ID | NRC Chronology Identification | Date Due for State Adoption | Incoming Letter | Outgoing Package | Notes |
|---------|---|--------------------------------|--------------------|--|-------|
| *2018-2 | Miscellaneous Corrections – Organizational Changes 10 CFR Parts 37, 40. 70 and 71 | 12/21/2021 | ML19175A257 | No Comments 08/01/2019 ML19164A100 | |

ⁱ IMPEP Team: verify that Georgia does not have any licensees subject to these regulations during each review. ⁱⁱ R stands for proposed changes to a final rule.