

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: 49-50 P.O. Box 468, Soldotna, Alaska Location Inspected: 38240 Kenal Spur Highway, Soldotna, Alaska REPORT NO.: 2019-001		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region IV, 1600 East Lamar Blvd Arlington, Texas 76011-4511	
3. DOCKET NUMBER 030-20575	4. LICENSE NUMBER 50-21374-01	5. DATE(S) OF INSPECTION June 6, 2019	

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):




- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
 10 CFR 30.34(l) requires each portable gauge licensee to use a minimum of two independent physical controls that form tangible barriers to secure portable gauges from unauthorized removal, whenever portable gauges are not under the control and constant surveillance of the licensee.

Contrary to the above, on June 6, 2019, the licensee failed to use a minimum of two independent physical controls that form tangible barriers to secure portable gauges from unauthorized removal, whenever portable gauges are not under the control and constant surveillance of the licensee. Specifically, the licensee failed to use two independent physical controls on two of three portable gauges that were in storage at the licensee's facility during the inspection on June 6, 2019. This was determined to be a SLIV violation in accordance with NRC Enforcement Policy Example 6.3.d.10.

To correct the above, the licensee immediately provided an additional barrier to both the improperly secured gauges and provided additional training to the gauge users on the securing of gauges in storage.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Sam McLane, RSO		7/3/2019
NRC INSPECTOR	Jason vonEhr		7/3/19
BRANCH CHIEF	Patricia A. Silva		7/12/19

Non-Public
 Sensitive - Security-Related
 Public
 Non-Sensitive