NRC FORM 366
(7-77)

U. S. NUCLEAR REGULATORY COMMISSION

LICENSEE EVENT REPORT

EXHIBIT A

CONTROL BLOCK: (1) (PLEASE PRINT OR TYPE ALL REQUIRED INFORMATION)

01 | ARANO2 (2) | 00-000-00 (3) | 4 1111 (4) | (5)
LICENSEE CODE — LICENSE NUMBER — LICENSE TYPE — CAT

CON'T
01 | REPORT SOURCE L (6) | 05000368 (7) | 080879 (8) | 090579 (9)
DOCKET NUMBER — EVENT DATE — REPORT DATE

EVENT DESCRIPTION AND PROBABLE CONSEQUENCES (10)

02 During Mode 1 operation, the Auxiliary Building Radwaste Area Radiation
03 Monitor, 2RE-E542, sample pump was found with a broken coupling. Gaseous
04 releases which utilize the subject monitor were sampled and analyzed
05 with less than minimum detectable results.
06 This occurrence is similar to LER 50-368/79-057.
07 Reportable per E.T.S. 5.6.2b.
08

09 SYSTEM CODE MC (11) | CAUSE CODE E (12) | CAUSE SUBCODE B (13) | COMPONENT CODE INSTRU (14) | COMP SUBCODE E (15) | VALVE SUBCODE Z (16)

(17) LER/RO REPORT NUMBER: EVENT YEAR 79 — SEQUENTIAL REPORT NO 067 — OCCURRENCE CODE 04 — REPORT TYPE L — REVISION NO 0

ACTION TAKEN A (18) | FUTURE ACTION Z (19) | EFFECT ON PLANT Z (20) | SHUTDOWN METHOD Z (21) | HOURS (22) 0000 | ATTACHMENT SUBMITTED N (23) | NPRD-4 FORM SUB N (24) | PRIME COMP SUPPLIER A (25) | COMPONENT MANUFACTURER W120 (26)

CAUSE DESCRIPTION AND CORRECTIVE ACTIONS (27)

10 Appropriate grab samples were taken while the monitor was out of service,
11 once per shift, as required by E.T.S. 2.2.2.7. The plastic coupling
12 was replaced with a metal coupling.
13
14

15 FACILITY STATUS R (28) | % POWER 050 (29) | OTHER STATUS (30) Pwr.Escal. Test | METHOD OF DISCOVERY A (31) Routine | DISCOVERY DESCRIPTION (32) operator tour

16 ACTIVITY RELEASED Z (33) | CONTENT OF RELEASE Z (34) | AMOUNT OF ACTIVITY (35) NA | LOCATION OF RELEASE (36) NA

PERSONNEL EXPOSURES
17 NUMBER 000 (37) | TYPE Z (38) | DESCRIPTION (39) NA

PERSONNEL INJURIES
18 NUMBER 000 (40) | DESCRIPTION (41) NA

LOSS OF OR DAMAGE TO FACILITY (43)
19 TYPE Z (42) | DESCRIPTION NA

PUBLICITY
20 ISSUED N (44) | DESCRIPTION (45) NA

NRC USE ONLY

NAME OF PREPARER Chris N. Shively PHONE 501/968-2519

943 137 7909120 567