

TIC

Superior Memorial Hospital, Inc.

3500 TOWER AVENUE - SUPERIOR, WISCONSIN 54880 - PHONE 392-8281
Area Code 715

CHARLES A. LAIBLE
Administrator

POOR ORIGINAL

Dec. 9, 1979

United States Nuclear Regulatory Commission
Region III
Glen Ellyn, Illinois

Gentlemen:

In response to your bulletin No. 79-19.

Your Question No. I.

Approximately twice a year, we send 20 Molybdenum MO99 Technicium 99m generators back to the manufactures E. R. Squibb in N. J.

Question No. 2.

These Tc99m generators are completely decayed before shipment and read Q mr on testing.

Question #3.

We do not generate any type of radioactive waste.

Our license #49-17367-01

Sincerely,

Curt Jensen
Nuclear Medicine Dept.
Superior, Memorial Hospital
Superior, Wisconsin

CJ:nl

~~2289 018~~

1744 324

DEC 12 1979

8001080

248