

DEC 11 1980

Canton Medical Care Center,
Incorporated
ATTN: Anselm A. Ofodile, M.D.
7288 North Sheldon Road
Canton Township, Michigan 48137

Gentlemen:

This is in reference to your application for an NRC license for medical diagnosis. So that we may complete our review, please supply the following additional information.

1. Your application does not include a personnel training program. This should include nuclear medicine technologists and ancillary personnel, such as housekeeping, and should include the subjects and frequency described in item 12 of the Medical Licensing Guide.
2. Your procedures for cleaning areas (after area surveys) specify an action level of 0.05 microcurie. The current standard (published in the Medical Licensing Guide) is 200 dpm/100 cm². You may use this standard or you may submit an alternate procedure for assaying the wipes with your survey meter. In the latter case, you should change the limit to any activity above background. This procedure should also include geometry and counting time (resolving time).
3. Your application should also include your ALARA program. We are enclosing a copy of the model ALARA program for your use.

We will continue our review upon receipt of the above information. Please reply in duplicate and refer to Control No. 04098.

Sincerely,

John W. Cooper, Ph.D., Chief
Regional Licensing Section
Material Licensing Branch
Division of Fuel Cycle and
Material Safety

81031100-14

Enclosures:

1. Model ALARA program
2. Regulatory Guide 10.8

OFFICE

RIII

SURNAME

Cooper/JW

DATE

12/10/80