

### SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Oakland Imaging Diagnostic Center 27483 Dequindre Road Suite 101 Madison Heights, MI 48071 REPORT NUMBER(S) 2019001		2. NRC/REGIONAL OFFICE Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
3. DOCKET NUMBER(S) 030-30646	4. LICENSE NUMBER(S) 21-25929-01	5. DATE(S) OF INSPECTION July 29, 2019

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:



- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, to exercise discretion, were satisfied.

\_\_\_\_\_ Non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.  
(Violations and Corrective Actions)

#### Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Edward F. Harvey		8/23/2019
BRANCH CHIEF	Aaron T. McCraw		08/26/2019

**Docket File Information**

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED:  Oakland Imaging Diagnostic Center 27483 Dequindre Road Suite 101 Madison Heights, MI 48071  REPORT NUMBER(S) 2019001	2. NRC/REGIONAL OFFICE  Region III U. S. Nuclear Regulatory Commission 2443 Warrenville Road, Suite 210 Lisle, IL 60532-4352
--	---

3. DOCKET NUMBER(S)  030-30646	4. LICENSE NUMBER(S)  21-25929-01	5. DATE(S) OF INSPECTION  July 29, 2019
--------------------------------------	---	---

6. INSPECTION PROCEDURES USED  87130	7. INSPECTION FOCUS AREAS  All
--	--------------------------------------

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S)  02201	2. PRIORITY  5	3. LICENSEE CONTACT  Patricia Ciotti, Manager	4. TELEPHONE NUMBER  (248) 544-9050
---------------------------------	----------------------	---	---

Main Office Inspection      Next Inspection Date: July 29, 2024

Field Office Inspection \_\_\_\_\_

Temporary Job Site Inspection \_\_\_\_\_

**PROGRAM SCOPE**

This was a routine, unannounced inspection of a private practice physician's owned clinic authorized to possess and use unsealed byproduct material for diagnostic medical procedures at its location in Madison Heights, Michigan. The licensee employed one full-time nuclear medicine technologist (NMT) that performed approximately four diagnostic procedures per day. The licensee retained the services of a medical physics consultant for instrument calibrations and quarterly program audits.

The inspector toured the licensee's facility in Madison Heights to evaluate the licensee's measures for material security, hazard communication, and exposure control. The inspector was unable to observe any administrations of licensed material as there were no procedures scheduled for the remainder of the day. The inspector observed the NMT demonstrate package receipt procedures, dose calibrator quality control, and area survey procedures. The NMT also demonstrated adequate knowledge of radiation protection principles and emergency procedures through interviews with the inspector.

The inspector reviewed a selection of licensee records, including quarterly audits, source inventories, survey meter calibration records, package receipt logs, and dosimetry with no issues noted. In addition, the inspector performed independent surveys, which revealed no readings that would indicate residual contamination or exposures to members of the public in excess of regulatory limits.

No violations of NRC requirements were identified during this inspection.