

Tran, Frank

From: Velez, Vanessa <vvelez@usgs.gov>
Sent: Monday, June 24, 2019 11:47 AM
To: Tran, Frank
Subject: [External_Sender] Re: [EXTERNAL] Request for additional information regarding NRC License No. 24-12728-01
Attachments: 20190620_122727.jpg; 20190620_122704.jpg; 20190620_122851.jpg

Mail Control Number 612424

Dear Mr. Tran,

This is a response to the email that you sent on Thursday, June 20, 2019 requesting additional information for Mail Control Number 612424. Please see attachments with evidence that the devices are labeled general licensed by the manufacturer.

Thanks,

Vanessa Velez
USGS Columbia Environmental Research Center
Radiation Safety Officer

Phone: 573-441-2958
Fax: 573-876-1896

On Thu, Jun 20, 2019 at 8:20 AM Tran, Frank <Frank.Tran@nrc.gov> wrote:

Dear Ms. Velez:

We have reviewed your amendment request dated June 6, 2019. In order to change a device which may be used under a general license in accordance with Title 10 of the Code of Federal Regulations (10 CFR) Part 31 from a specific license to a general license, a general license label must be adhered to the general device by the manufacturer or vendor and the instruction on that label must be followed. Please provide a copy (or image) of the label for each device which can be readable.

Please provide the above information by July 19, 2019. In addition, please reference in the response that the information is an additional information for Mail Control Number 612424. For this request, you could fax the response to 630-515-1078 or email to frank.tran@nrc.gov.

If you have any questions, please do not hesitate to contact me at 630-829-9623 or reply to this email.

Thank you,

Frank Tran

Health Physicist/License Reviewer

NRC Region III/Division of Nuclear Materials Safety

Phone: 630-829-9623

Fax: 630-515-1078

Email: Frank.Tran@nrc.gov





CAUTION

RADIOACTIVE MATERIAL

15 MILLICURIES OF RADIUM-226

G2337A ELECTRON CAPTURE DETECTOR

SERIAL NO.

U2340

DATE

07

00

Agilent Technology



SEE THE 6890 GC MANUAL FOR
OPERATING OR SERVICING

THIS DEVICE MUST BE WIPED (RADIATION)
TESTED AT INTERVALS NOT GREATER THAN 1
MONTHS.

THE RECEIPT, POSSESSION, USE AND TRANSFER
OF THIS DEVICE ARE SUBJECT TO A GENERAL
LICENSE OR THE EQUIVALENT, AND THE
REGULATIONS OF THE U.S. NRC OR OF A STATE
WITH WHICH THE NRC HAS ENTERED INTO AN
AGREEMENT FOR THE EXERCISE OF REGULATORY
AUTHORITY. OUTSIDE U.S., CONSULT THE
APPROPRIATE AGENCIES TO DETERMINE THE
EQUIVALENT REGULATIONS.

THE LABEL SHALL BE MAINTAINED ON THIS
DEVICE IN A LEGIBLE CONDITION. REMOVAL OF
THIS LABEL IS PROHIBITED.

MADE IN U.S.A.



**CAUTION
RADIOACTIVE MATERIAL**

15 MILLICURIES OF NICKEL 63

G2397A ELECTRON CAPTURE DETECTOR

SERIAL NO.

U8605

DATE

06 05



Agilent Technologies



SEE THE 6890 GC MANUALS BEFORE
OPERATING OR SERVICING THIS DEVICE.

THIS DEVICE MUST BE WIPE (RADIOACTIVE LEAK)
TESTED AT INTERVALS NOT GREATER THAN SIX
MONTHS.

THE RECEIPT, POSSESSION, USE AND TRANSFER
OF THIS DEVICE ARE SUBJECT TO A GENERAL
LICENSE OR THE EQUIVALENT, AND THE
REGULATIONS OF THE U.S. NRC OR OF A STATE
WITH WHICH THE NRC HAS ENTERED INTO AN
AGREEMENT FOR THE EXERCISE OF REGULATORY
AUTHORITY. OUTSIDE U.S., CONSULT THE
APPROPRIATE AGENCIES TO DETERMINE THE
EQUIVALENT REGULATIONS.

THE LABEL SHALL BE MAINTAINED ON THIS
DEVICE IN A LEGIBLE CONDITION. REMOVAL OF
THIS LABEL IS PROHIBITED.

MADE IN USA



CAUTION

RADIOACTIVE MATERIAL

15 MILLICURIES OF NICKEL 63

G2397A ELECTRON CAPTURE DETECTOR

SERIAL NO.

U3-127

DATE

2015-01



Agilent Technologies



SEE THE 6890 GC MANUALS BEFORE
OPERATING OR SERVICING THIS DEVICE.

THIS DEVICE MUST BE WIPE (RADIOACTIVE LEAK)
TESTED AT INTERVALS NOT GREATER THAN SIX
MONTHS.

THE RECEIPT, POSSESSION, USE AND TRANSFER
OF THIS DEVICE ARE SUBJECT TO A GENERAL
LICENSE OR THE EQUIVALENT, AND THE
REGULATIONS OF THE U.S. NRC OR OF A STATE
WITH WHICH THE NRC HAS ENTERED INTO AN
AGREEMENT FOR THE EXERCISE OF REGULATORY
AUTHORITY. OUTSIDE U.S., CONSULT THE
APPROPRIATE AGENCIES TO DETERMINE THE
EQUIVALENT REGULATIONS.

THE LABEL SHALL BE MAINTAINED ON THIS
DEVICE IN A LEGIBLE CONDITION. REMOVAL OF
THIS LABEL IS PROHIBITED.

MADE IN USA