

DOCKET NUMBER

PROPOSED RULE

PR-30,35 ①

DONALD R. HAMILTON

325 WESSLING CIRCLE

CATOCESVILLE, MARYLAND 21228

June 12, 1979

Testing of Radioisotope
Generators

NRC PUBLIC DOCUMENT

(44 FR 32394)

Secretary of the Commission
U.S. Nuclear Regulatory Commission
Washington, DC 2055


Attn: Docketing and Service Branch

Gentlemen:

I have read the proposed rule (Federal Register, 44, 32394-5
June 6, 1979) concerning the testing of radionuclide generators
elutions and agree with it.

The proposed changes in Parts 30 and 35 will clarify the condi-
tions and frequency that eluates must be checked before patient
administration. The ambiguities of current package inserts for
these products have led to confusion among users. Additionally,
since moly breakthroughs are infrequent, many nuclear medicine
facilities have discontinued testing. Your proposed rule should
be adopted without change.

Sincerely yours,



Donald R. Hamilton, CNMT



Acknowledged by card... 6-20-79, ers

343 360

7907230 779