WISCONSIN CLINICAL CANCER CENTER

University of Wisconsin • 600 Highland Ave. • Madison, WI 53792

DEPARTMENT OF HUMAN ONCOLOGY

Division of Radiation Oncology (608) 263-8500

G. W. Wirtanen, M.D., Director

A. L. Wiley, M.D., Ph.D., Assoc. Director

Section of

Radiation Therapy: 263-8500

A. L. Wiley, M.D., Ph.D.

D. A. Buchler, M.D.

A. D. Duszynski, D.D.S.

A. J. Greenberg, M.D.

M. T. Kademian, M.D.

J. C. (Kline) Puletti, M.D. R. A. Steeves, M.D., Ph.D.

G. W. Wirtanen, M.D.

December 17, 1980

Section of Radiation Therapy Physics: 263-8500 B. R. Paliwal, Ph.D., Director L. W. Asp, M.S. B. R. Thomadsen, M.S.

J. A. Zagzebski, Ph.D., 262-4921

Section of Radiation Biology: 263-5342 K. H. Clifton, Ph.D., Director M. N. Gould, Ph.D. E. D. Werts, Ph.D.

P.P.Carbone, M.D., Director

George W. Wirtanen, M.D.

Associate Director, WCCC

Barbara L. Burr, Administrative Assistant

M. B. Yatvin, Ph.D.

License No. 48-09843-31

A.B. Davis Fuel Facility and Materials Safety Branch U.S. Nuclear Regulatory Commission 799 Roosevelt Road Glen Ellyn, 11 60137

Dear Mr. Davis:

This letter is in response to the November 28, 1980 notice of noncompliance from your office as a result of the inspection conducted on November 18, 1980 at our institution, on our Cobalt 60 teletherapy treatment unit.

Item of noncompliance: determination of timer accuracy was performed less frequently than required.

- (i) Corrective action taken and results achieved: none
- Corrective action to be taken to avoid further noncompliance: determination of timer accuracy will be performed at intervals not exceeding one month.
- (3) The date when full compliance will be achieved: December 4, 1980

Sincerely,

Llord W. asp

Lloyd W. Asp, M.S.

Physicist

cot Dr. George Wirtanen Director, Radiation Oncology

LWA/mas

8108020030

I hereby affirm that the above stated information is true.

The above signature was affixed in my presence by one known to me to be Lloyd W. Asp. Barbara T. Bury 1-6-81 County of Dana State of Micronain my commission ount-4/22

JAN 0 7 1981