

Br. 2

NRC FORM 314 <small>(6-2004)</small> <small>10 CFR 30.26(f)(1); 40.42(f)(1);</small> <small>70.32(f)(1); and 72.54(f)(1)</small>	U.S. NUCLEAR REGULATORY COMMISSION	APPROVED BY OMB: NO. 3160-0028	EXPIRES: 06/30/2007 <small>Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 FSZ), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by Internet e-mail to infocentre@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NE08-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.</small>
--	---	---------------------------------------	--

CERTIFICATE OF DISPOSITION OF MATERIALS

LICENSEE NAME AND ADDRESS Dessarrolladora J.A. Inc. Box 343 Isabela, PR 00662	LICENSE NUMBER 52-31266-01	DOCKET NUMBER 030-37515 LICENSE EXPIRATION DATE 12/31/2032
---	--------------------------------------	---

A. LICENSE STATUS (Check the appropriate box)

This license has expired. This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL
(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

1. No radioactive materials have ever been procured or possessed by the licensee under this license.

2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:

a. Transfer of radioactive materials to the licensee listed below:
CPN, Division of Insirotek, Inc, 5052 Commercial Circle Concord, CA 94620 U.S.A.

b. Disposal of radioactive materials:

1. Directly by the licensee:

2. By licensed disposal site:

3. By waste contractor:

c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:

a. the absence of licensed radioactive materials

b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.

2. A copy of the radiation survey results:

a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date

3. A radiation survey is not required as only sealed sources were ever possessed under this license, and


a. The results of the latest leak test are attached; and/or b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME Alejandro J. Abrams, M.E.C.E.	TITLE	TELEPHONE (Include Area Code) (787) 382-4033	E-MAIL ADDRESS
--	--------------	--	-----------------------

Mail all future correspondence regarding this license to:
 Dessarrolladora J.A. Inc. Box 343 Isabela, PR 00662

C. CERTIFYING OFFICIAL
 I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE Alejandro J. Abrams, M.E.C.E.	SIGNATURE 	DATE 07/31/2018
--	--	---------------------------

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

REC'D IN LAT 8/28/19

614079
 NMSS/RGN1 MATERIALS-002

Leak Test Record
NRC License 52-25430-01

Leak Tested For:	Desarrolladora J. A. Inc.	
Leak Tested By:	David Rhoe	
Standard Source (NIST traceable):	Am-241	Cs-137 NES-139S
Standard Activity (uCi):	1.145	0.105
Standard Date	15-Nov-98	09-Sep-88
Date of the Leak Test:	10-Jun-19	
Decay Activity uCi (from decay chart):	1.10784	0.05161
Standard (dpm):	2459404.8	114574.2
Instrument used to count wipe sample:	Ludlum	
Instrument Model Number:	2200	
Instrument Serial Number:	185815	
NIST Traceable Standard (cpm)	147681	13593
Counting Efficiency:	0.06	0.12
Counting Efficiency in percentage (%):	6.01	11.86
Counting time (minutes)	1	1
Background (cpm)	172	172
Minimum Detectable Activity:	1.391E-04	7.042E-05

Wipe (Smear) Test: All external or accessible surfaces of the source or housing are wiped with a piece of filter paper or other absorbent material which has been moistened with an appropriate solvent and the activity removed is measured.
 Note: Background counts were not subtracted from wipe test sample to calculate sample activity.

Source ID and Serial Number	Wipe Test	Gamma	
		Sample Activity	Sample Activity
Am-241 & Cs-137 SnMD71108885	175	0.00131	0.00066

This test reveals that 0.005 microcuries or less was present as removable contamination. Should the removable contamination exceed 0.005 microcuries, the source must be removed from use and necessary measures taken according to NRC regulations.


 David Rhoe Health/Medical Physicist



CPN, Division of Instratek, Inc.
 5052 Commercial Circle
 Concord, CA 94520 U.S.A.
 Ph: 925-383-9770
 www.InstroTek.com

LETTER OF RECEIPT OF GAUGE / SEALED SOURCE

To Radioactive Material Licensing Agencies: CPN International a Division of Instratek, Inc. has taken possession of the following nuclear gauge(s). A certificate of leak test is on file.

Gauge Manufacturer	Model	Serial Number(s)	Nuclide	Nuclide Activity
CPN	MC1DR	MD71108885	Cesium137 Ambe241	10 mCi 50 mCi

Gauge(s) transferred from:

CRMI

David Rhoe

D-4 Calle Tivoli

Paseo de la fuente

San Juan, PR 00926

Date Received: 7/31/2019

Signature:

JENNIFER RICHARDSON



ACKNOWLEDGEMENT - RECEIPT OF CORRESPONDENCE

Name and Address of Applicant and/or Licensee

Alejandro J. Abrams, M.E.C.E.
Vice President
Desarrolladora J.A., Inc.
P.O. Box 343
Isabela, Puerto Rico 00662

Date

September 4, 2019

License Number(s)

52-31266-01

Mail Control Number(s)

614079

Licensing and/or Technical Reviewer or Branch

Commercial, Industrial, R&D, and Academic Branch
(NRC Form 314)

This is to acknowledge receipt of your: Letter and/or Application Dated: 07/31/2019

The initial processing, which included an administrative review, has been performed.

Amendment Termination New License Renewal

There were no administrative omissions identified during our initial review.

This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.

Your application for a new NRC license did not include your taxpayer identification number. Please fill out NRC Form 531, Request for Taxpayer Identification Number, located at the following link:
<http://www.nrc.gov/reading-rm/doc-collections/forms/nrc531.pdf>

Send the completed NRC Form 531, by facsimile, to the following number: (301) 415-5387

The following administrative omissions have been identified:

[Empty box for administrative omissions]

Your application has been assigned the above listed MAIL CONTROL NUMBER. When calling to inquire about this action, please refer to this control number. Your application has been forwarded to a technical reviewer. Please note that the technical review, which is normally completed within 180 days for a renewal application (90 days for all other requests), may identify additional omissions or require additional information. If you have any questions concerning the processing of your application, our contact information is listed below:

Select a location (Use keyboard arrows to select) . . .