

FRANKLIN and MARSHALL COLLEGE
LANCASTER, PENNSYLVANIA 17604

OFFICE of the PRESIDENT

May 15, 1979

Dear Mr. McClintock:


This letter is written in reply to your letter regarding the inspection conducted on March 30, 1979, by Mr. P. Jerman and the attached notice of violation. Mr. Jerman's inspection was thorough and professionally done, and we are appreciative of his efforts to help us maintain strict and careful control of radioactive sources.

With regard to Item A, a procedure has been established to schedule regular testing of sealed sources containing radioactive materials for contamination or leakage to insure that regular testing will occur within prescribed intervals. All such sources in the College's possession have been tested, and the College is now in full compliance with Condition 13.

With regard to item B, a proper label has been attached to the shielded container holding the sealed source, and the College is now in full compliance with 10 CFR 20.203 (f) (1) and (2).

Dr. Hood will regularly review the College's performance vis-a-vis NRC requirements to insure that the College continues its use and storage of radioactive materials in full compliance with those requirements.

Sincerely,



Richard Hoffman
Vice President, Budget & Planning

Mr. Robert O. McClintock, Chief
Materials Radiological Protection Section
United States Nuclear Regulatory Commission
Region 1
631 Park Avenue
King of Prussia, Pennsylvania 19406

cc: Dr. R.F. Hood

922 258
7909070 194